


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/19/2001 Time: 19:12:10

Sample: AD26321
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/6/01 19:10 Ended: 11/23/01 19:10 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.3		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzene		5.7		1.0
20	Surr_Triphenyl phosphate		5.9		1.0
21	Surr_Perylene-d12		5.0		1.0

Chromatogram Plot File: E:\26321A Date: Nov-24-1991 02:04:14
Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26321
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 8044517



Integration Report

Quanfile: 26321A

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26321

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:12	732	Auto	522,920	273,835
2	phenanthrene d-10 (IS)	16:48	1008	Auto	768,061	222,316
3	chrysene d-12 (IS)	22:35	1355	Auto	591,227	189,906
4	p-terphenyl d-14 (IS)	20:46	1246	Auto	3,599,366	2,524,330
5	acenaphthene d-10 (SURR)	12:12	732	Auto	522,920	273,835
6	phenanthrene d-10 (SURR)	16:48	1008	Auto	768,061	222,316
7	chrysene d-12 (SURR)	22:35	1355	Auto	591,227	189,906
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	359,529	187,998
9	triphenyl phosphate (S)	22:01	1321	Auto	386,573	152,500
10	perylene d-12 (SURR)	26:48	1608	Auto	343,669	48,982
11	bis(2-ethylhexyl)adipate	21:54	1314	Auto	27,314	4,772
12	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	267,630	70,384
13	2,6-dinitrotoluene	11:52	712	Auto	8,238	4,574
14	butachlor	20:22	1222	Auto	0	0

reversed IS/SURR's

Quantitation Report

Quanfile: 26321A

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26321

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	732	12:12	BB	0.959	UG/L
6	phenanthrene d-10(SURR)	A	1008	16:48	BV	0.980	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	0.982	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	5.730	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BB	5.898	UG/L
10	perylene d-12 (SURR)	A	1608	26:48	BB	5.035	UG/L
16	bis(2-ethylhexyl)adipate	A	1314	21:54	MM	0.206	UG/L
17	bis(2-ethylhexyl)phthalate	A	1373	22:53	BV	1.268	UG/L
20	2,6-dinitrotoluene	A	712	11:52	BB	0.270	UG/L
27	butachlor	A	1222	20:22	BB	0.001	UG/L

Comments for Sample AD26321 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 16:35:23

The breakdown for Endrin in the Breakdown Standard was 56%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.88 ug/L in the Laboratory Blank.