


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/20/2001 Time: 09:24:03

Sample: AD26337
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/6/01 09:22 Ended: 11/23/01 09:22 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		< MDL	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.0	1.0
20	Surr_Triphenyl phosphate		5.4	1.0
21	Surr_Perylene-d12		3.8	1.0

Chromatogram Plot File: E:\26337D Date: Nov-24-1991 03:47:28
Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26337
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6445245



Integration Report

Quanfile: 26337D

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26337

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	2,076,533	1,176,145
2	phenanthrene d-10 (IS)	16:48	1008	Auto	3,133,283	1,122,022
3	chrysene d-12(IS)	22:35	1355	Auto	2,336,323	970,853
4	p-terphenyl d-14(IS)	20:46	1246	Auto	3,746,531	1,910,938
5	acenaphthene d-10(SURR)	12:12	732	Auto	2,076,533	1,176,145
6	phenanthrene d-10(SURR)	16:48	1008	Auto	3,133,283	1,122,022
7	chrysene d-12 (SURR)	22:35	1355	Auto	2,336,323	970,853
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	1,257,320	668,572
9	triphenyl phosphate (S)	22:01	1321	Auto	1,390,256	699,510
10	perylene d-12 (SURR)	26:47	1607	Auto	1,012,132	179,329
11	bis(2-ethylhexyl)adipate	22:01	1321	Auto	54,990	12,596
12	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	478,740	150,934
13	2,4-dinitrotoluene	13:00	780	Auto	3,210	1,316
14	butachlor	20:20	1220	Auto	3,320	3,320

revised IS/SURR's

Quantitation Report

Quanfile: 26337D

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26337

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	732	12:12	BB	3.656	UG/L
6	phenanthrene d-10(SURR)	A	1008	16:48	BV	3.839	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	3.725	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	5.047	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BB	5.368	UG/L
10	perylene d-12 (SURR)	A	1607	26:47	MM	3.753	UG/L
16	bis(2-ethylhexyl)adipate	A	1321	22:01	MM	0.104	UG/L
17	bis(2-ethylhexyl)phthalate	A	1373	22:53	BV	0.557	UG/L
21	2,4-dinitrotoluene	A	780	13:00	BB	0.024	UG/L
27	butachlor	A	1220	20:20	BB	0.015	UG/L

Comments for Sample AD26337 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 16:39:00

The breakdown for Endrin in the Breakdown Standard was 56%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.88 ug/L in the Laboratory Blank.