


User: Bohik, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 09:36:40

Sample: AD26393
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/6/01 09:34 Ended: 11/23/01 09:34 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		2.2	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.9	1.0
20	Surr_Triphenyl phosphate		5.77	1.0
21	Surr_Perylene-d12		3.8	1.0

Chromatogram Plot File: E:\26393L Date: Nov-24-1991 13:49:06
Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26393
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5014296



Integration Report

Quanfile: 26393L

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26393

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	1,811,522	862,295
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,822,764	1,144,195
3	chrysene d-12(IS)	22:35	1355	Auto	1,899,503	717,471
4	p-terphenyl d-14(IS)	20:47	1247	Auto	2,425,935	1,108,392
5	acenaphthene d-10(SURR)	12:12	732	Auto	1,811,522	862,295
6	phenanthrene d-10(SURR)	16:48	1008	Auto	2,822,764	1,144,195
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,899,503	717,471
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	1,059,736	642,166
9	triphenyl phosphate (S)	22:01	1321	Auto	1,210,600	668,478
10	perylene d-12 (SURR)	26:47	1607	Auto	832,126	149,655
11	hexachlorocyclopentadiene	10:04	604	Auto	1,182	1,182
12	bis(2-ethylhexyl)adipate	21:54	1314	Auto	38,869	7,164
13	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	1,469,379	584,180
14	2,6-dinitrotoluene	11:53	713	Auto	124,152	69,825
15	2,4-dinitrotoluene	13:01	781	Auto	20,885	4,947
16	butachlor	20:22	1222	Auto	0	0

reviewed IS/ surr's

Quantitation Report

Quanfile: 26393L

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26393

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	732	12:12	BV	4.925	UG/L
6	phenanthrene d-10(SURR	A	1008	16:48	BV	5.342	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BB	4.677	UG/L
8	1,3-dimethyl-2-nitrobe	A	474	7:54	BV	4.876	UG/L
9	triphenyl phosphate (S	A	1321	22:01	BV	5.749	UG/L
10	perylene d-12 (SURR)	A	1607	26:47	BB	3.795	UG/L
11	hexachlorocyclopentadi	A	604	10:04	BB	0.015	UG/L
16	bis(2-ethylhexyl)adipa	A	1314	21:54	MM	0.091	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BB	2.251	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	1.169	UG/L
21	2,4-dinitrotoluene	A	781	13:01	BB	0.174	UG/L
27	butachlor	A	1222	20:22	BB	0.001	UG/L

Comments for Sample AD26393 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 16:47:19

The breakdown for Endrin in the Breakdown Standard was 56%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.88 ug/L in the Laboratory Blank.