


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 09:39:29

Sample: AD26433
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/6/01 09:38 Ended: 11/23/01 09:38 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		1.3	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.7	1.0
20	Surr_Triphenyl phosphate		5.2	1.0
21	Surr_Perylene-d12		3.6	1.0

Chromatogram Plot File: E:\26433N Date: Nov-24-1991 14:57:58
 Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26433
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 5310254



Integration Report

Quanfile: 26433N

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26433

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:12	732	Auto	1,848,387	1,241,911
2	phenanthrene d-10 (IS)	16:47	1007	Auto	2,765,571	1,164,412
3	chrysene d-12 (IS)	22:35	1355	Auto	1,986,582	952,008
4	p-terphenyl d-14 (IS)	20:47	1247	Auto	14,146	6,229
5	acenaphthene d-10 (SURR)	12:12	732	Auto	1,848,387	1,241,911
6	phenanthrene d-10 (SURR)	16:47	1007	Auto	2,765,571	1,164,412
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,986,582	952,008
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	1,043,575	461,111
9	triphenyl phosphate (S)	22:01	1321	Auto	1,141,992	409,939
10	perylene d-12 (SURR)	26:47	1607	Auto	825,750	159,612
11	bis(2-ethylhexyl)adipate	21:54	1314	Auto	49,546	13,397
12	bis(2-ethylhexyl)phthalate	22:52	1372	Auto	915,665	306,224
13	butachlor	20:18	1218	Auto	937	937

reviewed IS/surr's

Quantitation Report

Quanfile: 26433N

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/23/01; SAMPLE 26433

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1007	16:47	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	732	12:12	BB	861.706	UG/L
6	phenanthrene d-10(SURR)	A	1007	16:47	BV	897.386	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BB	838.744	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	4.706	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BB	5.185	UG/L
10	perylene d-12 (SURR)	A	1607	26:47	MM	3.601	UG/L
16	bis(2-ethylhexyl)adipate	A	1314	21:54	MM	0.111	UG/L
17	bis(2-ethylhexyl)phthalate	A	1372	22:52	MM	1.292	UG/L
27	butachlor	A	1218	20:18	BB	0.005	UG/L

Comments for Sample AD26433 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 16:49:39

The breakdown for Endrin in the Breakdown Standard was 56%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.88 ug/L in the Laboratory Blank.