


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 15:47:32

Sample: AD26434
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/7/01 15:45 Ended: 11/24/01 15:45 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		1.1	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.6	1.0
20	Surr_Triphenyl phosphate		5.6	1.0
21	Surr_Perylene-d12		4.9	1.0

Chromatogram Plot File: E:\26434A Date: Nov-24-1991 21:11:06
Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26434
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5931621



Integration Report Quanfile: 26434A Quan Entries: 17
 Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26434
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	1,338,741	859,189
2	phenanthrene d-10 (IS)	16:47	1007	Auto	1,932,677	694,895
3	chrysene d-12(IS)	22:35	1355	Auto	1,522,917	583,019
4	p-terphenyl d-14(IS)	20:46	1246	Auto	2,963,951	1,874,688
5	acenaphthene d-10(SURR)	12:12	732	Auto	1,338,741	859,189
6	phenanthrene d-10(SURR)	16:47	1007	Auto	1,932,677	694,895
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,522,917	583,019
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	741,330	377,108
9	triphenyl phosphate (S)	22:01	1321	Auto	939,924	456,093
10	perylene d-12 (SURR)	26:48	1608	Auto	864,932	169,860
11	hexachlorobenzene	15:12	912	Auto	5,801	2,243
12	bis(2-ethylhexyl)adipate	22:01	1321	Auto	19,352	4,521
13	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	600,063	201,726
14	2,6-dinitrotoluene	11:52	712	Auto	50,308	30,008
15	2,4-dinitrotoluene	13:00	780	Auto	4,385	1,920
16	butachlor	20:19	1219	Auto	1,395	1,395
17	4,4'-dichlorodiphenyl ether	20:42	1242	Auto	3,949	3,099

reversed IS/surrs

Quantitation Report

Quanfile: 26434A

Quan Entries: 17

Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26434

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1007	16:47	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	732	12:12	BB	2.979	UG/L
6	phenanthrene d-10 (SURR)	A	1007	16:47	BV	2.994	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	3.069	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BB	4.615	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BV	5.567	UG/L
10	perylene d-12 (SURR)	A	1608	26:48	BV	4.920	UG/L
12	hexachlorobenzene	A	912	15:12	BB	0.036	UG/L
16	bis(2-ethylhexyl)adipate	A	1321	22:01	MM	0.056	UG/L
17	bis(2-ethylhexyl)phthalate	A	1373	22:53	BV	1.096	UG/L
20	2,6-dinitrotoluene	A	712	11:52	BB	0.643	UG/L
21	2,4-dinitrotoluene	A	780	13:00	BB	0.050	UG/L
27	butachlor	A	1219	20:19	BB	0.011	UG/L
28	4,4'-dichlorodiphenyl ether	A	1242	20:42	BB	0.025	UG/L

Comments for Sample AD26434 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 11:49:43

The breakdown for Endrin in the Breakdown Standard was 39%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges.