


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 15:50:32

Sample: AD26472
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/7/01 15:49 Ended: 11/24/01 15:49 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		0.80	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.3	1.0
20	Surr_Triphenyl phosphate		5.3	1.0
21	Surr_Perylene-d12		3.7	1.0

Chromatogram Plot File: C:\26472C Date: Nov-24-1991 22:19:54
Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26472
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5693056



Integration Report

Quanfile: 26472C

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26472

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:12	732	Auto	1,334,273	857,772
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,195,063	912,838
3	chrysene d-12 (IS)	22:35	1355	Auto	1,395,426	547,737
4	p-terphenyl d-14 (IS)	20:46	1246	Auto	2,979,782	1,742,863
5	acenaphthene d-10 (SURR)	12:12	732	Auto	1,334,273	857,772
6	phenanthrene d-10 (SURR)	16:48	1008	Auto	2,195,063	912,838
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,395,426	547,737
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	845,898	497,444
9	triphenyl phosphate (S)	22:01	1321	Auto	817,284	404,773
10	perylene d-12 (SURR)	26:48	1608	Auto	599,754	113,225
11	bis(2-ethylhexyl)adipate	21:54	1314	Auto	15,430	3,627
12	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	408,066	139,832
13	2,6-dinitrotoluene	11:52	712	Auto	121,891	54,091
14	2,4-dinitrotoluene	13:00	780	Auto	12,251	3,786
15	butachlor	20:19	1219	Auto	1,212	1,212

reviewed IS/surrs

Quantitation Report

Quanfile: 26472C

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26472

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	732	12:12	BB	2.953	UG/L
6	phenanthrene d-10(SURR	A	1008	16:48	BV	3.382	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BB	2.797	UG/L
8	1,3-dimethyl-2-nitrobe	A	474	7:54	BV	5.284	UG/L
9	triphenyl phosphate (S	A	1321	22:01	BV	5.283	UG/L
10	perylene d-12 (SURR)	A	1608	26:48	BB	3.723	UG/L
16	bis(2-ethylhexyl)adipa	A	1314	21:54	MM	0.049	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BV	0.803	UG/L
20	2,6-dinitrotoluene	A	712	11:52	BB	1.554	UG/L
21	2,4-dinitrotoluene	A	780	13:00	BB	0.138	UG/L
27	butachlor	A	1219	20:19	BB	0.008	UG/L

Comments for Sample AD26472 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 11:48:28

The breakdown for Endrin in the Breakdown Standard was 39%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges.