


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 15:53:08

Sample: AD26474
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/7/01 15:52 Ended: 11/24/01 15:52 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		< MDL	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.0	1.0
20	Surr_Triphenyl phosphate		5.3	1.0
21	Surr_Perylene-d12		3.9	1.0

Chromatogram Plot File: C:\26474E Date: Nov-24-1991 23:28:44
Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26474
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 5788779



Integration Report

Quanfile: 26474E

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26474

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	1,740,864	1,198,387
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,625,831	1,021,951
3	chrysene d-12(IS)	22:35	1355	Auto	1,863,116	791,413
4	p-terphenyl d-14(IS)	20:46	1246	Auto	2,570,666	1,765,429
5	acenaphthene d-10(SURR)	12:12	732	Auto	1,740,864	1,198,387
6	phenanthrene d-10(SURR)	16:48	1008	Auto	2,625,831	1,021,951
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,863,116	791,413
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	1,047,278	504,532
9	triphenyl phosphate (S)	22:01	1321	Auto	1,090,408	534,600
10	perylene d-12 (SURR)	26:47	1607	Auto	832,711	167,002
11	alachlor	18:18	1098	Auto	1,109	1,109
12	bis(2-ethylhexyl)adipate	21:54	1314	Auto	52,732	14,947
13	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	346,992	110,981
14	2,4-dinitrotoluene	13:00	780	Auto	9,907	3,062
15	metolachlor	19:04	1144	Auto	10,902	5,222
16	butachlor	20:18	1218	Auto	6,410	2,140

reversed IS/surr's

Quantitation Report

Quanfile: 26474E

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/24/01; SAMPLE 26474

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	732	12:12	BB	4.466	UG/L
6	phenanthrene d-10(SURR)	A	1008	16:48	BV	4.689	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	4.329	UG/L
8	1,3-dimethyl-2-nitrobe	A	474	7:54	BV	5.014	UG/L
9	triphenyl phosphate (S	A	1321	22:01	BV	5.279	UG/L
10	perylene d-12 (SURR)	A	1607	26:47	BB	3.872	UG/L
15	alachlor	A	1098	18:18	BB	0.007	UG/L
16	bis(2-ethylhexyl)adipa	A	1314	21:54	MM	0.126	UG/L
17	bis(2-ethylhexyl)phtha	A	1373	22:53	BV	0.505	UG/L
21	2,4-dinitrotoluene	A	780	13:00	BB	0.086	UG/L
26	metolachlor	A	1144	19:04	BB	0.021	UG/L
27	butachlor	A	1218	20:18	BB	0.035	UG/L

Comments for Sample AD26474 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 12-22-2001 Time: 11:52:51

The breakdown for Endrin in the Breakdown Standard was 39%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 2 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges.