


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/21/2001 Time: 16:06:13

Sample: AD26536
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/9/01 16:01 Ended: 11/28/01 16:01 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		< MDL	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.4	1.0
20	Surr_Triphenyl phosphate		5.6	1.0
21	Surr_Perylene-d12		4.6	1.0

Chromatogram Plot
 File: E:\26536A Date: Nov-28-1991 16:17:51
 Comment: BOHLK: METHOD 525; INST 204; 11/28/01; SAMPLE 26536
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 6040792



Integration Report

Quanfile: 26536A

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26536

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:12	732	Auto	1,305,142	678,784
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,153,792	889,027
3	chrysene d-12(IS)	22:35	1355	Auto	1,772,175	651,585
4	p-terphenyl d-14(IS)	20:46	1246	Auto	3,463,082	1,855,421
5	acenaphthene d-10(SURR)	12:12	732	Auto	1,305,142	678,784
6	phenanthrene d-10(SURR)	16:48	1008	Auto	2,153,792	889,027
7	chrysene d-12 (SURR)	22:35	1355	Auto	1,772,175	651,585
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	837,891	432,739
9	triphenyl phosphate (S)	22:01	1321	Auto	1,093,027	523,397
10	perylene d-12 (SURR)	26:48	1608	Auto	940,646	197,708
11	hexachlorocyclopentadiene	10:04	604	Auto	3,527	2,628
12	simazine	15:59	959	Auto	2,083	1,224
13	bis(2-ethylhexyl)adipate	22:01	1321	Auto	32,378	10,008
14	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	339,061	112,017
15	2,4-dinitrotoluene	12:50	770	Auto	8,086	1,852
16	butachlor	20:19	1219	Auto	2,331	1,250

Quantitation Report Quanfile: 26536A Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 11/28/01; SAMPLE 26536
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	732	12:12	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1355	22:35	BV	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1246	20:46	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	732	12:12	BB	2.486	UG/L
6	phenanthrene d-10 (SURR)	A	1008	16:48	BV	2.855	UG/L
7	chrysene d-12 (SURR)	A	1355	22:35	BV	3.057	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	5.351	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BB	5.564	UG/L
10	perylene d-12 (SURR)	A	1608	26:48	BV	4.598	UG/L
11	hexachlorocyclopentadiene	A	604	10:04	BB	0.062	UG/L
13	simazine	A	959	15:59	BB	0.032	UG/L
16	bis(2-ethylhexyl)adipate	A	1321	22:01	MM	0.081	UG/L
17	bis(2-ethylhexyl)phthalate	A	1373	22:53	BB	0.519	UG/L
21	2,4-dinitrotoluene	A	770	12:50	BB	0.094	UG/L
27	butachlor	A	1219	20:19	BB	0.016	UG/L

Comments for Sample AD26536 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 12:50:17

The breakdown for Endrin in the Breakdown Standard was 40%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the LCS Standard were below their acceptance ranges.