


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/20/2001 Time: 16:00:41

Sample: AD26470
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 11/9/01 15:58 Ended: 11/29/01 15:58 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	MDL
1	Hexachlorocyclopentadiene		< MDL	0.10
2	Hexachlorobenzene		< MDL	0.10
3	Simazine		< MDL	0.40
4	Atrazine		< MDL	0.20
5	Alachlor		< MDL	0.20
6	bis(2-Ethylhexyl)adipate		< MDL	0.60
7	bis(2-Ethylhexyl)phthalate		0.61	0.60
8	Benzo(a)pyrene		< MDL	0.20
9	EPTC		< MDL	1.0
10	2,6-Dinitrotoluene		< MDL	2.0
11	2,4-Dinitrotoluene		< MDL	2.0
12	Molinate		< MDL	0.90
13	Terbacil		< MDL	2.0
14	Acetochlor		< MDL	2.0
15	Metribuzin		< MDL	0.15
16	Metolachlor		< MDL	0.75
17	Butachlor		< MDL	0.40
18	4,4-DDE		< MDL	0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.3	1.0
20	Surr_Triphenyl phosphate		5.4	1.0
21	Surr_Perylene-d12		5.6	1.0

Chromatogram Plot File: E:\26470A Date: Nov-29-1991 20:05:26
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26470
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 6545977



Integration Report Quanfile: 26470A Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26470
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:13	733	Auto	1,436,838	776,098
2	phenanthrene d-10 (IS)	16:48	1008	Auto	2,340,358	816,526
3	chrysene d-12 (IS)	22:36	1356	Auto	1,848,836	774,982
4	p-terphenyl d-14 (IS)	20:47	1247	Auto	3,140,047	1,850,735
5	acenaphthene d-10 (SURR)	12:13	733	Auto	1,436,838	776,098
6	phenanthrene d-10 (SURR)	16:48	1008	Auto	2,340,358	816,526
7	chrysene d-12 (SURR)	22:36	1356	Auto	1,848,836	774,982
8	1,3-dimethyl-2-nitrobenzene	7:54	474	Auto	907,198	504,308
9	triphenyl phosphate (S)	22:01	1321	Auto	1,107,642	498,731
10	perylene d-12 (SURR)	26:49	1609	Auto	1,195,325	245,644
11	bis(2-ethylhexyl)adipate	21:54	1314	Auto	26,135	7,616
12	bis(2-ethylhexyl)phthalate	22:53	1373	Auto	416,733	150,548
13	2,6-dinitrotoluene	11:53	713	Auto	118,904	66,728
14	2,4-dinitrotoluene	13:01	781	Auto	10,937	4,222
15	butachlor	20:21	1221	Auto	1,188	1,188

reversed IS/ SURR's

Quantitation Report

Quanfile: 26470A

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 11/29/01; SAMPLE 26470

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	733	12:13	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1008	16:48	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1356	22:36	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1247	20:47	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	733	12:13	BV	3.018	UG/L
6	phenanthrene d-10 (SURR)	A	1008	16:48	BV	3.422	UG/L
7	chrysene d-12 (SURR)	A	1356	22:36	BB	3.517	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	474	7:54	BV	5.262	UG/L
9	triphenyl phosphate (S)	A	1321	22:01	BV	5.404	UG/L
10	perylene d-12 (SURR)	A	1609	26:49	BV	5.600	UG/L
16	bis(2-ethylhexyl)adipate	A	1314	21:54	MM	0.063	UG/L
17	bis(2-ethylhexyl)phthalate	A	1373	22:53	VV	0.614	UG/L
20	2,6-dinitrotoluene	A	713	11:53	BB	1.409	UG/L
21	2,4-dinitrotoluene	A	781	13:01	BB	0.115	UG/L
27	butachlor	A	1221	20:21	BB	0.008	UG/L

• **Comments for Sample AD26470 - Analysis \$525**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-22-2001 Time: 12:41:43

The breakdown for Endrin in the Breakdown Standard was 50%. Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were below their acceptance ranges. Recoveries for 5 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.64 ug/L.