


User: Fakhouri, Morgan
Westchester Dept. of Labs & Research
Date: 12/18/2001 Time: 15:50:37

Sample: AD26251
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 12/18/01 15:48 Ended: 12/18/01 15:48 Analyst: MF
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.84		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.6		1.0
20	Surr_Triphenyl phosphate		4.8		1.0
21	Surr_Perylene-d12		3.6		1.0

Chromatogram Plot
 File: E:\26251L Date: Nov-21-1991 15:07:15
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26251
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 3962567
 100%



Integration Report

Quanfile: 26251L

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26251

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:19	739	Auto	1,345,450	621,058
2	phenanthrene d-10 (IS)	16:57	1017	Auto	1,944,411	755,673
3	chrysene d-12 (IS)	22:42	1362	Auto	1,312,389	529,606
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	1,912,049	960,581
5	acenaphthene d-10 (SURR)	12:19	739	Auto	1,345,450	621,058
6	phenanthrene d-10 (SURR)	16:57	1017	Auto	1,944,411	755,673
7	chrysene d-12 (SURR)	22:42	1362	Auto	1,312,389	529,606
8	1,3-dimethyl-2-nitrobenzene	7:59	479	Auto	764,749	379,066
9	triphenyl phosphate (S)	22:07	1327	Auto	910,022	412,199
10	perylene d-12 (SURR)	27:02	1622	Auto	565,200	95,605
11	simazine	16:13	973	Auto	0	0
12	bis(2-ethylhexyl)adipate	21:59	1319	Auto	35,538	3,544
13	bis(2-ethylhexyl)phthalate	23:00	1380	Auto	424,388	125,771
14	2,6-dinitrotoluene	11:58	718	Auto	56,724	24,547

reversed IS/sur's

Quantitation Report

Quanfile: 26251L

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26251

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	739	12:19	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	739	12:19	BV	4.630	UG/L
6	phenanthrene d-10(SURR)	A	1017	16:57	BV	4.364	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	4.416	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	479	7:59	BB	4.563	UG/L
9	triphenyl phosphate (S)	A	1327	22:07	BV	4.837	UG/L
10	perylene d-12 (SURR)	A	1622	27:02	BB	3.568	UG/L
13	simazine	A	973	16:13	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipate	A	1319	21:59	MM	0.225	UG/L
17	bis(2-ethylhexyl)phthalate	A	1380	23:00	BV	0.840	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.654	UG/L

Comments for Sample AD26251 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 12-19-2001 Time: 10:52:42

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 6 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the LPC Standard was 53%. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.74 ug/L.