

ser: Fakhouri, Morgan
 /estchester Dept. of Labs & Research
 ate: 12/18/2001 Time: 15:35:23

ample: AD26128
 alysis: \$525 Organic Chemicals - 525.2
 its: ug
 arted: 12/18/01 15:34 Ended: 12/18/01 15:34 Analyst: MF
 sult source: manual entry
 ge: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		2.5		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimeyl-2-nitrobenzen		4.9		1.0
20	Surr_Triphenyhosphate		4.7		1.0
21	Surr_Perylene12		3.5		1.0

Quantitation Report

Quanfile: 26128A

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26128

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	738	12:18	BV	2.848	UG/L
6	phenanthrene d-10(SURR)	A	1017	16:57	BV	2.776	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	2.565	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	479	7:59	BV	4.896	UG/L
9	triphenyl phosphate (S)	A	1326	22:06	BV	4.749	UG/L
10	perylene d-12 (SURR)	A	1621	27:01	BB	3.456	UG/L
16	bis(2-ethylhexyl)adipate	A	1319	21:59	MM	0.403	UG/L
17	bis(2-ethylhexyl)phthalate	A	1380	23:00	MM	2.476	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.475	UG/L
21	2,4-dinitrotoluene	A	775	12:55	BB	0.021	UG/L

Comments for Sample AD26128 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-19-2001 Time: 10:56:52

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 6 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the LPC Standard was 53%. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.74 ug/L.