


User: Fakhouri, Morgan
Westchester Dept. of Labs & Research
Date: 12/18/2001 Time: 15:44:18

Sample: AD26163
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 12/18/01 15:43 Ended: 12/18/01 15:43 Analyst: MF
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.81		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.6		1.0
20	Surr_Triphenyl phosphate		4.6		1.0
21	Surr_Perylene-d12		3.9		1.0

Chromatogram Plot File: E:\26163G Date: Nov-21-1991 05:44:12
Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26163
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 6083251



Integration Report Quanfile: 26163G Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26163
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	1,324,707	777,004
2	phenanthrene d-10 (IS)	16:57	1017	Auto	2,015,399	771,772
3	chrysene d-12 (IS)	22:42	1362	Auto	1,275,697	481,268
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	2,564,229	1,817,742
5	acenaphthene d-10 (SURR)	12:18	738	Auto	1,324,707	777,004
6	phenanthrene d-10 (SURR)	16:57	1017	Auto	2,015,399	771,772
7	chrysene d-12 (SURR)	22:42	1362	Auto	1,275,697	481,268
8	1,3-dimethyl-2-nitrobenzene	7:59	479	Auto	753,976	398,733
9	triphenyl phosphate (S)	22:06	1326	Auto	849,819	368,297
10	perylene d-12 (SURR)	27:01	1621	Auto	603,209	104,367
11	bis(2-ethylhexyl)adipate	21:59	1319	Auto	31,265	9,620
12	bis(2-ethylhexyl)phthalate	22:59	1379	Auto	397,808	104,649
13	2,6-dinitrotoluene	11:58	718	Auto	36,344	23,149
14	2,4-dinitrotoluene	12:56	776	Auto	884	884

reversed IS/surrs

Quantitation Report Quanfile: 26163G Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26163
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BV	5.000	UG/L
5	acenaphthene d-10(SURR	A	738	12:18	BB	3.399	UG/L
6	phenanthrene d-10(SURR	A	1017	16:57	BV	3.373	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	3.201	UG/L
8	1,3-dimethyl-2-nitrobe	A	479	7:59	BV	4.570	UG/L
9	triphenyl phosphate (S	A	1326	22:06	BB	4.647	UG/L
10	perylene d-12 (SURR)	A	1621	27:01	BB	3.918	UG/L
16	bis(2-ethylhexyl)adipa	A	1319	21:59	MM	0.204	UG/L
17	bis(2-ethylhexyl)phtha	A	1379	22:59	BB	0.809	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.423	UG/L
21	2,4-dinitrotoluene	A	776	12:56	BB	0.010	UG/L

Comments for Sample AD26163 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 12-19-2001 Time: 11:04:16

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 6 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the LPC Standard was 53%. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.74 ug/L.