


User: Fakhouri, Morgan
 Westchester Dept. of Labs & Research
 Date: 12/18/2001 Time: 15:55:31

Sample: AD26165
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 12/18/01 15:54 Ended: 12/18/01 15:54 Analyst: MF
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.6		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.5		1.0
20	Surr_Triphenyl phosphate		4.5		1.0
21	Surr_Perylene-d12		3.7		1.0

Chromatogram Plot File: E:\26165I Date: Nov-21-1991 13:23:52
Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26165
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 4984263



Integration Report Quanfile: 26165I Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26165
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:19	739	Auto	1,246,262	742,098
2	phenanthrene d-10 (IS)	16:57	1017	Auto	1,999,859	712,844
3	chrysene d-12(IS)	22:43	1363	Auto	1,143,151	397,048
4	p-terphenyl d-14(IS)	20:52	1252	Auto	1,962,765	1,176,551
5	acenaphthene d-10(SURR)	12:19	739	Auto	1,246,262	742,098
6	phenanthrene d-10(SURR)	16:57	1017	Auto	1,999,859	712,844
7	chrysene d-12 (SURR)	22:43	1363	Auto	1,143,151	397,048
8	1,3-dimethyl-2-nitrobenzene	7:59	479	Auto	700,625	302,863
9	triphenyl phosphate (S)	22:07	1327	Auto	737,432	341,365
10	perylene d-12 (SURR)	27:02	1622	Auto	515,742	84,121
11	simazine	16:10	970	Auto	866	866
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	33,571	4,830
13	bis(2-ethylhexyl)phthalate	23:00	1380	Auto	689,422	259,527
14	2,6-dinitrotoluene	11:59	719	Auto	21,737	12,020
15	butachlor	20:24	1224	Auto	2,020	1,133

reviewed IS/ SURR'S

Quantitation Report Quanfile: 26165I Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26165
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	739	12:19	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1363	22:43	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1252	20:52	BV	5.000	UG/L
5	acenaphthene d-10(SURR	A	739	12:19	BV	4.178	UG/L
6	phenanthrene d-10(SURR	A	1017	16:57	BV	4.372	UG/L
7	chrysene d-12 (SURR)	A	1363	22:43	BV	3.747	UG/L
8	1,3-dimethyl-2-nitrobe	A	479	7:59	BV	4.514	UG/L
9	triphenyl phosphate (S	A	1327	22:07	BB	4.500	UG/L
10	perylene d-12 (SURR)	A	1622	27:02	BB	3.738	UG/L
13	simazine	A	970	16:10	BB	0.016	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	BV	0.244	UG/L
17	bis(2-ethylhexyl)phtha	A	1380	23:00	BB	1.617	UG/L
20	2,6-dinitrotoluene	A	719	11:59	BB	0.268	UG/L
27	butachlor	A	1224	20:24	BB	0.017	UG/L

Comments for Sample AD26165 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-19-2001 Time: 11:06:37

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 6 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the LPC Standard was 53%. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.74 ug/L.