


User: Fakhouri, Morgan
Westchester Dept. of Labs & Research
Date: 12/18/2001 Time: 15:56:49

Sample: AD26202
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 12/18/01 15:55 Ended: 12/18/01 15:55 Analyst: MF
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.93		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.5		1.0
20	Surr_Triphenyl phosphate		4.3		1.0
21	Surr_Perylene-d12		3.3		1.0

Chromatogram Plot File: E:\26202J Date: Nov-21-1991 13:58:21
Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26202
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 4244228



Integration Report Quanfile: 26202J Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26202
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:18	738	Auto	1,445,046	659,189
2	phenanthrene d-10 (IS)	16:57	1017	Auto	2,081,542	885,877
3	chrysene d-12 (IS)	22:42	1362	Auto	1,534,170	619,101
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	1,984,107	1,284,482
5	acenaphthene d-10 (SURR)	12:18	738	Auto	1,445,046	659,189
6	phenanthrene d-10 (SURR)	16:57	1017	Auto	2,081,542	885,877
7	chrysene d-12 (SURR)	22:42	1362	Auto	1,534,170	619,101
8	1,3-dimethyl-2-nitrobenzene	7:59	479	Auto	817,811	443,421
9	triphenyl phosphate (S)	22:07	1327	Auto	945,260	345,850
10	perylene d-12 (SURR)	27:02	1622	Auto	611,909	103,048
11	simazine	16:12	972	Auto	0	0
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	33,235	3,303
13	bis(2-ethylhexyl)phthalate	23:00	1380	Auto	544,027	154,591
14	2,6-dinitrotoluene	11:58	718	Auto	79,284	36,564
15	2,4-dinitrotoluene	13:07	787	Auto	3,180	1,341

renewed IS/ Surrs

Quantitation Report Quanfile: 26202J Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 11/20/01; SAMPLE 26202
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	738	12:18	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	738	12:18	BB	4.792	UG/L
6	phenanthrene d-10(SURR	A	1017	16:57	BV	4.502	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	4.975	UG/L
8	1,3-dimethyl-2-nitrobe	A	479	7:59	BV	4.544	UG/L
9	triphenyl phosphate (S	A	1327	22:07	BB	4.298	UG/L
10	perylene d-12 (SURR)	A	1622	27:02	BB	3.305	UG/L
13	simazine	A	972	16:12	BB	0.001	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.180	UG/L
17	bis(2-ethylhexyl)phtha	A	1380	23:00	BV	0.925	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.856	UG/L
21	2,4-dinitrotoluene	A	787	13:07	BB	0.031	UG/L

Comments for Sample AD26202 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-19-2001 Time: 11:08:47

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges. Recoveries for 6 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges. The breakdown for Endrin in the LPC Standard was 53%. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.74 ug/L. The recovery for 1 of the 3 Surrogate Standards in this sample was below its acceptance range.