


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 12/18/2001 Time: 18:55:47

Sample: AD26204
Analysis: SLP525 Organic Extractables
Units: ug
Started: 11/21/01 18:53 Ended: 11/21/01 18:53 Analyst: TB
Page: 1


	Component Name	Viol	Result
1	DFTPP		PASSED
2	ENDRIN *		53
3	DDT		8.4

blog

Chromatogram Plot File: E:\DFTA1121 Date: Nov-21-1991 19:09:19
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3653593



Chromatogram Plot File: E:\DFTA1121 Date: Nov-21-1991 19:09:19
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN
Scan No: 1080 Retention Time: 18:00 RIC: 23004 Mass Range: 47 - 86
Plotted: 1080 to 1120 Range: 1 to 1860 100% = 3653593
9.25%



Mass List

File: E:\DFTA1121 Date: Nov-21-1991 19:09:19

Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 3653593 Mass Range: 45 - 444

Peaks: 199 Mass of Base Peak: 442 Int: 579868 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
45	1,184	0.20	79	18,947	3.27	111	17,500	3.02
47	2,631	0.45	80	8,684	1.50	112	2,368	0.41
48	1,710	0.29	81	10,921	1.88	116	3,421	0.59
49	9,078	1.57	82	1,973	0.34	117	70,921	12.23
50	79,473	13.71	83	2,236	0.39	118	3,815	0.66
51	227,631	39.26	84	6,052	1.04	120	1,052	0.18
52	7,894	1.36	85	2,368	0.41	122	4,605	0.79
53	921	0.16	86	7,631	1.32	123	6,578	1.13
55	4,210	0.73	87	2,236	0.39	124	2,368	0.41
56	5,131	0.88	91	4,605	0.79	125	1,447	0.25
57	8,421	1.45	92	2,894	0.50	126	1,578	0.27
61	3,947	0.68	93	22,894	3.95	127	126,973	21.90
62	4,342	0.75	94	1,578	0.27	128	7,236	1.25
63	6,842	1.18	97	1,315	0.23	129	54,342	9.37
65	3,289	0.57	98	21,447	3.70	130	1,973	0.34
67	1,447	0.25	99	9,605	1.66	131	1,052	0.18
68	3,421	0.59	100	1,052	0.18	132	1,052	0.18
69	147,105	25.37	101	7,105	1.23	134	2,105	0.36
70	1,710	0.29	103	4,473	0.77	135	3,684	0.64
73	4,078	0.70	104	4,736	0.82	136	3,684	0.64
74	20,394	3.52	105	3,157	0.54	137	2,368	0.41
75	26,973	4.65	107	56,052	9.67	141	12,105	2.09
76	6,052	1.04	108	4,605	0.79	142	2,894	0.50
77	162,631	28.05	109	1,842	0.32	143	2,763	0.48
78	9,605	1.66	110	103,421	17.84	146	1,578	0.27

Mass List

File: E:\DFTA1121 Date: Nov-21-1991 19:09:19

Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 3653593 Mass Range: 45 - 444

Peaks: 199 Mass of Base Peak: 442 Int: 579868 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
147	6,315	1.09	185	3,421	0.59	223	6,315	1.09
148	8,815	1.52	186	43,552	7.51	224	35,131	6.06
149	2,236	0.39	187	14,078	2.43	225	12,894	2.22
151	1,184	0.20	189	1,842	0.32	227	11,184	1.93
153	1,447	0.25	192	1,973	0.34	228	2,763	0.48
154	1,973	0.34	193	3,289	0.57	229	4,078	0.70
155	4,736	0.82	196	3,947	0.68	231	1,710	0.29
156	4,868	0.84	197	2,105	0.36	235	2,763	0.48
157	1,315	0.23	198	337,894	58.27	236	1,973	0.34
158	1,710	0.29	199	20,394	3.52	238	921	0.16
160	1,842	0.32	200	3,421	0.59	239	1,447	0.25
161	2,500	0.43	203	1,842	0.32	241	921	0.16
162	1,184	0.20	204	10,526	1.82	242	1,447	0.25
165	3,289	0.57	205	17,500	3.02	243	2,500	0.43
166	2,236	0.39	206	61,842	10.66	244	19,605	3.38
167	17,368	3.00	207	7,236	1.25	245	3,289	0.57
168	6,710	1.16	208	1,184	0.20	246	6,973	1.20
174	3,026	0.52	210	1,842	0.32	247	1,710	0.29
175	5,921	1.02	211	3,289	0.57	249	1,710	0.29
176	1,578	0.27	215	1,710	0.29	250	921	0.16
177	5,000	0.86	216	1,842	0.32	255	222,763	38.42
178	1,184	0.20	217	15,526	2.68	256	33,421	5.76
179	9,868	1.70	218	1,710	0.29	257	2,763	0.48
180	9,078	1.57	221	5,263	0.91	258	15,789	2.72
181	4,078	0.70	222	1,842	0.32	265	3,157	0.54

Mass List

File: E:\DFTA1121 Date: Nov-21-1991 19:09:19


Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN

Scan No: 1101 Retention Time: 18:21 RIC: 3653593 Mass Range: 45 - 444


Peaks: 199 Mass of Base Peak: 442 Int: 579868 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
222	1,842	0.32	265	3,157	0.54	334	10,921	1.88
223	6,315	1.09	273	5,000	0.86	335	3,947	0.68
224	35,131	6.06	274	26,447	4.56	342	1,710	0.29
225	12,894	2.22	275	114,736	19.79	346	2,894	0.50
227	11,184	1.93	276	12,631	2.18	352	3,157	0.54
228	2,763	0.48	277	14,078	2.43	353	3,289	0.57
229	4,078	0.70	278	2,236	0.39	354	6,447	1.11
231	1,710	0.29	285	1,973	0.34	355	3,157	0.54
235	2,763	0.48	293	3,684	0.64	365	21,710	3.74
236	1,973	0.34	294	921	0.16	366	3,421	0.59
238	921	0.16	296	26,842	4.63	372	6,578	1.13
239	1,447	0.25	297	5,000	0.86	373	2,236	0.39
241	921	0.16	302	1,052	0.18	383	2,236	0.39
242	1,447	0.25	303	2,368	0.41	391	3,421	0.59
243	2,500	0.43	304	2,368	0.41	402	1,447	0.25
244	19,605	3.38	314	2,105	0.36	403	5,263	0.91
245	3,289	0.57	315	5,394	0.93	404	2,236	0.39
246	6,973	1.20	316	1,842	0.32	421	2,763	0.48
247	1,710	0.29	321	2,368	0.41	422	5,526	0.95
249	1,710	0.29	323	12,763	2.20	423	33,289	5.74
250	921	0.16	324	1,710	0.29	424	4,736	0.82
255	222,763	38.42	327	1,315	0.23	441	95,394	16.45
256	33,421	5.76	328	1,710	0.29	442	579,868	100.00
257	2,763	0.48	332	1,447	0.25	443	108,289	18.67
258	15,789	2.72	333	1,710	0.29	444	9,868	1.70


Chromatogram Plot File: E:\DFTA1121 Date: Nov-21-1991 19:09:19
 Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN
 Scan No: 1280 Retention Time: 21:20 RIC: 23589 Mass Range: 47 - 207
 Plotted: 1250 to 1280 Range: 1 to 1860 100% = 3653593




Chromatogram Plot File: E:\DFTA1121 Date: Nov-21-1991 19:09:19
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN
Scan No: 1302 Retention Time: 21:42 RIC: 23573 Mass Range: 47 - 207
Plotted: 1282 to 1302 Range: 1 to 1860 100% = 3653593




Chromatogram Plot File: E:\DFTA1121 Date: Nov-21-1991 19:09:19
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DFTPP/ENDRIN
Scan No: 1367 Retention Time: 22:47 RIC: 21479 Mass Range: 47 - 207
Plotted: 1347 to 1367 Range: 1 to 1860 100% = 3653593




Chromatogram Plot File: E:\DDTA1121 Date: Nov-21-1991 19:47:38
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DDT
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3684585



Chromatogram Plot File: E:\DDTA1121 Date: Nov-21-1991 19:47:38
 Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DDT
 Scan No: 1259 Retention Time: 20:59 RIC: 19677 Mass Range: 47 - 207
 Plotted: 1239 to 1259 Range: 1 to 1860 100% = 3684585
 6.94%



Chromatogram Plot File: E:\DDTA1121 Date: Nov-21-1991 19:47:38
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DDT
Scan No: 1292 Retention Time: 21:32 RIC: 23264 Mass Range: 47 - 237
Plotted: 1272 to 1292 Range: 1 to 1860 100% = 3684585



Chromatogram Plot File: E:\DDTA1121 Date: Nov-21-1991 19:47:38
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; DDT
Scan No: 1324 Retention Time: 22:04 RIC: 59694 Mass Range: 47 - 429
Plotted: 1304 to 1324 Range: 1 to 1860 100% = 3684585

