


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/19/2001 Time: 14:46:29

Sample: AD26206
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/5/01 14:45 Ended: 11/21/01 14:45 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.65		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.6		1.0
20	Surr_Triphenyl phosphate		4.9		1.0
21	Surr_Perylene-d12		3.6		1.0

Chromatogram Plot File: E:\26206C Date: Nov-21-1991 23:14:14
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26206
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 2752027



Integration Report Quanfile: 26206C Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26206
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:19	739	Auto	933,462	418,137
2	phenanthrene d-10 (IS)	16:57	1017	Auto	1,610,935	670,793
3	chrysene d-12 (IS)	22:42	1362	Auto	1,007,484	412,600
4	p-terphenyl d-14 (IS)	20:51	1251	Auto	1,126,383	638,035
5	acenaphthene d-10 (SURR)	12:19	739	Auto	933,462	418,137
6	phenanthrene d-10 (SURR)	16:57	1017	Auto	1,610,935	670,793
7	chrysene d-12 (SURR)	22:42	1362	Auto	1,007,484	412,600
8	1,3-dimethyl-2-nitrobenzene	7:59	479	Auto	532,430	285,945
9	triphenyl phosphate (S)	22:07	1327	Auto	712,562	258,579
10	perylene d-12 (SURR)	27:01	1621	Auto	442,678	67,124
11	hexachlorocyclopentadiene	10:10	610	Auto	1,058	1,058
12	bis(2-ethylhexyl)adipate	22:02	1322	Auto	16,831	2,328
13	bis(2-ethylhexyl)phthalate	23:00	1380	Auto	253,063	68,541
14	2,6-dinitrotoluene	11:58	718	Auto	36,285	18,897

reviewed IS/ surr's

Quantitation Report

Quanfile: 26206C

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26206

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	739	12:19	BV	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1251	20:51	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	739	12:19	BV	5.453	UG/L
6	phenanthrene d-10(SURR	A	1017	16:57	BV	6.137	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	5.754	UG/L
8	1,3-dimethyl-2-nitrobe	A	479	7:59	BB	4.579	UG/L
9	triphenyl phosphate (S	A	1327	22:07	BV	4.934	UG/L
10	perylene d-12 (SURR)	A	1621	27:01	BB	3.641	UG/L
11	hexachlorocyclopentadi	A	610	10:10	BB	0.024	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.139	UG/L
17	bis(2-ethylhexyl)phtha	A	1380	23:00	BV	0.647	UG/L
20	2,6-dinitrotoluene	A	718	11:58	BB	0.602	UG/L

Comments for Sample AD26206 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 13:56:27

The breakdown for Endrin in the Breakdown Standard is 53%. Recoveries for 6 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.73 ug/L in the Laboratory Blank.