


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 12/19/2001 Time: 14:57:57

Sample: AD26315
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 11/5/01 14:49 Ended: 11/21/01 14:49 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.20
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.9		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.7		1.0
20	Surr_Triphenyl phosphate		4.5		1.0
21	Surr_Perylene-d12		3.6		1.0

Chromatogram Plot File: E:\26315F Date: Nov-22-1991 00:57:30
Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26315
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3605118



Integration Report Quanfile: 26315F Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26315
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:19	739	Auto	1,181,517	652,238
2	phenanthrene d-10 (IS)	16:57	1017	Auto	1,803,008	635,107
3	chrysene d-12(IS)	22:42	1362	Auto	1,298,818	513,291
4	p-terphenyl d-14(IS)	20:52	1252	Auto	1,802,829	817,597
5	acenaphthene d-10(SURR)	12:19	739	Auto	1,181,517	652,238
6	phenanthrene d-10(SURR)	16:57	1017	Auto	1,803,008	635,107
7	chrysene d-12 (SURR)	22:42	1362	Auto	1,298,818	513,291
8	1,3-dimethyl-2-nitrobe	7:59	479	Auto	695,516	324,387
9	triphenyl phosphate (S	22:07	1327	Auto	836,113	389,912
10	perylene d-12 (SURR)	27:02	1622	Auto	562,821	94,548
11	bis(2-ethylhexyl)adipa	22:02	1322	Auto	23,532	3,023
12	bis(2-ethylhexyl)phtha	23:00	1380	Auto	919,776	315,595
13	2,6-dinitrotoluene	11:59	719	Auto	32,530	16,141

reviewed IS/suror's

Quantitation Report

Quanfile: 26315F

Quan Entries: 13

Comment: BOHLK; METHOD 525; INST 204; 11/21/01; SAMPLE 26315

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	739	12:19	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1017	16:57	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1362	22:42	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1252	20:52	BB	5.000	UG/L
5	acenaphthene d-10 (SURR	A	739	12:19	BB	4.312	UG/L
6	phenanthrene d-10 (SURR	A	1017	16:57	BV	4.292	UG/L
7	chrysene d-12 (SURR)	A	1362	22:42	BB	4.635	UG/L
8	1,3-dimethyl-2-nitrobe	A	479	7:59	BB	4.726	UG/L
9	triphenyl phosphate (S	A	1327	22:07	BB	4.491	UG/L
10	perylene d-12 (SURR)	A	1622	27:02	BB	3.591	UG/L
16	bis(2-ethylhexyl)adipa	A	1322	22:02	MM	0.151	UG/L
17	bis(2-ethylhexyl)phtha	A	1380	23:00	BB	1.920	UG/L
20	2,6-dinitrotoluene	A	719	11:59	BB	0.424	UG/L

Comments for Sample AD26315 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-21-2001 Time: 14:00:18

The breakdown for Endrin in the Breakdown Standard is 53%. Recoveries for 6 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 4 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate was 0.73 ug/L in the Laboratory Blank.