

User: Fakhouri, Morgan  
 Westchester Dept. of Labs & Research  
 Date: 01/16/2002 Time: 15:57:07

Sample: AD28319  
 Analysis: \$LP525 Organic Extractables  
 Units: ug  
 Started: 1/16/02 15:46 Ended: 1/16/02 15:46 Analyst: MF  
 Page: 1

|   | Component Name | Viol. | Result |
|---|----------------|-------|--------|
| 1 | DFTPP          |       | PASSED |
| 2 | ENDRIN         |       | 13     |
| 3 | DDT            |       | 18     |

## DFTPP

Data File : C:\MSDCHEM\1\DATA\121701\DFTPP2.D

Acq On : 18 Dec 2001 2:02 am

Sample :

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\011402.M (RTE Integrator)

Title : Quantitation For Method 525.2

## Spectrum Information: Scan 1408

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 26.6         | 285760     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.5          | 5185       | PASS                |
| 69             | 198             | 0.00            | 100             | 31.6         | 340416     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.7          | 2427       | PASS                |
| 127            | 198             | 10              | 80              | 50.8         | 546880     | PASS                |
| 197            | 198             | 0.00            | 1               | 0.0          | 0          | PASS                |
| 198            | 198             | 50              | 100             | 100.0        | 1076224    | PASS                |
| 199            | 198             | 5               | 9               | 6.1          | 65336      | PASS                |
| 275            | 198             | 10              | 60              | 17.4         | 187072     | PASS                |
| 365            | 198             | 1               | 100             | 1.5          | 16270      | PASS                |
| 441            | 443             | 0.01            | 100             | 68.7         | 95984      | PASS                |
| 442            | 198             | 50              | 100             | 67.5         | 726592     | PASS                |
| 443            | 442             | 15              | 24              | 19.2         | 139648     | PASS                |

DFTPP2.D 011402.M

Tue Jan 15 12:35:27 2002

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\121701\DFTPP2.D  
 Acq On : 18 Dec 2001 2:02 am  
 Sample :  
 Misc :  
 MS Integration Params: BENZO.P

Vial: 1  
 Operator: McLean  
 Inst : Instrumen  
 Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\011402.M (RTE Integrator)  
 Title : Quantitation For Method 525.2  
 Smoothing : ON Filtering: 5  
 Sampling : 1 Min Area: 2500 Area counts  
 Start Thrs: 0.15 Max Peaks: 10  
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >  
 Peak separation: 5

Signal : TIC

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 27.883      | 1846          | 1853        | 1863         | rM       | 5317048        | 10569913      | 100.00%         | 88.003%       |
| 2         | 28.472      | 1915          | 1918        | 1920         | rM 3     | 84929          | 135385        | 1.28%           | 1.127%        |
| 3         | 30.105      | 2093          | 2098        | 2102         | rM 2     | 578926         | 1305491       | 12.35%          | 10.869%       |

Sum of corrected areas: 12010789

*13% of total*

Area Percent Report

Data File : C:\MSDCHEM\1\DATA\121701\DDT2.D

Vial: 2

Acq On : 18 Dec 2001 2:47 am

Operator: McLean

Sample : 1126

Inst : Instrumen

Misc :

Multiplr: 1.00

MS Integration Params: BENZO.P

Method : C:\MSDCHEM\1\5973N\011402.M (RTE Integrator)

Title : Quantitation For Method 525.2

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 2500 Area counts

Start Thrs: 0.15

Max Peaks: 10

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Signal : TIC

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 28.273      | 1890          | 1896        | 1901         | rM       | 1050548        | 2062498       | 21.42%          | 17.639%       |
| 2         | 29.144      | 1984          | 1992        | 2003         | rM       | 5333919        | 9630323       | 100.00%         | 82.361%       |

*-DDT*

Sum of corrected areas: 11692821