

DFTPP

Data File : C:\MSDCHEM\1\DATA\032202\DFTJ0322.D

Acq On : 22 Mar 2002 7:23 pm

Sample : DFTPP

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: McLean

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080302.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 3531

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	57.3	351424	PASS
68	69	0.00	2	1.4	4297	PASS
69	198	0.00	100	48.7	298688	PASS
70	69	0.00	2	0.5	1443	PASS
127	198	40	60	51.1	313216	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	612800	PASS
199	198	5	9	6.8	41496	PASS
275	198	10	30	21.7	133056	PASS
365	198	1	100	2.6	16065	PASS
441	443	0.01	100	70.2	68664	PASS
442	198	40	110	85.4	523264	PASS
443	442	17	23	18.7	97840	PASS

DFTJ0322.D 5080302.M

Fri Apr 05 07:39:55 2002