

DFTPP

Data File : C:\MSDCHEM\1\DATA\040502\DFAAA405.D
Acq On : 5 Apr 2002 7:38 pm
Sample : DFTPP
Misc :
MS Integration Params: BENZO.P

Vial: 1
Operator: Bohlk
Inst : Instrumen
Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080402.M (RTE Integrator)
Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 3518

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	44.9	291200	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	47.8	310080	PASS
70	69	0.00	2	0.4	1306	PASS
127	198	40	60	50.5	327552	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	648448	PASS
199	198	5	9	6.6	42704	PASS
275	198	10	30	19.4	125928	PASS
365	198	1	100	2.1	13641	PASS
441	443	0.01	100	73.7	52216	PASS
442	198	40	110	57.0	369408	PASS
443	442	17	23	19.2	70824	PASS

DFAAA405.D 5080402.M Mon Apr 08 05:04:34 2002

File : C:\MSDCHEM\1\DATA\040502\DFAAA405.D
 Operator : Bohlk
 Acquired : 5 Apr 2002 7:38 pm using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name: DFTPP
 Misc Info :
 Vial Number: 1

