

Reviewed by _____
Date _____

5085CAW

Method 525.2

File Directory 041802

Storage Directory _____

Instrument 213

Lab #	Filename	Description	Vial #		Inj.	Standards used - Lot #			Int. Std.	Remarks	Initials
			front	rear		Calibration	Reference	Surrogate			
	DFTAD418	DFTPP	1		1	21106					TB
	C5A0418	C5	2			020412					
	C25A0418	C25	3			1					
	SURAD418	Surrogate	4			2249				DDD	
	LBAD418	LB	5						2286m		
	REFAD418	Ref 1	6						1		
	REFBO418	Ref 2	7								
	MCAD418	MC	8								
	T173	sample	9						2286m		
	MCIS MC	with IS	10								
	T174B	sample	11						2286m		
	T175B		12								
	T176		13								
	T177		14								
	T178		15								
	MCBO418	MC	16								
	DFTBO418	DFTPP	17			21106					
	C40B0418	C40	18			020412					
	C40CD418	C40	18			1					
	C100B418	C100	19								
	SURBO418	Surrogate	20			2249				DDD	
	LBBO418	LB	21						2286b		
	REFCO418	Ref 1	22						1		
	REFDO418	Ref 2	23								
	MCCD418	MC	24								
	T376	sample	25						2286b		
	T377		26						1		
	T378		27								

Date 4/18/02Extraction date 3/27/02

Page _____

3/28/02

Method 525.2

Instrument 213

File Directory V 7 1 D 2 5

Storage Directory

[illegible]

Date_____

4/18/02

Extraction date

3/28/02

Page_

Sequence Name: C:\MSDCHEM\1\SEQUENCE\041802.S

Comment:

Operator: Bohlk

Data Path: C:\MSDCHEM\1\DATA\041802\

Pre-Seq Cmd:

Post-Seq Cmd:

Method Sections To Run On A Barcode Mismatch
(X) Full Method (X) Inject Anyway
() Reprocessing Only () Don't Inject

Line	Type	Vial	DataFile	Method	Sample Name
1	Sample	1	DFTA0418	508SCAN	DFTPP
2	Sample	2	C5A0418	508SCAN	C5
3	Sample	3	C25A0418	508SCAN	C25
4	Sample	4	SURA0418	508SCAN	SURROGATE, DDD
5	Sample	5	LBA0418	508SCAN	LB 3/27/02
6	Sample	6	REFA0418	508SCAN	REF1 3/27/02
7	Sample	7	REFB0418	508SCAN	REF2 3/27/02
8	Sample	8	MCA0418	508SCAN	MC
9	Sample	9	7173	508SCAN	SAMPLE 7173 3/27/02
10	Sample	10	MCIS	508SCAN	MC WITH IS
11	Sample	11	7174B	508SCAN	SAMPLE 7174 3/27/02
12	Sample	12	7175B	508SCAN	SAMPLE 7175 3/27/02
13	Sample	13	7176	508SCAN	SAMPLE 7176 3/27/02
14	Sample	14	7177	508SCAN	SAMPLE 7177 3/27/02
15	Sample	15	7178	508SCAN	FB, SAMPLE 7178 3/27/02
16	Sample	16	MCB0418	508SCAN	MC
17	Sample	17	DFTB0418	508SCAN	DFTPP
18	Sample	18	C40B0418	508SCAN	C40
19	Sample	18	C40C0418	508SCAN	C40
20	Sample	19	C100B418	508SCAN	C100
21	Sample	20	SURB0418	508SCAN	SURROGATE, DDD
22	Sample	21	LBB0418	508SCAN	LB 3/28/02
23	Sample	22	REFC0418	508SCAN	REF1 3/28/02
24	Sample	23	REFD0418	508SCAN	REF2 3/28/02
25	Sample	24	MCC0418	508SCAN	MC
26	Sample	25	7376	508SCAN	SAMPLE 7376 3/28/02
27	Sample	26	7377	508SCAN	SAMPLE 7377 3/28/02
28	Sample	27	7378	508SCAN	SAMPLE 7378 3/28/02
29	Sample	28	7379	508SCAN	SAMPLE 7379 3/28/02, FB
30	Sample	29	7380	508SCAN	SAMPLE 7380 3/28/02
31	Sample	30	7381	508SCAN	SAMPLE 7381 3/28/02
32	Sample	31	7382	508SCAN	SAMPLE 7382 3/28/02
33	Sample	32	7383	508SCAN	SAMPLE 7383 3/28/02, FB

DFTPP

Data File : C:\MSDCHEM\1\DATA\041802\DFTA0418.D

Acq On : 18 Apr 2002 8:24 am

Sample : DFTPP

Misc :

MS Integration Params: BENZO.P

Vial: 1

Operator: Bohlk

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080402BBC.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 3506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	47.4	157952	PASS
68	69	0.00	2	1.3	2079	PASS
69	198	0.00	100	49.1	163584	PASS
70	69	0.00	2	0.6	936	PASS
127	198	40	60	49.6	165184	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	332928	PASS
199	198	5	9	7.1	23768	PASS
275	198	10	30	19.3	64272	PASS
365	198	1	100	2.2	7488	PASS
441	443	0.01	100	78.7	25200	PASS
442	198	40	110	52.3	174080	PASS
443	442	17	23	18.4	32016	PASS

DFTA0418.D 5080402BBC.M

Thu Apr 18 09:08:28 2002

File : C:\MSDCHEM\1\DATA\041802\DFTA0418.D
 Operator : Bohlk
 Acquired : 18 Apr 2002 8:24 am using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name: DFTPP
 Misc Info :
 Vial Number: 1

