

Method 508 mod

Date Extracted 3-28-02

Page 33

Ref

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH	Count		
LB		500ml		1ml	1ml		1ml	7A	7			
Ref					1ml			7B				
Ref					x			4B				
7376								1A				
7377								6A				
7378								2B				
7379	FB							1B				
7380								4A				
7381								2A				
7382								8A				
7383	FB							8B				
Remarks												
Reviewed by, <i>LB 8/14/02</i>												
drip rate must be between 5-15 ml/min for liq-liq extractors												
Extraction type <u>Sep Fun</u> Solv. lot# <u>X01E19</u> NaCl lot# <u>V39618</u> Na2SO4 lot# <u>V46623</u>												
STANDARD INT STD LOG#												
CALIBRATION												
2193												
2286												

Verified by _____

Reviewed by _____
Date 4/24/02

5085CAN

Method 525.2

Instrument 213

File Directory 041802

Storage Directory

Lab #	Filename	Description	Vial #		Inj.	Standards used - Lot #			Int. Std.	Remarks	Initials
			front	rear		ul	Calibration	Reference	Surrogate		
	DFTAD418	DFTPP	1		1		21106				TB
	C5A0418	C5	2				020412				
	C25A0418	C25	3				1				
	SWRAD418	Surrogate	4				2249			DDD	
	LBAD418	LB	5						2286m		
	REFAD418	Ref 1	6								
	REFBD418	Ref 2	7								
	MCAD418	MC	8								
	7173	Sample	9						2286m		
	MCE6	MC with IS	10								
	7174B	Sample	11						2286m		
	7175B		12								
	7176		13								
	7177		14								
	7178		15								
	MCB0418	MC	16								TB
	DFTB0418	DFTPP	17				21106				
	C40B0418	C40	18				020412				
	C40ED418	C40	18				1				
	C100B418	C100	19								
	SWRB0418	Surrogate	20				2249			DDD	
	LBBO418	LB	21						2286b		
	REFCO418	Ref 1	22								
	REFDO418	Ref 2	23								
	MCC0418	MC	24								
	7376	Sample	25						2286b		
	7377		26								
	7378		27								

Date 4/18/02

Extraction date 3/27/02
3/28/02

Page

DFTPP

Data File : C:\MSDCHEM\1\DATA\041802\DFTB0418.D

Acq On : 18 Apr 2002 9:57 pm

Sample : DFTPP

Misc :

MS Integration Params: BENZO.P

Vial: 17

Operator: Bohlk

Inst : Instrumen

Multiplr: 1.00

Method : C:\MSDCHEM\1\5973N\5080402BBC.M (RTE Integrator)

Title : Quantitation For Method 508gcscan

Spectrum Information: Scan 3506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	46.3	169932	PASS
68	69	0.00	2	1.4	2509	PASS
69	198	0.00	100	48.8	179136	PASS
70	69	0.00	2	0.4	696	PASS
127	198	40	60	50.6	185664	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	367098	PASS
199	198	5	9	6.6	24136	PASS
275	198	10	30	19.0	69752	PASS
365	198	1	100	1.8	6646	PASS
441	443	0.01	100	78.7	28520	PASS
442	198	40	110	50.5	185216	PASS
443	442	17	23	19.6	36232	PASS

DFTB0418.D 5080402BBC.M

Fri Apr 19 08:13:18 2002

File : C:\MSDCHEM\1\DATA\041802\DFTB0418.D
 Operator : Bohlk
 Acquired : 18 Apr 2002 9:57 pm using AcqMethod 508SCAN
 Instrument : Instrumen
 Sample Name: DFTPP
 Misc Info :
 Vial Number: 17

