

User: Cai, Jian
 Westchester Dept. of Labs & Research
 Date: 09/18/2001 Time: 08:24:36

Sample: AD22471
 Analysis: SSD549 Diquat
 Units: ug
 Started: 9/17/01 09:00 Ended: 9/17/01 13:00 Analyst: CAI
 Result source: manual entry
 Page: 1

[Handwritten signatures]

Component/Name	Viol	Result	Qualifiers	MDL
1. Diquat		33.8		2.0

User: Cai, Jian
 Westchester Dept. of Labs & Research
 Date: 09/18/2001 Time: 08:24:40

Sample: AD22471
 Analysis: SP5549 Diquat
 Units: ug
 Started: 9/18/01 08:21 Ended: 9/18/01 08:21 Analyst: CAI
 Result source: calculation
 Page: 1

	Component Name	Viol.	Result	Qualifier	MDL
1	Diquat		0.20		2.0

Current Date 9/17/2001


1 of 1

Sample Information

SampleName AD22471-SD
Vial 21
Injection 1
Injection Volume 100.00 ul
Channel 996
Run Time 5.0 Minutes

Sample Type Unknown
Date Acquired 9/17/2001 11:37:39 AM
Acq Method Set DIQUAT
Processing Method Diquat010917
Date Processed 9/17/2001 12:56:23 PM

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	Height	Amount	Units
1	Diquat	2.254	24530	3730	33.752	ug/L

PDA Result Table

	Name	RT	Purity1 Angle	Purity1 Threshold	Match1 Spect. Name	Match1 Angle	Match1 Threshold
1	Diquat	2.254	15.116	0.711	Diquat	3.248	