

User: Nefliu, Marcela
Westchester Dept. of Labs & Research
Date: 09/26/2001 Time: 11:00:24

Sample: AD22471
Analysis: \$515 Herbicides - 515.1
Units: ug
Started: 9/17/01 10:58 Ended: 9/25/01 10:58 Analyst: MN
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Dalapon		< MDL		0.80
2	Dicamba		< MDL		0.080
3	2,4-D		< MDL		0.080
4	Pentachlorophenol		< MDL		0.040
5	Silvex		< MDL		0.080
6	2,4,5-T		< MDL		0.10
7	Dinoseb		< MDL		0.16
8	DCPA, di-acid		< MDL		1.0
9	Picloram		< MDL		0.08
10	Surr - 2,4-DCAA		1.4		

Quantitation Report (QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\515F\010925\22371.D\ECD1A.CH Vial: 28
Signal #2 : C:\HPCHEM\1\DATA\515F\010925\22371.D\ECD2B.CH
Acq On : 25 Sep 2001 3:07 pm Operator: Nefliu
Sample : sample Inst : 104
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Sep 26 8:55 2001 Quant Results File: 010925H.RES

Quant Method : C:\HPCHEM\1\METHODS\010925H.M (Chemstation Integrator)
Title : CHLORINATED ACIDS 515.1 1/25/98
Last Update : Wed Sep 26 08:35:24 2001
Response via : Initial Calibration
DataAcq Meth : 515AF.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/l	ug/l

System Monitoring Compounds						
2) S DCAA SURROGATE	13.58	12.81	325959	497559	1.731	1.438m
Spiked Amount 2.000			Recovery	=	86.55%	71.90%
Target Compounds						
1) DALAPON	4.91	4.84	72354	59311	0.156	0.113 #
4) 2,4-D	15.43	14.77	15294	26199	0.025	0.038 #

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\515F\010925\22371.D\ECD1A.CH Vial: 28
Signal #2 : C:\HPCHEM\1\DATA\515F\010925\22371.D\ECD2B.CH
Acq On : 25 Sep 2001 3:07 pm Operator: Nefliu
Sample : sample Inst : 104
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Sep 26 8:55 2001 Quant Results File: 010925H.RES

Quant Method : C:\HPCHEM\1\METHODS\010925H.M (Chemstation Integrator)
Title : CHLORINATED ACIDS 515.1 1/25/98
Last Update : Wed Sep 26 08:35:24 2001
Response via : Multiple Level Calibration
DataAcq Meth : 515AF.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

