


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/24/2001 Time: 11:08:53

Sample: AD23653
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/5/01 11:01 Ended: 10/16/01 11:01 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.1		1.0
20	Surr_Triphenyl phosphate		5.3		1.0
21	Surr_Perylene-d12		5.8		1.0

Chromatogram Plot File: E:\23653TTF Date: Oct-17-1991 01:21:12
Comment: BOHLK; METHOD 525; INST 204; 10/16/01; SAMPLE 23653
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7445884



Integration Report

Quanfile: 23653TTF

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 10/16/01; SAMPLE 23653

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:31	751	Auto	2,091,330	1,047,979
2	phenanthrene d-10 (IS)	17:15	1035	Auto	3,373,793	1,557,180
3	chrysene d-12(IS)	22:57	1377	Auto	2,606,737	1,027,017
4	p-terphenyl d-14(IS)	21:01	1261	Auto	3,676,204	2,337,018
5	acenaphthene d-10(SURR)	12:31	751	Auto	2,091,330	1,047,979
6	phenanthrene d-10(SURR)	17:15	1035	Auto	3,373,793	1,557,180
7	chrysene d-12 (SURR)	22:57	1377	Auto	2,606,737	1,027,017
8	1,3-dimethyl-2-nitrobenzene	8:11	491	Auto	1,287,562	716,140
9	triphenyl phosphate (S)	22:19	1339	Auto	1,958,344	880,694
10	perylene d-12 (SURR)	27:30	1650	Auto	1,865,051	356,920
11	hexachlorocyclopentadiene	10:21	621	Auto	904	904
12	bis(2-ethylhexyl)adipate	22:19	1339	Auto	28,086	7,708
13	bis(2-ethylhexyl)phthalate	23:14	1394	Auto	320,090	90,748
14	2,6-dinitrotoluene	12:10	730	Auto	76,204	39,631
15	2,4-dinitrotoluene	13:18	798	Auto	5,251	2,436
16	4,4'-dichlorodiphenyl ether	20:57	1257	Auto	4,979	3,702

Quantitation Report

Quanfile: 23653TTF

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 10/16/01; SAMPLE 23653

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	751	12:31	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1035	17:15	BB	5.000	UG/L
3	chrysene d-12(IS)	I	1377	22:57	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	751	12:31	BB	4.084	UG/L
6	phenanthrene d-10(SURR)	A	1035	17:15	BB	4.472	UG/L
7	chrysene d-12 (SURR)	A	1377	22:57	BB	4.993	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	491	8:11	BV	5.138	UG/L
9	triphenyl phosphate (S)	A	1339	22:19	BV	5.258	UG/L
10	perylene d-12 (SURR)	A	1650	27:30	BB	5.765	UG/L
11	hexachlorocyclopentadiene	A	621	10:21	BB	0.016	UG/L
16	bis(2-ethylhexyl)adipate	A	1339	22:19	MM	0.066	UG/L
17	bis(2-ethylhexyl)phthalate	A	1394	23:14	VV	0.399	UG/L
20	2,6-dinitrotoluene	A	730	12:10	BB	0.516	UG/L
21	2,4-dinitrotoluene	A	798	13:18	BB	0.030	UG/L
28	4,4'-dichlorodiphenyl ether	A	1257	20:57	BB	0.018	UG/L

renewed IS/sur's

Comments for Sample AD23653 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-04-2001 Time: 12:06:41

The recovery for Terbacil in the LCS Standard was above its acceptance range. Recoveries for four and three of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample respectively, were outside of their acceptance ranges.