

Action Taken On Failed QC Analyses

LAB #	DATE EXT.	COMMENT:	NAME
10/1/01		525	TB
23260		-	
23264			
		calibrated/quantitated simazine using masses 201 + 55 instead of 201	
		calibrated/quantitated EPTC using total mass instead of 128	
		terbacid based on 4 concentrations only, C5 - C8 (includes MRL level).	
		sample 22775 was re-injected - it belongs with batch 9/25-9/26/01 It will be re-extracted.	
		phthalate in the LB is > 10% of the phthalate in the positive samples.	