


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/10/2001 Time: 12:20:57

Sample: AD23261
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/1/01 12:19 Ended: 10/6/01 12:19 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.2		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.0		1.0
20	Surr_Triphenyl phosphate		5.0		1.0
21	Surr_Perylene-d12		6.5		1.0

flag ↑

Chromatogram Plot File: C:\23261BB Date: Oct-07-1991 03:44:31
Comment: BOHLK; METHOD 525; INST 204; 10/06/01; SAMPLE 23261
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 7043874



Integration Report Quanfile: 23261BB Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/06/01; SAMPLE 23261
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:34	754	Auto	1,703,640	946,649
2	phenanthrene d-10 (IS)	17:18	1038	Auto	2,561,139	1,123,790
3	chrysene d-12 (IS)	22:58	1378	Auto	1,759,092	601,696
4	p-terphenyl d-14 (IS)	21:03	1263	Auto	2,930,390	2,207,661
5	acenaphthene d-10 (SURR)	12:34	754	Auto	1,703,640	946,649
6	phenanthrene d-10 (SURR)	17:18	1038	Auto	2,561,139	1,123,790
7	chrysene d-12 (SURR)	22:58	1378	Auto	1,759,092	601,696
8	1,3-dimethyl-2-nitrobenzene	8:13	493	Auto	984,447	530,217
9	triphenyl phosphate (S)	22:21	1341	Auto	1,300,892	539,826
10	perylene d-12 (SURR)	27:33	1653	Auto	1,357,497	256,127
11	hexachlorocyclopentadiene	10:24	624	Auto	2,570	1,630
12	bis(2-ethylhexyl)adipate	22:21	1341	Auto	31,842	5,633
13	bis(2-ethylhexyl)phthalate	23:16	1396	Auto	613,068	155,879
14	2,6-dinitrotoluene	12:13	733	Auto	21,610	10,757
15	4,4'-dichlorodiphenyl ether	20:59	1259	Auto	2,906	1,641

Quantitation Report

Quanfile: 23261BB

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/06/01; SAMPLE 23261

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	754	12:34	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1038	17:18	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1378	22:58	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1263	21:03	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	754	12:34	BB	3.682	UG/L
6	phenanthrene d-10(SURR)	A	1038	17:18	BV	3.938	UG/L
7	chrysene d-12 (SURR)	A	1378	22:58	BB	4.074	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	493	8:13	BV	4.962	UG/L
9	triphenyl phosphate (S)	A	1341	22:21	BB	5.004	UG/L
10	perylene d-12 (SURR)	A	1653	27:33	BB	6.529	UG/L
11	hexachlorocyclopentadiene	A	624	10:24	BB	0.034	UG/L
16	bis(2-ethylhexyl)adipate	A	1341	22:21	MM	0.118	UG/L
17	bis(2-ethylhexyl)phthalate	A	1396	23:16	BV	1.166	UG/L
20	2,6-dinitrotoluene	A	733	12:13	BB	0.183	UG/L
28	4,4'-dichlorodiphenyl ether	A	1259	20:59	BB	0.012	UG/L

reviewed TS/ SWR's

Comments for Sample AD23261 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-12-2001 Time: 11:41:00

The recovery for one of the three Surrogate Standards in this sample was above its acceptance range. The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. The recovery for Terbacil in the Spiked Sample was above its acceptance range.