

User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 10/15/2001 Time: 11:14:01

Sample: AD23345
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 10/4/01 11:10 Ended: 10/10/01 11:10 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.4		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21					


Chromatogram Plot

File: C:\23345CC Date: Oct-10-1991 23:10:40

Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23344

Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0

Plotted: 1 to 1860 Range: 1 to 1860 100% = 7760308



Integration Report Quanfile: 23345CC Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23344
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:33	753	Auto	1,548,470	778,512
2	phenanthrene d-10 (IS)	17:16	1036	Auto	2,474,095	1,145,208
3	chrysene d-12(IS)	22:57	1377	Auto	1,617,864	654,205
4	p-terphenyl d-14(IS)	21:02	1262	Auto	3,550,155	2,496,115
5	acenaphthene d-10(SURR	12:33	753	Auto	1,548,470	778,512
6	phenanthrene d-10(SURR	17:16	1036	Auto	2,474,095	1,145,208
7	chrysene d-12 (SURR)	22:57	1377	Auto	1,617,864	654,205
8	1,3-dimethyl-2-nitrobe	8:12	492	Auto	969,860	525,217
9	triphenyl phosphate (S	22:20	1340	Auto	1,322,083	491,862
10	perylene d-12 (SURR)	27:30	1650	Auto	1,171,196	209,903
11	bis(2-ethylhexyl)adipa	22:20	1340	Auto	24,938	4,197
12	bis(2-ethylhexyl)phtha	23:15	1395	Auto	553,206	133,878

Quantitation Report Quanfile: 23345CC Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23344
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	753	12:33	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1036	17:16	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1377	22:57	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1262	21:02	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	753	12:33	BB	2.762	UG/L
6	phenanthrene d-10(SURR)	A	1036	17:16	BV	3.140	UG/L
7	chrysene d-12 (SURR)	A	1377	22:57	BB	3.093	UG/L
8	1,3-dimethyl-2-nitrobe	A	492	8:12	BV	5.378	UG/L
9	triphenyl phosphate (S	A	1340	22:20	BB	5.529	UG/L
10	perylene d-12 (SURR)	A	1650	27:30	BB	6.125	UG/L
16	bis(2-ethylhexyl)adipa	A	1340	22:20	MM	0.101	UG/L
17	bis(2-ethylhexyl)phtha	A	1395	23:15	BB	1.144	UG/L

revised IS/SURR'S
 checked manually for 5 images at
 miss 201 NTP

Comments for Sample AD23345 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-16-2001 Time: 16:18:11

Recoveries for three of the eighteen compounds in the MDL and LCS Standards were outside of their acceptance ranges. The recovery for one of the three Surrogate Standards in this sample was above its acceptance range. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 2.2 ug/L.