


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 10/15/2001 Time: 11:16:18

Sample: AD23471
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 10/4/01 11:10 Ended: 10/10/01 11:10 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.9		1.0
20	Surr_Triphenyl phosphate		5.5		1.0
21	Surr_Paraldehyde		USPCC 5.5		1.0

Chromatogram Plot File: C:\23471FF Date: Oct-11-1991 00:53:51
Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23471
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6961857



Integration Report Quanfile: 23471FF Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23471
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:32	752	Auto	1,672,429	902,102
2	phenanthrene d-10 (IS)	17:15	1035	Auto	2,518,751	1,071,212
3	chrysene d-12 (IS)	22:56	1376	Auto	1,637,516	660,935
4	p-terphenyl d-14 (IS)	21:01	1261	Auto	3,081,756	2,207,692
5	acenaphthene d-10 (SURR)	12:32	752	Auto	1,672,429	902,102
6	phenanthrene d-10 (SURR)	17:15	1035	Auto	2,518,751	1,071,212
7	chrysene d-12 (SURR)	22:56	1376	Auto	1,637,516	660,935
8	1,3-dimethyl-2-nitrobenzene	8:11	491	Auto	958,689	477,083
9	triphenyl phosphate (S)	22:19	1339	Auto	1,335,370	507,863
10	perylene d-12 (SURR)	27:29	1649	Auto	1,329,666	245,684
11	bis(2-ethylhexyl)adipate	22:19	1339	Auto	13,903	4,340
12	bis(2-ethylhexyl)phthalate	23:14	1394	Auto	213,026	48,813
13	2,6-dinitrotoluene	12:11	731	Auto	17,850	8,746
14	2,4-dinitrotoluene	13:18	798	Auto	1,047	1,047

Quantitation Report

Quanfile: 23471FF

Quan Entries: 14

Comment: BOHLK; METHOD 525; INST 204; 10/10/01; SAMPLE 23471

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	752	12:32	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1035	17:15	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1376	22:56	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	752	12:32	BB	3.437	UG/L
6	phenanthrene d-10(SURR)	A	1035	17:15	BV	3.682	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BB	3.606	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	491	8:11	BB	4.922	UG/L
9	triphenyl phosphate (S)	A	1339	22:19	BB	5.517	UG/L
10	perylene d-12 (SURR)	A	1649	27:29	BB	6.870	UG/L
16	bis(2-ethylhexyl)adipate	A	1339	22:19	MM	0.056	UG/L
17	bis(2-ethylhexyl)phthalate	A	1394	23:14	VB	0.441	UG/L
20	2,6-dinitrotoluene	A	731	12:11	BB	0.154	UG/L
21	2,4-dinitrotoluene	A	798	13:18	BB	0.008	UG/L

reviewed IS/Surr's
 Checked manually for Sumazine
 at mass 201 N.P

Comments for Sample AD23471 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 10-16-2001 Time: 16:39:34

Recoveries for three of the eighteen compounds in the MDL and LCS Standards were outside of their acceptance ranges. The recovery for one of the three Surrogate Standards in this sample was above its acceptance range.