


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 09/22/2001 Time: 10:50:38

Sample: AD22488
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 9/18/01 10:46 Ended: 9/19/95 10:46 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.5		1.0
20	Surr_Triphenyl phosphate		6.1		1.0
21	Surr_Perylene-d12		4.7		1.0

Chromatogram Plot File: E:\22488C Date: Sep-19-1991 21:36:41
Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22488
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 3820412



Integration Report Quanfile: 22488C Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22488
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:37	757	Auto	817,887	442,378
2	phenanthrene d-10 (IS)	17:22	1042	Auto	1,390,016	590,960
3	chrysene d-12(IS)	23:03	1383	Auto	1,029,384	443,995
4	p-terphenyl d-14(IS)	21:06	1266	Auto	1,726,536	1,307,180
5	acenaphthene d-10(SURR)	12:37	757	Auto	817,887	442,378
6	phenanthrene d-10(SURR)	17:22	1042	Auto	1,390,016	590,960
7	chrysene d-12 (SURR)	23:03	1383	Auto	1,029,384	443,995
8	1,3-dimethyl-2-nitrobenzene	8:16	496	Auto	546,844	232,224
9	triphenyl phosphate (S)	22:25	1345	Auto	872,759	379,344
10	perylene d-12 (SURR)	27:43	1663	Auto	625,582	124,858
11	bis(2-ethylhexyl)phthalate	23:21	1401	Auto	78,526	23,038
12	2,6-dinitrotoluene	12:15	735	Auto	9,071	5,427
13	2,4-dinitrotoluene	13:23	803	Auto	1,016	1,016

Quantitation Report Quanfile: 22488C Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22488
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	757	12:37	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1042	17:22	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1383	23:03	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1266	21:06	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	757	12:37	BB	3.278	UG/L
6	phenanthrene d-10(SURR	A	1042	17:22	BV	3.751	UG/L
7	chrysene d-12 (SURR)	A	1383	23:03	BB	3.882	UG/L
8	1,3-dimethyl-2-nitrobe	A	496	8:16	BB	5.459	UG/L
9	triphenyl phosphate (S	A	1345	22:25	BV	6.059	UG/L
10	perylene d-12 (SURR)	A	1663	27:43	BB	4.735	UG/L
17	bis(2-ethylhexyl)phtha	A	1401	23:21	MM	0.337	UG/L
20	2,6-dinitrotoluene	A	735	12:15	BB	0.180	UG/L
21	2,4-dinitrotoluene	A	803	13:23	BB	0.016	UG/L

Comments for Sample AD22488 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 09-24-2001 Time: 13:35:44

The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. The recoveries for Terbacil in the LCS Standard and the Spiked Sample were above their acceptance ranges.

Comments for Sample AD22488 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-04-2001 Time: 14:29:00

The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. The recoveries for Terbacil in the LCS Standard and the Spiked Sample were above their acceptance ranges.