

User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 09/22/2001 Time: 10:52:37


Sample: AD22490
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 9/18/01 10:46 Ended: 9/19/95 10:46 Analyst: TB
 Result source: manual entry
 Page: 1

not on bottles

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.68		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.2		1.0
20	Surr_Triphenyl phosphate		11		1.0
21	Surr_Perylene-d12		3.5		1.0

flag

Chromatogram Plot File: E:\22490F Date: Sep-20-1991 01:03:00
Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22490
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 49815850



Integration Report Quanfile: 22490F Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22490
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:38	758	Auto	512,230	264,692
2	phenanthrene d-10 (IS)	17:23	1043	Auto	1,072,734	470,886
3	chrysene d-12 (IS)	23:05	1385	Auto	419,353	159,838
4	p-terphenyl d-14 (IS)	21:08	1268	Auto	2,996,037	1,633,620
5	acenaphthene d-10 (SURR)	12:38	758	Auto	512,230	264,692
6	phenanthrene d-10 (SURR)	17:23	1043	Auto	1,072,734	470,886
7	chrysene d-12 (SURR)	23:05	1385	Auto	419,353	159,838
8	1,3-dimethyl-2-nitrobenzene	8:16	496	Auto	324,882	184,929
9	triphenyl phosphate (S)	22:26	1346	Auto	649,183	285,802
10	perylene d-12 (SURR)	27:44	1664	Auto	186,034	34,902
11	bis(2-ethylhexyl)phthalate	23:22	1402	Auto	66,168	20,999
12	2,6-dinitrotoluene	12:16	736	Auto	3,043	2,120

Quantitation Report Quanfile: 22490F Quan Entries: 12
 Comment: BOHLK; METHOD 525; INST 204; 09/19/01; SAMPLE 22490
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	758	12:38	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1043	17:23	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1385	23:05	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1268	21:08	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	758	12:38	BB	1.183	UG/L
6	phenanthrene d-10 (SURR)	A	1043	17:23	BV	1.668	UG/L
7	chrysene d-12 (SURR)	A	1385	23:05	BB	0.912	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	496	8:16	BB	5.179	UG/L
9	triphenyl phosphate (S)	A	1346	22:26	BB	11.062	UG/L
10	perylene d-12 (SURR)	A	1664	27:44	BB	3.457	UG/L
17	bis(2-ethylhexyl)phthalate	A	1402	23:22	MM	0.679	UG/L
20	2,6-dinitrotoluene	A	736	12:16	BB	0.097	UG/L

Comments for Sample AD22490 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 09-25-2001 Time: 08:48:16

Bis(2-Ethylhexyl)phthalate was present in the Laboratory Blank at 0.37 ug/L. The recovery for bis(2-Ethylhexyl)phthalate in the MDL Standard was above its acceptance range. The recoveries for Terbacil in the LCS Standard and the Spiked Sample were above their acceptance ranges. The recovery for one of the three Surrogate Standards in this sample was above its acceptance range. This sample was submitted to our laboratory using NYCDEP bottles.