


User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 09/20/2001 Time: 10:21:36

Sample: AD22359
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 9/15/01 10:18 Ended: 9/18/01 10:18 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_ 1,3-Dimethyl-2-nitrobenzen		4.9		
20	Surr_ Triphenyl phosphate		5.9		
21	Surr_ Perylene-d12		5.0		

Chromatogram Plot File: E:\22359B Date: Sep-18-1991 16:09:33
Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22359
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 3173063



Integration Report Quanfile: 22359B Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22359
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:38	758	Auto	928,013	494,303
2	phenanthrene d-10 (IS)	17:23	1043	Auto	1,489,825	707,284
3	chrysene d-12(IS)	23:04	1384	Auto	1,061,867	440,159
4	p-terphenyl d-14(IS)	21:07	1267	Auto	1,594,018	1,031,074
5	acenaphthene d-10(SURR)	12:38	758	Auto	928,013	494,303
6	phenanthrene d-10(SURR)	17:23	1043	Auto	1,489,825	707,284
7	chrysene d-12 (SURR)	23:04	1384	Auto	1,061,867	440,159
8	1,3-dimethyl-2-nitrobenzene	8:16	496	Auto	553,229	296,929
9	triphenyl phosphate (S)	22:26	1346	Auto	879,823	356,104
10	perylene d-12 (SURR)	27:44	1664	Auto	683,477	136,346
11	bis(2-ethylhexyl)phthalate	23:22	1402	Auto	272,119	83,208
12	2,6-dinitrotoluene	12:16	736	Auto	20,668	12,943
13	2,4-dinitrotoluene	13:23	803	Auto	2,190	1,254
14	4,4'-dichlorodiphenyl ether	21:02	1262	Auto	1,197	1,197

Quantitation Report Quanfile: 22359B Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22359
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	758	12:38	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1043	17:23	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1384	23:04	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1267	21:07	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	758	12:38	BB	4.028	UG/L
6	phenanthrene d-10(SURR	A	1043	17:23	BV	4.354	UG/L
7	chrysene d-12 (SURR)	A	1384	23:04	BB	4.337	UG/L
8	1,3-dimethyl-2-nitrobe	A	496	8:16	BB	4.867	UG/L
9	triphenyl phosphate (S	A	1346	22:26	BB	5.921	UG/L
10	perylene d-12 (SURR)	A	1664	27:44	BB	5.015	UG/L
17	bis(2-ethylhexyl)phtha	A	1402	23:22	VB	1.072	UG/L
20	2,6-dinitrotoluene	A	736	12:16	BB	0.357	UG/L
21	2,4-dinitrotoluene	A	803	13:23	BB	0.030	UG/L
28	4,4'-dde	A	1262	21:02	BB	0.009	UG/L

Comments for Sample AD22359 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 09-20-2001 Time: 13:56:29

The recoveries for 4 of the 18 compounds in the MDL Standard were above their acceptance ranges. This was due to contamination in the MDL Standard analysis. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.35 ug/L. The recovery for Terbacil in the LCS Standard was above its acceptance range.