

User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 09/20/2001 Time: 10:23:54

Sample: AD22470
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 9/15/01 10:18 Ended: 9/18/01 10:18 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.73		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.1		
20	Surr_Triphenyl phosphate		6.0		
21	Surr_Perylene-d12		4.5		

Comments for Sample AD22470 - Analysis \$525


Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-04-2001 Time: 14:27:29

The recoveries for 4 of the 18 compounds in the MDL Standard were above their acceptance ranges. This was due to contamination in the MDL Standard analysis. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.35 ug/L. The recovery for Terbacil in the LCS Standard was above its acceptance range.

Chromatogram Plot File: E:\22470D Date: Sep-18-1991 17:21:23
Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22470
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 4128549



Integration Report Quanfile: 22470D Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22470
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:37	757	Auto	927,850	624,152
2	phenanthrene d-10 (IS)	17:22	1042	Auto	1,508,012	817,857
3	chrysene d-12 (IS)	23:04	1384	Auto	1,068,126	392,931
4	p-terphenyl d-14 (IS)	21:06	1266	Auto	1,826,364	1,374,397
5	acenaphthene d-10 (SURR)	12:37	757	Auto	927,850	624,152
6	phenanthrene d-10 (SURR)	17:22	1042	Auto	1,508,012	817,857
7	chrysene d-12 (SURR)	23:04	1384	Auto	1,068,126	392,931
8	1,3-dimethyl-2-nitrobenzene	8:16	496	Auto	577,049	313,194
9	triphenyl phosphate (S)	22:25	1345	Auto	903,365	393,137
10	perylene d-12 (SURR)	27:43	1663	Auto	612,211	117,473
11	bis(2-ethylhexyl)phthalate	23:21	1401	Auto	181,510	49,012
12	2,6-dinitrotoluene	12:16	736	Auto	30,919	15,727
13	2,4-dinitrotoluene	13:22	802	Auto	5,554	2,617
14	4,4'-dichlorodiphenyl ether	21:02	1262	Auto	2,592	1,513

Quantitation Report Quanfile: 22470D Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/17/01; SAMPLE 22470
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	757	12:37	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1042	17:22	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1384	23:04	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1266	21:06	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	757	12:37	BB	3.515	UG/L
6	phenanthrene d-10(SURR	A	1042	17:22	BV	3.847	UG/L
7	chrysene d-12 (SURR)	A	1384	23:04	BB	3.808	UG/L
8	1,3-dimethyl-2-nitrobe	A	496	8:16	BB	5.078	UG/L
9	triphenyl phosphate (S	A	1345	22:25	BB	6.044	UG/L
10	perylene d-12 (SURR)	A	1663	27:43	BB	4.466	UG/L
17	bis(2-ethylhexyl)phtha	A	1401	23:21	VB	0.729	UG/L
20	2,6-dinitrotoluene	A	736	12:16	BB	0.530	UG/L
21	2,4-dinitrotoluene	A	802	13:22	BB	0.075	UG/L
28	4,4'-dde	A	1262	21:02	BB	0.019	UG/L

Comments for Sample AD22470 - Analysis \$525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 09-20-2001 Time: 14:02:31

The recoveries for 4 of the 18 compounds in the MDL Standard were above their acceptance ranges. This was due to contamination in the MDL Standard analysis. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.35 ug/L. The recovery for Terbacil in the LCS Standard was above its acceptance range.