

User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 10/02/2001 Time: 11:50:52

Sample: AD22582
 Analysis: \$M_525 Organic Chemicals - 525.2; mdl
 Units: ug
 Started: 9/24/01 11:40 Ended: 9/29/01 11:40 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		0.13		0.10
2	Hexachlorobenzene		0.18		0.10
3	Simazine		0.20		0.07
4	Atrazine		0.17		0.10
5	Alachlor		0.24		0.20
6	bis(2-Ethylhexyl)adipate		0.19		0.60
7	bis(2-Ethylhexyl)phthalate		0.40		0.60
8	Benzo(a)pyrene		0.087		0.20
9	EPTC		0.86		1.0
10	2,6-Dinitrotoluene		2.1		2.0
11	2,4-Dinitrotoluene		1.9		2.0
12	Molinate		1.0		0.90
13	Terbacil		3.2		2.0
14	Acetochlor		2.1		2.0
15	Metribuzin		0.20		0.15
16	Metolachlor		0.24		0.75
17	Butachlor		0.26		0.40
18	4,4-DDE		0.69		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.2		
20	Surr_Triphenyl phosphate		4.9		
21	Surr_Perylene-d12		5.4		

User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/02/2001 Time: 11:50:44

Sample: AD22582
Analysis: \$MR525 Organic Chemicals - 525.2; mdl
Units: ug
Started: 9/24/01 11:50 Ended: 9/29/01 11:50 Analyst: TB
Result source: calculation
Page: 1

*No limit
I like 0-200%*

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		65		0.10
2	Hexachlorobenzene		90		0.10
3	Simazine		100		0.07
4	Atrazine		85		0.10
5	Alachlor		120		0.20
6	bis(2-Ethylhexyl)adipate		95		0.60
7	bis(2-Ethylhexyl)phthalate		200		0.60
8	Benzo(a)pyrene		43.5		0.20
9	EPTC		95.55556		1.0
10	2,6-Dinitrotoluene		105		2.0
11	2,4-Dinitrotoluene		95		2.0
12	Molinate		100		0.90
13	Terbacil		160		2.0
14	Acetochlor		105		2.0
15	Metribuzin		100		0.15
16	Metolachlor		120		0.75
17	Butachlor		130		0.40
18	4,4-DDE		86.25		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		104		
20	Surr_Triphenyl phosphate		98		
21	Surr_Perylene-d12		108		

Chromatogram Plot

File: E:\MDLA929

Date: Sep-29-1991 22:21:45

Comment: BOHLK; METHOD 525; INST 204; 09/29/01; MDL

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1860

Range: 1 to 1860

100% = 8121221



Integration Report

Quanfile: MDLA929

Quan Entries: 28

Comment: BOHLK; METHOD 525; INST 204; 09/29/01; MDL

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
3	chrysene d-12(IS)	22:56	1376	Auto	2,135,521	897,453
4	p-terphenyl d-14(IS)	21:01	1261	Auto	3,322,676	2,520,833
5	acenaphthene d-10(SURR)	12:32	752	Auto	1,897,898	1,200,312
6	phenanthrene d-10(SURR)	17:15	1035	Auto	3,062,217	1,418,518
7	chrysene d-12 (SURR)	22:56	1376	Auto	2,135,521	897,453
8	1,3-dimethyl-2-nitrobenzene	8:12	492	Auto	1,212,842	545,974
9	triphenyl phosphate (S)	22:19	1339	Auto	1,443,087	573,640
10	perylene d-12 (SURR)	27:29	1649	Auto	1,320,716	236,717
11	hexachlorocyclopentadiene	10:23	623	Auto	7,010	3,867
12	hexachlorobenzene	15:42	942	Auto	37,665	15,824
13	simazine	16:28	988	Auto	22,340	7,068
14	atrazine	16:39	999	Auto	16,485	7,564
15	alachlor	18:37	1117	Auto	45,945	26,772
16	bis(2-ethylhexyl)adipate	22:11	1331	Auto	67,893	14,427
17	bis(2-ethylhexyl)phthalate	23:14	1394	Auto	252,381	60,582
18	benzo(a)pyrene	27:14	1634	Auto	32,328	3,930
19	epicloric acid	10:47	647	Auto	892,694	436,464
20	2,6-dinitrotoluene	12:09	729	Auto	296,746	167,147
21	2,4-dinitrotoluene	13:08	788	Auto	349,087	191,760
22	molinic acid	13:17	797	Auto	386,254	232,291
23	terbacil	17:37	1057	Auto	229,021	89,066
24	acetochlor	18:26	1106	Auto	289,111	184,638
25	metribuzin	18:28	1108	Auto	41,222	18,323
26	metolachlor	19:21	1161	Auto	130,114	59,348
27	butachlor	20:33	1233	Auto	52,107	22,529
28	4,4'-dichlorodiphenyl ether	20:57	1257	Auto	188,365	116,606

Quantitation Report Quanfile: MDLA929
 Comment: BOHLK; METHOD 525; INST 204; 09/29/01; MDL
 Sorted via: Entry Number ↑

Quan Entries: 28

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
3	chrysene d-12(IS)	I	1376	22:56	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	752	12:32	BB	4.187	UG/L
6	phenanthrene d-10(SURR)	A	1035	17:15	BV	4.491	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BB	4.560	UG/L
8	1,3-dimethyl-2-nitrobe	A	492	8:12	BV	5.204	UG/L
9	triphenyl phosphate (S	A	1339	22:19	BB	4.893	UG/L
10	perylene d-12 (SURR)	A	1649	27:29	BB	5.390	UG/L
11	hexachlorocyclopentadi	A	623	10:23	BB	0.126	UG/L
12	hexachlorobenzene	A	942	15:42	BB	0.182	UG/L
13	simazine	A	988	16:28	BB	0.195	UG/L
14	atrazine	A	999	16:39	BB	0.167	UG/L
15	alachlor	A	1117	18:37	BB	0.243	UG/L
16	bis(2-ethylhexyl)adipa	A	1331	22:11	MM	0.188	UG/L
17	bis(2-ethylhexyl)phtha	A	1394	23:14	MM	0.397	UG/L
18	benzo(a)pyrene	A	1634	27:14	MM	0.087	UG/L
19	eptc	A	647	10:47	MM	0.865	UG/L
20	2,6-dinitrotoluene	A	729	12:09	BB	2.132	UG/L
21	2,4-dinitrotoluene	A	788	13:08	BV	1.876	UG/L
22	molinate	A	797	13:17	BB	1.025	UG/L
23	terbacil	A	1057	17:37	BB	3.161	UG/L
24	acetochlor	A	1106	18:26	BB	2.076	UG/L
25	metribuzin	A	1108	18:28	MM	0.205	UG/L
26	metolachlor	A	1161	19:21	BB	0.244	UG/L
27	butachlor	A	1233	20:33	MM	0.259	UG/L
28	4,4'-dde	A	1257	20:57	BV	0.694	UG/L

IS/sure areas OK