

User: Bohlk, Ted
 Westchester Dept. of Labs & Research
 Date: 10/02/2001 Time: 11:56:24

Sample: AD22654
 Analysis: \$525 Organic Chemicals - 525.2
 Units: ug
 Started: 9/24/01 11:53 Ended: 9/29/01 11:53 Analyst: TB
 Result source: manual entry
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.4		1.0
20	Surr_Triphenyl phosphate		5.1		1.0
21	Surr_Perylene-d12		4.4		1.0

Chromatogram Plot

File: F:\22654CC

Date: Sep-30-1991 00:04:58

Comment: BOHLK; METHOD 525; INST 204; 09/29/01; SAMPLE 22654

Scan No: 1

Retention Time: 0:01


RIC: 0

Mass Range: 0 - 0

Plotted: 1 to 1859

Range: 1 to 1859

100% = 13627055



Integration Report Quanfile: 22654CC Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/29/01; SAMPLE 22654
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:32	752	Auto	2,097,320	1,270,666
2	phenanthrene d-10 (IS)	17:15	1035	Auto	3,530,386	1,452,450
3	chrysene d-12 (IS)	22:56	1376	Auto	2,457,891	998,635
4	p-terphenyl d-14 (IS)	21:01	1261	Auto	4,117,002	3,002,822
5	acenaphthene d-10 (SURR)	12:32	752	Auto	2,097,320	1,270,666
6	phenanthrene d-10 (SURR)	17:15	1035	Auto	3,530,386	1,452,450
7	chrysene d-12 (SURR)	22:56	1376	Auto	2,457,891	998,635
8	1,3-dimethyl-2-nitrobenzene	8:12	492	Auto	1,379,409	626,682
9	triphenyl phosphate (S)	22:19	1339	Auto	1,741,458	735,020
10	perylene d-12 (SURR)	27:29	1649	Auto	1,254,470	238,015
11	hexachlorobenzene	15:42	942	Auto	4,027	2,002
12	bis(2-ethylhexyl)phthalate	23:14	1394	Auto	322,443	89,829
13	2,4-dinitrotoluene	13:18	798	Auto	9,334	3,918
14	4,4'-dichlorodiphenyl ether	20:57	1257	Auto	5,315	3,347

Quantitation Report Quanfile: 22654CC Quan Entries: 14
 Comment: BOHLK; METHOD 525; INST 204; 09/29/01; SAMPLE 22654
 Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	752	12:32	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1035	17:15	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1376	22:56	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BV	5.000	UG/L
5	acenaphthene d-10(SURR)	A	752	12:32	BB	3.734	UG/L
6	phenanthrene d-10(SURR)	A	1035	17:15	BV	4.179	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BB	4.236	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	492	8:12	BB	5.356	UG/L
9	triphenyl phosphate (S)	A	1339	22:19	BV	5.130	UG/L
10	perylene d-12 (SURR)	A	1649	27:29	BB	4.448	UG/L
12	hexachlorobenzene	A	942	15:42	BB	0.018	UG/L
17	bis(2-ethylhexyl)phthalate	A	1394	23:14	BB	0.441	UG/L
21	2,4-dinitrotoluene	A	798	13:18	BB	0.049	UG/L
28	4,4'-dichlorodiphenyl ether	A	1257	20:57	BB	0.017	UG/L
IS/SURR areas OK							

Comments for Sample AD22654 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-04-2001 Time: 11:59:43

The recovery for Terbacil in the LCS Standard was above its acceptance range.