


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/05/2001 Time: 08:51:36

Sample: AD22754
Analysis: \$LP525 Organic Extractables
Units: ug
Started: 10/1/01 08:50 Ended: 10/1/01 08:50 Analyst: TB
Page: 1

	Component Name	Viol	Result
1	DFTPP		PASSED
2	DDT		1.9
3	DDT		1.9

Chromatogram Plot File: E:\DFTA1001 Date: Oct-01-1991 17:13:37
Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1859 Range: 1 to 1859 100% = 5111466



Mass List

File: E:\DFTA1001 Date: Oct-01-1991 17:13:37

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP

Scan No: 1116 Retention Time: 18:36 RIC: 5111466 Mass Range: 47 - 444

Peaks: 176 Mass of Base Peak: 442 Int: 876923 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
47	2,644	0.30	83	5,528	0.63	118	8,413	0.96
49	17,067	1.95	84	5,528	0.63	122	7,211	0.82
50	116,105	13.24	85	4,326	0.49	123	6,730	0.77
51	312,740	35.66	86	10,576	1.21	124	3,605	0.41
52	13,701	1.56	87	2,644	0.30	125	3,125	0.36
55	5,048	0.58	91	3,365	0.38	127	156,971	17.90
56	11,538	1.32	92	5,048	0.58	128	14,663	1.67
57	13,221	1.51	93	31,490	3.59	129	71,394	8.14
61	5,769	0.66	94	3,125	0.36	130	7,692	0.88
62	5,769	0.66	98	31,971	3.65	131	3,125	0.36
63	8,413	0.96	99	13,701	1.56	135	8,413	0.96
65	2,884	0.33	100	1,923	0.22	136	3,365	0.38
67	3,125	0.36	101	5,769	0.66	137	5,528	0.63
68	4,326	0.49	103	5,528	0.63	141	12,500	1.43
69	222,115	25.33	104	6,009	0.69	142	5,048	0.58
73	3,846	0.44	105	5,769	0.66	143	2,644	0.30
74	37,259	4.25	106	2,403	0.27	146	3,605	0.41
75	34,375	3.92	107	70,913	8.09	147	5,528	0.63
76	13,461	1.54	108	8,413	0.96	148	18,269	2.08
77	239,423	27.30	109	5,288	0.60	149	2,884	0.33
78	11,778	1.34	110	137,740	15.71	153	4,326	0.49
79	27,644	3.15	111	18,990	2.17	154	3,125	0.36
80	16,105	1.84	112	3,125	0.36	155	4,567	0.52
81	14,663	1.67	116	4,326	0.49	156	8,413	0.96
82	2,163	0.25	117	84,375	9.62	158	2,163	0.25

Mass List

File: E:\DFTA1001 Date: Oct-01-1991 17:13:37

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP

Scan No: 1116 Retention Time: 18:36 RIC: 5111466 Mass Range: 47 - 444

Peaks: 176 Mass of Base Peak: 442 Int: 876923 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
158	2,163	0.25	198	490,144	55.89	247	2,884	0.33
161	4,086	0.47	199	32,211	3.67	249	2,403	0.27
165	2,884	0.33	203	4,086	0.47	255	312,500	35.64
166	4,086	0.47	204	12,740	1.45	256	42,788	4.88
167	24,278	2.77	205	17,548	2.00	257	2,644	0.30
168	9,615	1.10	206	83,894	9.57	258	12,980	1.48
172	1,923	0.22	207	11,298	1.29	259	3,605	0.41
173	4,086	0.47	208	3,125	0.36	265	5,048	0.58
174	4,086	0.47	211	3,125	0.36	266	1,442	0.16
175	6,490	0.74	215	2,644	0.30	273	7,932	0.90
176	1,442	0.16	217	22,836	2.60	274	34,615	3.95
177	3,605	0.41	218	3,846	0.44	275	165,625	18.89
178	2,644	0.30	221	5,769	0.66	276	25,480	2.91
179	10,817	1.23	223	3,846	0.44	277	19,951	2.28
180	6,250	0.71	224	43,990	5.02	278	3,605	0.41
181	6,250	0.71	225	12,980	1.48	283	2,163	0.25
184	1,442	0.16	227	23,076	2.63	285	2,163	0.25
185	6,730	0.77	228	5,048	0.58	293	3,365	0.38
186	51,442	5.87	229	4,567	0.52	296	31,971	3.65
187	17,067	1.95	230	1,682	0.19	297	3,846	0.44
188	3,605	0.41	233	2,884	0.33	303	4,326	0.49
192	3,846	0.44	243	2,644	0.30	304	2,644	0.30
193	3,125	0.36	244	33,413	3.81	315	3,365	0.38
196	7,692	0.88	245	4,807	0.55	316	2,644	0.30
197	7,211	0.82	246	9,855	1.12	322	1,923	0.22

Mass List

File: E:\DFTA1001 Date: Oct-01-1991 17:13:37

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP

Scan No: 1116 Retention Time: 18:36 RIC: 5111466 Mass Range: 47 - 444

Peaks: 176 Mass of Base Peak: 442 Int: 876923 Threshold: 0 %

Mass	Intensity	% Base	Mass	Intensity	% Base	Mass	Intensity	% Base
203	4,086	0.47	255	312,500	35.64	335	4,567	0.52
204	12,740	1.45	256	42,788	4.88	336	1,682	0.19
205	17,548	2.00	257	2,644	0.30	341	2,403	0.27
206	83,894	9.57	258	12,980	1.48	346	2,403	0.27
207	11,298	1.29	259	3,605	0.41	352	4,807	0.55
208	3,125	0.36	265	5,048	0.58	353	4,326	0.49
211	3,125	0.36	266	1,442	0.16	354	11,298	1.29
215	2,644	0.30	273	7,932	0.90	355	3,365	0.38
217	22,836	2.60	274	34,615	3.95	364	2,163	0.25
218	3,846	0.44	275	165,625	18.89	365	36,298	4.14
221	5,769	0.66	276	25,480	2.91	366	6,490	0.74
223	3,846	0.44	277	19,951	2.28	372	11,538	1.32
224	43,990	5.02	278	3,605	0.41	373	3,365	0.38
225	12,980	1.48	283	2,163	0.25	401	1,442	0.16
227	23,076	2.63	285	2,163	0.25	402	4,567	0.52
228	5,048	0.58	293	3,365	0.38	403	6,490	0.74
229	4,567	0.52	296	31,971	3.65	404	2,403	0.27
230	1,682	0.19	297	3,846	0.44	421	3,365	0.38
233	2,884	0.33	303	4,326	0.49	422	6,250	0.71
243	2,644	0.30	304	2,644	0.30	423	43,750	4.99
244	33,413	3.81	315	3,365	0.38	424	11,057	1.26
245	4,807	0.55	316	2,644	0.30	441	133,173	15.19
246	9,855	1.12	322	1,923	0.22	442	876,923	100.00
247	2,884	0.33	323	13,942	1.59	443	151,923	17.32
249	2,403	0.27	334	9,375	1.07	444	17,067	1.95

Chromatogram Plot

File: E:\DFTA1001

Date: Oct-01-1991 17:13:37

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFIPP

Scan No: 1290

Retention Time: 21:30


RIC: 28748

Mass Range: 45 - 281


Plotted: 1270 to 1290

Range: 1 to 1859


100% = 5111466




Chromatogram Plot File: E:\DFTA1001 Date: Oct-01-1991 17:13:37
 Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP
 Scan No: 1310 Retention Time: 21:50 RIC: 64767 Mass Range: 45 - 431
 Plotted: 1290 to 1310 Range: 1 to 1859 100% = 5111466



Chromatogram Plot File: E:\DFTA1001 Date: Oct-01-1991 17:13:37
Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DFTPP
Scan No: 1380 Retention Time: 23:00 RIC: 26023 Mass Range: 45 - 281
Plotted: 1365 to 1380 Range: 1 to 1859 100% = 5111466



Chromatogram Plot File: E:\DDTA1001 Date: Oct-01-1991 17:48:06
 Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DDT
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 3254960



Chromatogram Plot

File: E:\DDTA1001

Date: Oct-01-1991 17:48:06

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DDT

Scan No: 1280

Retention Time: 21:20


RIC: 26476

Mass Range: 47 - 207


Plotted: 1240 to 1280

Range: 1 to 1860

100% = 3254960



Chromatogram Plot File: E:\DDTA1001 Date: Oct-01-1991 17:48:06
 Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DDT
 Scan No: 1300 Retention Time: 21:40 RIC: 26480 Mass Range: 45 - 207
 Plotted: 1280 to 1300 Range: 1 to 1860 100% = 3254960



Chromatogram Plot

File: E:\DDTA1001

Date: Oct-01-1991 17:48:06

Comment: BOHLK; METHOD 525; INST 204; 10/01/01; DDT

Scan No: 1350

Retention Time: 22:30

RIC: 26717

Mass Range: 45 - 207

Plotted: 1310 to 1350

Range: 1 to 1860

100% = 3254960

