NYC 9/11 Public Portal Document

User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/05/2001 Time: 10:17:47

Sample: AD22773
Analysis: $525 Organic Chemicals - 525.2
Units: ug
Started: 9/25/01 10:15 Ended: 10/1/01 10:15 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		4.7		1.0
20	Surr_Triphenyl phosphate		6.1		1.0
21	Surr_Perylene-d12		4.4		1.0

NYC-WTC_000063677

NYC 9/11 Public Portal Document



NYC-WTC_000063678

NYC 9/11 Public Portal Document

Integration Report Quanfile: 22773BBB Quan Entries: 12
Comment: BOHLK; METHOD 525; INST 204; 10/01/01; SAMPLE 22773
Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:33	753	Auto	1,415,799	953,125
2	phenanthrene d-10 (IS)	17:17	1037	Auto	2,150,983	1,071,455
3	chrysene d-12(IS)	22:58	1378	Auto	1,503,725	648,270
4	p-terphenyl d-14(IS)	21:02	1262	Auto	2,646,177	1,871,428
5	acenaphthene d-10(SURR	12:33	753	Auto	1,415,799	953,125
6	phenanthrene d-10(SURR	17:17	1037	Auto	2,150,983	1,071,455
7	chrysene d-12 (SURR)	22:58	1378	Auto	1,503,725	648,270
8	1,3-dimethyl-2-nitrobe	8:13	493	Auto	866,286	417,881
9	triphenyl phosphate (S	22:20	1340	Auto	1,225,811	506,600
10	perylene d-12 (SURR)	27:33	1653	Auto	815,637	155,908
11	bis(2-ethylhexyl)phtha	23:15	1395	Auto	134,944	32,048
12	2,6-dinitrotoluene	12:12	732	Auto	7,993	4,799

NYC-WTC_000063679

NYC 9/11 Public Portal Document

Quantitation Report Quanfile: 22773BBB Quan Entries: 12
Comment: BOHLK; METHOD 525; INST 204; 10/01/01; SAMPLE 22773
Sorted via: Entry Number ↑ (S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	753	12:33	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1037	17:17	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1378	22:58	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1262	21:02	BB	5.000	UG/L
5	acenaphthene d-10(SURR	A	753	12:33	BB	3.574	UG/L
6	phenanthrene d-10(SURR	A	1037	17:17	BV	3.541	UG/L
7	chrysene d-12 (SURR)	A	1378	22:58	BB	3.651	UG/L
8	1,3-dimethyl-2-nitrobe	A	493	8:13	BB	4.692	UG/L
9	triphenyl phosphate (S	A	1340	22:20	BB	6.054	UG/L
10	perylene d-12 (SURR)	A	1653	27:33	BB	4.366	UG/L
17	bis(2-ethylhexyl)phtha	A	1395	23:15	VB	0.314	UG/L
20	2,6-dinitrotoluene	A	732	12:12	BB	0.082	UG/L

received IS/Surr's

NYC 9/11 Public Portal Document

Comments for Sample AD22773 - Analysis $525

Westchester Dept. of Labs & Research
User: Vinci, David L
Date: 10-08-2001 Time: 12:53:53

Recoveries for three of the eighteen compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for six of the eighteen compounds in the LCS Standard were outside of their acceptance ranges.