


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/05/2001 Time: 16:29:21

Sample: AD22975
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 9/25/01 16:11 Ended: 10/2/01 16:11 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.07
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenzen		5.2		1.0
20	Surr_Triphenyl phosphate		6.2		1.0
21	Surr_Perylene-d12		4.3		1.0

Chromatogram Plot File: E:\229751 Date: Oct-03-1991 01:32:57
 Comment: BOHLK; METHOD 525; INST 204; 10/02/01; SAMPLE 22975
 Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
 Plotted: 1 to 1860 Range: 1 to 1860 100% = 4179852



Integration Report Quanfile: 22975I Quan Entries: 13
 Comment: BOHLK; METHOD 525; INST 204; 10/02/01; SAMPLE 22975
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:32	752	Auto	1,266,546	728,649
2	phenanthrene d-10 (IS)	17:16	1036	Auto	2,094,536	997,302
3	chrysene d-12 (IS)	22:57	1377	Auto	1,345,123	578,022
4	p-terphenyl d-14 (IS)	21:02	1262	Auto	2,193,081	1,179,166
5	acenaphthene d-10 (SURR)	12:32	752	Auto	1,266,546	728,649
6	phenanthrene d-10 (SURR)	17:16	1036	Auto	2,094,536	997,302
7	chrysene d-12 (SURR)	22:57	1377	Auto	1,345,123	578,022
8	1,3-dimethyl-2-nitrobenzene	8:12	492	Auto	865,726	512,701
9	triphenyl phosphate (S)	22:19	1339	Auto	1,117,400	386,287
10	perylene d-12 (SURR)	27:31	1651	Auto	713,578	133,737
11	bis(2-ethylhexyl)phthalate	23:15	1395	Auto	80,901	18,986
12	2,6-dinitrotoluene	12:11	731	Auto	11,816	7,120
13	4,4'-dichlorodiphenyl ether	20:57	1257	Auto	1,277	1,277

Quan Entries: 13

22975

(S) = Standard

[illegible]

Comments for Sample AD22975 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 10-08-2001 Time: 09:27:22

Recoveries for three of the eighteen compounds in the MDL Standard were above their acceptance ranges. The recovery for Terbacil in the LCS Standard was above its acceptance range. The recoveries for six of the eighteen compounds in the Spiked Sample were outside their acceptance ranges.