


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/27/2001 Time: 16:33:55

Sample: AD23834
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/9/01 15:31 Ended: 10/17/01 15:31 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		0.63		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.3		1.0
20	Surr_Triphenyl phosphate		5.8		1.0
21	Surr_Perylene-d12		4.4		1.0

Chromatogram Plot File: F:\23834LL Date: Oct-18-1991 03:46:25
Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23834
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 9752151



Integration Report

Quanfile: 23834LL

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23834

Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10 (IS)	12:31	751	Auto	1,758,000	1,124,709
2	phenanthrene d-10 (IS)	17:15	1035	Auto	2,684,017	1,011,656
3	chrysene d-12 (IS)	22:56	1376	Auto	1,962,605	832,892
4	p-terphenyl d-14 (IS)	21:01	1261	Auto	4,232,584	3,111,551
5	acenaphthene d-10 (SURR)	12:31	751	Auto	1,758,000	1,124,709
6	phenanthrene d-10 (SURR)	17:15	1035	Auto	2,684,017	1,011,656
7	chrysene d-12 (SURR)	22:56	1376	Auto	1,962,605	832,892
8	1,3-dimethyl-2-nitrobe	8:11	491	Auto	1,108,671	472,882
9	triphenyl phosphate (S	22:19	1339	Auto	1,620,411	602,087
10	perylene d-12 (SURR)	27:29	1649	Auto	1,069,724	192,282
11	bis(2-ethylhexyl)adipa	22:10	1330	Auto	35,635	8,497
12	bis(2-ethylhexyl)phtha	23:13	1393	Auto	380,471	107,726
13	2,6-dinitrotoluene	12:10	730	Auto	102,682	61,778
14	2,4-dinitrotoluene	13:17	797	Auto	11,396	5,135
15	4,4'-dde	20:56	1256	Auto	3,595	1,825

Quantitation Report

Quanfile: 23834LL

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23834

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	751	12:31	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1035	17:15	BV	5.000	UG/L
3	chrysene d-12(IS)	I	1376	22:56	BV	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	751	12:31	BB	2.982	UG/L
6	phenanthrene d-10(SURR)	A	1035	17:15	BV	3.090	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BV	3.265	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	491	8:11	BV	5.263	UG/L
9	triphenyl phosphate (S)	A	1339	22:19	BV	5.778	UG/L
10	perylene d-12 (SURR)	A	1649	27:29	BB	4.392	UG/L
16	bis(2-ethylhexyl)adipate	A	1330	22:10	MM	0.110	UG/L
17	bis(2-ethylhexyl)phthalate	A	1393	23:13	BB	0.626	UG/L
20	2,6-dinitrotoluene	A	730	12:10	BB	0.821	UG/L
21	2,4-dinitrotoluene	A	797	13:17	BB	0.077	UG/L
28	4,4'-dichlorodiphenyl ether	A	1256	20:56	BB	0.016	UG/L

reversed IS/ SURR's

Comments for Sample AD23834 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-04-2001 Time: 10:54:16

Recoveries for three of the eighteen compounds in the MDL Standard and the LCS Standard were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.76 ug/L.