


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/27/2001 Time: 15:40:55

Sample: AD23835
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/9/01 15:31 Ended: 10/17/01 15:31 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		< MDL		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		4.7		1.0
20	Surr_Triphenyl phosphate		5.4		1.0
21	Surr_Perylene-d12		4.5		1.0

Chromatogram Plot File: F:\23835MM Date: Oct-18-1991 04:20:51
Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23835
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 6753068



Integration Report Quanfile: 23835MM Quan Entries: 16
 Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23835
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:31	751	Auto	2,372,858	1,077,808
2	phenanthrene d-10 (IS)	17:14	1034	Auto	3,424,193	1,548,393
3	chrysene d-12(IS)	22:56	1376	Auto	2,701,782	1,055,780
4	p-terphenyl d-14(IS)	21:01	1261	Auto	3,812,335	1,944,584
5	acenaphthene d-10(SURR)	12:31	751	Auto	2,372,858	1,077,808
6	phenanthrene d-10(SURR)	17:14	1034	Auto	3,424,193	1,548,393
7	chrysene d-12 (SURR)	22:56	1376	Auto	2,701,782	1,055,780
8	1,3-dimethyl-2-nitrobenzene	8:10	490	Auto	1,346,761	769,117
9	triphenyl phosphate (S)	22:18	1338	Auto	2,076,471	1,000,547
10	perylene d-12 (SURR)	27:28	1648	Auto	1,519,486	278,285
11	hexachlorocyclopentadiene	10:21	621	Auto	1,335	1,335
12	bis(2-ethylhexyl)adipate	22:18	1338	Auto	40,839	10,904
13	bis(2-ethylhexyl)phthalate	23:13	1393	Auto	256,982	75,231
14	2,6-dinitrotoluene	12:10	730	Auto	75,164	30,920
15	2,4-dinitrotoluene	13:17	797	Auto	4,797	2,275
16	4,4'-dichlorodiphenyl ether	20:56	1256	Auto	2,061	2,061

Quantitation Report

Quanfile: 23835MM

Quan Entries: 16

Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23835

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10(IS)	I	751	12:31	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1034	17:14	BU	5.000	UG/L
3	chrysene d-12(IS)	I	1376	22:56	BB	5.000	UG/L
4	p-terphenyl d-14(IS)	I	1261	21:01	BB	5.000	UG/L
5	acenaphthene d-10(SURR)	A	751	12:31	BB	4.468	UG/L
6	phenanthrene d-10(SURR)	A	1034	17:14	BU	4.376	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BB	4.990	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	490	8:10	BB	4.737	UG/L
9	triphenyl phosphate (S)	A	1338	22:18	BB	5.379	UG/L
10	perylene d-12 (SURR)	A	1648	27:28	BB	4.532	UG/L
11	hexachlorocyclopentadiene	A	621	10:21	BB	0.020	UG/L
16	bis(2-ethylhexyl)adipate	A	1338	22:18	MM	0.092	UG/L
17	bis(2-ethylhexyl)phthalate	A	1393	23:13	VV	0.310	UG/L
20	2,6-dinitrotoluene	A	730	12:10	BB	0.450	UG/L
21	2,4-dinitrotoluene	A	797	13:17	BB	0.024	UG/L
28	4,4'-dichlorodiphenyl ether	A	1256	20:56	BB	0.007	UG/L

reviewed IS/Surrogate

Comments for Sample AD23835 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-04-2001 Time: 10:57:25

Recoveries for three of the eighteen compounds in the MDL Standard and the LCS Standard were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges.