


User: Bohlk, Ted
Westchester Dept. of Labs & Research
Date: 10/27/2001 Time: 16:41:22

Sample: AD23838
Analysis: \$525 Organic Chemicals - 525.2
Units: ug
Started: 10/9/01 16:05 Ended: 10/17/01 16:05 Analyst: TB
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Hexachlorocyclopentadiene		< MDL		0.10
2	Hexachlorobenzene		< MDL		0.10
3	Simazine		< MDL		0.40
4	Atrazine		< MDL		0.10
5	Alachlor		< MDL		0.20
6	bis(2-Ethylhexyl)adipate		< MDL		0.60
7	bis(2-Ethylhexyl)phthalate		1.1		0.60
8	Benzo(a)pyrene		< MDL		0.20
9	EPTC		< MDL		1.0
10	2,6-Dinitrotoluene		< MDL		2.0
11	2,4-Dinitrotoluene		< MDL		2.0
12	Molinate		< MDL		0.90
13	Terbacil		< MDL		2.0
14	Acetochlor		< MDL		2.0
15	Metribuzin		< MDL		0.15
16	Metolachlor		< MDL		0.75
17	Butachlor		< MDL		0.40
18	4,4-DDE		< MDL		0.80
19	Surr_1,3-Dimethyl-2-nitrobenze		5.2		1.0
20	Surr_Triphenyl phosphate		5.4		1.0
21	Surr_Perylene-d12		4.3		1.0

Chromatogram Plot File: C:\23838PP Date: Oct-18-1991 11:14:08
Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23838
Scan No: 1 Retention Time: 0:01 RIC: 0 Mass Range: 0 - 0
Plotted: 1 to 1860 Range: 1 to 1860 100% = 10446952



Integration Report Quanfile: 23838PP Quan Entries: 15
 Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23838
 Sorted via: Entry Number ↑

No	Name of Compound	R Time	Scan#	Mode	Peak Area	Pk Height
1	acenaphthene d-10(IS)	12:31	751	Auto	2,054,558	1,166,250
2	phenanthrene d-10 (IS)	17:14	1034	Auto	3,387,757	1,534,796
3	chrysene d-12(IS)	22:56	1376	Auto	2,583,543	1,045,211
4	p-terphenyl d-14(IS)	21:01	1261	Auto	4,415,473	3,217,739
5	acenaphthene d-10(SURR)	12:31	751	Auto	2,054,558	1,166,250
6	phenanthrene d-10(SURR)	17:14	1034	Auto	3,387,757	1,534,796
7	chrysene d-12 (SURR)	22:56	1376	Auto	2,583,543	1,045,211
8	1,3-dimethyl-2-nitrobe	8:10	490	Auto	1,274,508	692,469
9	triphenyl phosphate (S	22:18	1338	Auto	1,999,694	883,877
10	perylene d-12 (SURR)	27:28	1648	Auto	1,379,262	257,999
11	hexachlorocyclopentadi	10:21	621	Auto	1,147	1,147
12	bis(2-ethylhexyl)adipa	22:10	1330	Auto	44,279	8,653
13	bis(2-ethylhexyl)phtha	23:13	1393	Auto	860,111	279,163
14	2,6-dinitrotoluene	12:10	730	Auto	110,910	58,411
15	4,4'-dde	20:56	1256	Auto	2,788	1,701

Quantitation Report

Quanfile: 23838PP

Quan Entries: 15

Comment: BOHLK; METHOD 525; INST 204; 10/17/01; SAMPLE 23838

Sorted via: Entry Number ↑

(S) = Standard

Cal	Name of Compound	S	Scan#	R Time	Me	Calc Amt(A)	Units
1	acenaphthene d-10 (IS)	I	751	12:31	BB	5.000	UG/L
2	phenanthrene d-10 (IS)	I	1034	17:14	BV	5.000	UG/L
3	chrysene d-12 (IS)	I	1376	22:56	BB	5.000	UG/L
4	p-terphenyl d-14 (IS)	I	1261	21:01	BB	5.000	UG/L
5	acenaphthene d-10 (SURR)	A	751	12:31	BB	3.341	UG/L
6	phenanthrene d-10 (SURR)	A	1034	17:14	BV	3.739	UG/L
7	chrysene d-12 (SURR)	A	1376	22:56	BB	4.120	UG/L
8	1,3-dimethyl-2-nitrobenzene	A	490	8:10	BV	5.177	UG/L
9	triphenyl phosphate (S)	A	1338	22:18	BB	5.417	UG/L
10	perylene d-12 (SURR)	A	1648	27:28	BB	4.302	UG/L
11	hexachlorocyclopentadiene	A	621	10:21	BB	0.020	UG/L
16	bis(2-ethylhexyl)adipate	A	1330	22:10	MM	0.104	UG/L
17	bis(2-ethylhexyl)phthalate	A	1393	23:13	BB	1.065	UG/L
20	2,6-dinitrotoluene	A	730	12:10	BV	0.760	UG/L
28	4,4'-dichlorodiphenyl ether	A	1256	20:56	BB	0.010	UG/L

revised IS/surr's

Comments for Sample AD23838 - Analysis \$525

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 11-04-2001 Time: 11:10:24

Recoveries for three of the eighteen compounds in the MDL Standard and the LCS Standard were outside of their acceptance ranges. Recoveries for two of the eighteen compounds in the Spiked Sample and the Spiked Duplicate Sample were outside of their acceptance ranges. The level for bis(2-Ethylhexyl)phthalate in the Laboratory Blank was 0.76 ug/L.