

IT TOOK ~1 hr to check

255 ~~OK~~ good analyses. & update tracking.

AS

9/18/01 Stop messing with a cal curve that meets criteria

Messed up an ok curve and then ran samples

Print out seq. numbers

CCV before restart.

last QC failed.

9/20/01 Good

9/24/01 Initial QC rerun

9/26/01 Good

9/27/01 Initial QC rerun

10/4/01 Initial spike fail

Why is Reporting limit 20 if
IPC is 20 μ g/L

10/9/01 Good

10/11/01 Run a QC after technical problem notes

10/23 Initial QC fail

CCV before restart

Why not running 200.8?
get DL's down