

Sample Information File Untitled

Description : Sb 9/21/01
 Batch ID :
 Volume Units :
 Weight Units :
 Analyst :
 Sample Volume : 1.000
 Nominal Weight : 1.000000

DATE ANALYZED	9/21/01
TIME ANALYZED	2.00
ANALYST INITIALS	MS
REVIEWED BY	AB

AS Sample ID Loc	Sample Sample Weight Units	User Dilution	Remarks
---------------------	-------------------------------	------------------	---------

1 QCS - Fail

2 IPC

3 CB

4 LFB

5 LRB

6 22753 - 98 ✓

7 22753dup

8 22754

9 22760

10 22761

11 22762

12 22913

13 22914

14 22915

15 22916

16 22917 - 99 ✓

17 22918

18 22919

19 22974

20 22975

21 22977

22 22978

23 22979

24 23079

25 23081

26 23082 - 99 ✓

27 23082dup

28 23083

29 23084

30 23085

31 23123

32 23124

33 23125

~~34 23080~~

~~35 23081~~

AU DEP

QC Reviewed by,
 QB 11/7/01

QC PASS

QC FAIL

QC PASS

QC PASS

=====

Element: Sb Seq. No.: 55 AS Loc.: 36 Date: 09/21/2001

Sample ID: Calib Blank

µL dispensed: 10 from 36, 5 from 37, 20 from 36

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.000	0.001	0.013	0.010	0.014	14:30:46	No

Auto-zero performed.

=====

Element: Sb Seq. No.: 56 AS Loc.: 38 Date: 09/21/2001

Sample ID: S5

µL dispensed: 25 from 36, 5 from 37, 5 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.013	0.014	0.022	0.013	0.015	14:33:24	No

Standard number 1 applied. [5]

Correlation Coefficient: 1.0000

Slope: 0.0026

=====

Element: Sb Seq. No.: 57 AS Loc.: 38 Date: 09/21/2001

Sample ID: S10

µL dispensed: 20 from 36, 5 from 37, 10 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.027	0.028	0.039	0.014	0.018	14:36:08	No

Standard number 2 applied. [10]

Correlation Coefficient: 0.9998

Slope: 0.0027

=====

Element: Sb Seq. No.: 58 AS Loc.: 38 Date: 09/21/2001

Sample ID: S15

µL dispensed: 15 from 36, 5 from 37, 15 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.043	0.044	0.044	0.019	0.018	14:38:50	No

Standard number 3 applied. [15]

Correlation Coefficient: 0.9994

Slope: 0.0028

=====

Element: Sb Seq. No.: 59 AS Loc.: 38 Date: 09/21/2001

Sample ID: S20

µL dispensed: 10 from 36, 5 from 37, 20 from 38


Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.058	0.059	0.055	0.016	0.015	14:41:32	No

The calibration curve may not be linear.

Standard number 4 applied. [20]

Correlation Coefficient: 0.9993

Slope: 0.0029



Calibration data for Sb

Standard ID	Mean Signal (Pk Area)	Entered Concentration (µg/L)	Calculated Concentration (µg/L)	Standard Deviation	%RSD
Calib Blank	0.000	----	0.3164	----	----
S5	0.013	5	4.836	----	----
S10	0.027	10	9.648	----	----
S15	0.043	15	14.93	----	----
S20	0.058	20	20.27	----	----
Correlation Coefficient: 0.99934 Slope: 0.00292 Intercept: -0.001					

=====
 Element: Sb Seq. No.: 60 AS Loc.: 1 Date: 09/21/2001
 Sample ID: QCS
 µL dispensed: 10 from 36, 5 from 37, 20 from 1

Repl #	Sample Conc µg/L	Std Conc µg/L	Blk Corr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	12.35	12.35	0.035	0.036	0.036	0.010	0.017	14:44:37	No

QC out of limits. QC will be repeated based on Retry setting.

=====
 Element: Sb Seq. No.: 61 AS Loc.: 1 Date: 09/21/2001
 Sample ID: QCS
 µL dispensed: 10 from 36, 5 from 37, 20 from 1

Repl #	Sample Conc µg/L	Std Conc µg/L	Blk Corr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.067	9.067	0.026	0.026	0.034	0.013	0.019	14:47:14	No

QC value within specified limits.

Method Name: SbHGA Element: Sb
Method Description: Sbhga SM3113 B

Date: 09/21/2001

Technique: Furnace

Wavelength: 217.6 nm

Lamp Current: 0

Sample Info File: Untitled

Calibration Equation: Calc. Intercept: Linear

Slit Width: 0.70 nm

Energy: 55

Results Data Set:

Element: Sb Seq. No.: 62 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.77	10.77	0.030	0.031	0.035	0.011	0.015	14:52:53	No

QC value within specified limits.

Element: Sb Seq. No.: 63 AS Loc.: 3 Date: 09/21/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.9355	0.9355	0.002	0.003	0.013	0.010	0.018	14:55:31	No

Element: Sb Seq. No.: 64 AS Loc.: 4 Date: 09/21/2001

Sample ID: LFB

μL dispensed: 10 from 36, 5 from 37, 20 from 4

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	22.09	22.09	0.063	0.064	0.065	0.013	0.018	14:58:08	No

QC value within specified limits.

Element: Sb Seq. No.: 65 AS Loc.: 5 Date: 09/21/2001

Sample ID: LRB

μL dispensed: 10 from 36, 5 from 37, 20 from 5

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.2586	0.2586	0.000	0.001	0.011	0.011	0.015	15:00:46	No

Element: Sb Seq. No.: 66 AS Loc.: 6 Date: 09/21/2001

Sample ID: 22753

μL dispensed: 10 from 36, 5 from 37, 20 from 6

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.8188	0.8188	0.001	0.002	0.013	0.015	0.016	15:03:25	No

=====
 Element: Sb Seq. No.: 67 AS Loc.: 6 Date: 09/21/2001

Sample ID: 22753

μL dispensed: 5 from 37, 10 from 38, 20 from 6

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	8.432	8.432	0.024	0.025	0.032	0.009	0.014	15:06:01	No

Recovery is 76.1%

=====
 Element: Sb Seq. No.: 68 AS Loc.: 7 Date: 09/21/2001

Sample ID: 22753dup

μL dispensed: 10 from 36, 5 from 37, 20 from 7

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.892	1.892	0.005	0.006	0.016	0.015	0.016	15:08:42	No

=====
 Element: Sb Seq. No.: 69 AS Loc.: 8 Date: 09/21/2001

Sample ID: 22754

μL dispensed: 10 from 36, 5 from 37, 20 from 8

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.4539	0.4539	0.000	0.001	0.015	0.011	0.016	15:11:24	No

=====
 Element: Sb Seq. No.: 70 AS Loc.: 9 Date: 09/21/2001

Sample ID: 22760

μL dispensed: 10 from 36, 5 from 37, 20 from 9

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.2621	-0.2621	-0.002	-0.001	0.016	0.008	0.013	15:14:07	No

=====
 Element: Sb Seq. No.: 71 AS Loc.: 10 Date: 09/21/2001

Sample ID: 22761

μL dispensed: 10 from 36, 5 from 37, 20 from 10

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.0003	0.0003	-0.001	0.000	0.011	0.008	0.015	15:16:51	No

=====
 Element: Sb Seq. No.: 72 AS Loc.: 11 Date: 09/21/2001

Sample ID: 22762

μL dispensed: 10 from 36, 5 from 37, 20 from 11

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.270	-1.270	-0.005	-0.004	0.012	0.012	0.013	15:19:36	No

=====
 Element: Sb Seq. No.: 73 AS Loc.: 12 Date: 09/21/2001

Sample ID: 22913

µL dispensed: 10 from 36, 5 from 37, 20 from 12

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.4196	0.4196	0.000	0.001	0.011	0.017	0.015	15:22:21	No

=====
 Element: Sb Seq. No.: 74 AS Loc.: 13 Date: 09/21/2001

Sample ID: 22914

µL dispensed: 10 from 36, 5 from 37, 20 from 13

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.8848	-0.8848	-0.004	-0.003	0.011	0.018	0.017	15:25:08	No

=====
 Element: Sb Seq. No.: 75 AS Loc.: 14 Date: 09/21/2001

Sample ID: 22915

µL dispensed: 10 from 36, 5 from 37, 20 from 14

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.1605	0.1605	0.000	0.001	0.016	0.017	0.019	15:27:54	No

=====
 Element: Sb Seq. No.: 76 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

µL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.141	9.141	0.026	0.027	0.032	0.015	0.015	15:30:35	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 77 AS Loc.: 3 Date: 09/21/2001

Sample ID: CB

µL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.566	-1.566	-0.005	-0.005	0.012	0.015	0.016	15:33:12	No

=====
 Element: Sb Seq. No.: 78 AS Loc.: 15 Date: 09/21/2001

Sample ID: 22916

µL dispensed: 10 from 36, 5 from 37, 20 from 15

Repl #	SampleConc µg/L	StdConc µg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.8156	0.8156	0.001	0.002	0.013	0.015	0.017	15:35:56	No

=====

Element: Sb Seq. No.: 79 AS Loc.: 16 Date: 09/21/2001

Sample ID: 22917

µL dispensed: 10 from 36, 5 from 37, 20 from 16

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.0832	-0.0832	-0.001	0.000	0.013	0.017	0.019	15:38:39	No

=====

Element: Sb Seq. No.: 80 AS Loc.: 16 Date: 09/21/2001

Sample ID: 22917

µL dispensed: 5 from 37, 10 from 38, 20 from 16

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	7.933	7.933	0.022	0.023	0.031	0.009	0.014	15:41:24	No

Recovery is 79.3%

=====

Element: Sb Seq. No.: 81 AS Loc.: 17 Date: 09/21/2001

Sample ID: 22918

µL dispensed: 10 from 36, 5 from 37, 20 from 17

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.598	1.598	0.004	0.005	0.013	0.017	0.017	15:44:06	No

=====

Element: Sb Seq. No.: 82 AS Loc.: 18 Date: 09/21/2001

Sample ID: 22919

µL dispensed: 10 from 36, 5 from 37, 20 from 18

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.8910	0.8910	0.002	0.003	0.012	0.014	0.015	15:46:47	No

=====

Element: Sb Seq. No.: 83 AS Loc.: 19 Date: 09/21/2001

Sample ID: 22974

µL dispensed: 10 from 36, 5 from 37, 20 from 19

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.250	1.250	0.003	0.004	0.013	0.014	0.015	15:49:27	No

=====

Element: Sb Seq. No.: 84 AS Loc.: 20 Date: 09/21/2001

Sample ID: 22975

µL dispensed: 10 from 36, 5 from 37, 20 from 20

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.9539	-0.9539	-0.004	-0.003	0.012	0.016	0.017	15:52:05	No

=====
 Element: Sb Seq. No.: 85 AS Loc.: 21 Date: 09/21/2001

Sample ID: 22977

μL dispensed: 10 from 36, 5 from 37, 20 from 21

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.2090	0.2090	0.000	0.001	0.016	0.018	0.020	15:54:43	No

=====
 Element: Sb Seq. No.: 86 AS Loc.: 22 Date: 09/21/2001

Sample ID: 22978

μL dispensed: 10 from 36, 5 from 37, 20 from 22

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.4405	0.4405	0.000	0.001	0.014	0.013	0.017	15:57:20	No

=====
 Element: Sb Seq. No.: 87 AS Loc.: 23 Date: 09/21/2001

Sample ID: 22979

μL dispensed: 10 from 36, 5 from 37, 20 from 23

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.146	-1.146	-0.004	-0.003	0.013	0.016	0.019	15:59:58	No

=====
 Element: Sb Seq. No.: 88 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	8.817	8.817	0.025	0.026	0.035	0.014	0.018	16:02:36	No

QC out of limits. QC will be repeated based on Retry setting.

=====
 Element: Sb Seq. No.: 89 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.26	10.26	0.029	0.030	0.038	0.015	0.016	16:05:14	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 90 AS Loc.: 3 Date: 09/21/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.2657	-0.2657	-0.002	-0.001	0.014	0.006	0.014	16:07:51	No

=====
 Element: Sb Seq. No.: 91 AS Loc.: 24 Date: 09/21/2001

Sample ID: 23079

µL dispensed: 10 from 36, 5 from 37, 20 from 24

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.6665	0.6665	0.001	0.002	0.012	0.015	0.016	16:10:29	No

=====
 Element: Sb Seq. No.: 92 AS Loc.: 25 Date: 09/21/2001

Sample ID: 23081

µL dispensed: 10 from 36, 5 from 37, 20 from 25

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.7134	-0.7134	-0.003	-0.002	0.014	0.017	0.019	16:13:08	No

=====
 Element: Sb Seq. No.: 93 AS Loc.: 26 Date: 09/21/2001

Sample ID: 23082

µL dispensed: 10 from 36, 5 from 37, 20 from 26

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.7604	-0.7604	-0.003	-0.002	0.013	0.020	0.018	16:15:49	No

=====
 Element: Sb Seq. No.: 94 AS Loc.: 26 Date: 09/21/2001

Sample ID: 23082

µL dispensed: 5 from 37, 10 from 38, 20 from 26

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	7.281	7.281	0.020	0.021	0.035	0.012	0.018	16:18:26	No

Recovery is 72.8%

=====
 Element: Sb Seq. No.: 95 AS Loc.: 27 Date: 09/21/2001

Sample ID: 23082dup

µL dispensed: 10 from 36, 5 from 37, 20 from 27

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.7606	0.7606	0.001	0.002	0.012	0.019	0.019	16:21:09	No

=====
 Element: Sb Seq. No.: 96 AS Loc.: 28 Date: 09/21/2001

Sample ID: 23083

µL dispensed: 10 from 36, 5 from 37, 20 from 28

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.6673	-0.6673	-0.003	-0.002	0.015	0.019	0.019	16:23:54	No

Element: Sb Seq. No.: 97 AS Loc.: 29 Date: 09/21/2001

Sample ID: 23084

μL dispensed: 10 from 36, 5 from 37, 20 from 29

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.2536	-0.2536	-0.002	-0.001	0.014	0.018	0.017	16:26:39	No

Element: Sb Seq. No.: 98 AS Loc.: 30 Date: 09/21/2001

Sample ID: 23085

μL dispensed: 10 from 36, 5 from 37, 20 from 30

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.017	-1.017	-0.004	-0.003	0.011	0.020	0.017	16:29:26	No

Element: Sb Seq. No.: 99 AS Loc.: 31 Date: 09/21/2001

Sample ID: 23123

μL dispensed: 10 from 36, 5 from 37, 20 from 31

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.1751	0.1751	0.000	0.001	0.015	0.014	0.017	16:32:12	No

Element: Sb Seq. No.: 100 AS Loc.: 32 Date: 09/21/2001

Sample ID: 23124

μL dispensed: 10 from 36, 5 from 37, 20 from 32

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.237	1.237	0.003	0.004	0.016	0.015	0.014	16:34:56	No

Element: Sb Seq. No.: 101 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.914	9.914	0.028	0.029	0.041	0.011	0.018	16:37:36	No

QC value within specified limits.

Element: Sb Seq. No.: 102 AS Loc.: 3 Date: 09/21/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.5017	-0.5017	-0.002	-0.001	0.012	0.010	0.018	16:40:14	No

=====
 Element: Sb Seq. No.: 103 AS Loc.: 33 Date: 09/21/2001

Sample ID: 23125

μL dispensed: 10 from 36, 5 from 37, 20 from 33

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.8811	0.8811	0.002	0.003	0.014	0.020	0.018	16:42:55	No

=====
 Element: Sb Seq. No.: 104 AS Loc.: 2 Date: 09/21/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.421	9.421	0.027	0.028	0.042	0.013	0.017	16:45:34	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 105 AS Loc.: 3 Date: 09/21/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.310	-1.310	-0.005	-0.004	0.013	0.013	0.014	16:48:11	No