

=====

Element: Sb Seq. No.: 68 AS Loc.: 36 Date: 10/02/2001

Sample ID: Calib Blank

µL dispensed: 10 from 36, 5 from 37, 20 from 36

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			-0.001	0.001	0.015	-0.031	0.018	14:52:56	No

Auto-zero performed. ✓

=====

Element: Sb Seq. No.: 69 AS Loc.: 38 Date: 10/02/2001

Sample ID: S5

µL dispensed: 25 from 36, 5 from 37, 5 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.022	0.023	0.037	-0.032	0.014	14:55:33	No

Standard number 1 applied. [5]

Correlation Coefficient: 1.0000 ✓

Slope: 0.0044

=====

Element: Sb Seq. No.: 70 AS Loc.: 38 Date: 10/02/2001

Sample ID: S10

µL dispensed: 20 from 36, 5 from 37, 10 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.037	0.038	0.049	-0.030	0.016	14:58:15	No

Standard number 2 applied. [10]

Correlation Coefficient: 0.9948 ✓

Slope: 0.0037

=====

Element: Sb Seq. No.: 71 AS Loc.: 38 Date: 10/02/2001

Sample ID: S15

µL dispensed: 15 from 36, 5 from 37, 15 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.048	0.049	0.060	-0.022	0.015	15:00:58	No

Standard number 3 applied. [15]

Correlation Coefficient: 0.9880 ✓

Slope: 0.0032

=====

Element: Sb Seq. No.: 72 AS Loc.: 38 Date: 10/02/2001

Sample ID: S20

µL dispensed: 10 from 36, 5 from 37, 20 from 38


Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.069	0.069	0.072	-0.026	0.014	15:03:41	No

The calibration curve may not be linear. ✓

Standard number 4 applied. [20]

Correlation Coefficient: 0.9940 ✓

Slope: 0.0033



Calibration data for Sb

Standard ID	Mean Signal (Pk Area)	Entered Concentration (µg/L)	Calculated Concentration (µg/L)	Standard Deviation	%RSD
Calib Blank	0.000	----	-0.7784	----	----
S5	0.022	5	5.951	----	----
S10	0.037	10	10.65	----	----
S15	0.048	15	13.97	----	----
S20	0.069	20	20.21	----	----
Correlation Coefficient: 0.99396 Slope: 0.00327 Intercept: 0.003					

The calibration curve has been edited.

=====

Element: Sb Seq. No.: 73 AS Loc.: 38 Date: 10/02/2001

Sample ID: S5

µL dispensed: 25 from 36, 5 from 37, 5 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.024	0.024	0.026	-0.026	0.016	15:06:57	No
Standard number 1 applied. [5]									
Correlation Coefficient: 0.9947					Slope: 0.0033				

=====
 Element: Sb Seq. No.: 74 AS Loc.: 38 Date: 10/02/2001

Sample ID: S15

µL dispensed: 15 from 36, 5 from 37, 15 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.051	0.052	0.049	-0.023	0.013	15:09:41	No

Standard number 3 applied. [15]

Correlation Coefficient: 0.9946

Slope: 0.0033

The calibration curve has been edited.

=====
 Element: Sb Seq. No.: 75 AS Loc.: 38 Date: 10/02/2001

Sample ID: S5

µL dispensed: 25 from 36, 5 from 37, 5 from 38

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.015	0.015	0.023	-0.024	0.016	15:13:30	No

Standard number 1 applied. [5]

Correlation Coefficient: 0.9975

Slope: 0.0035

=====
 Element: Sb Seq. No.: 76 AS Loc.: 1 Date: 10/02/2001

Sample ID: QCS

µL dispensed: 15 from 36, 5 from 37, 20 from 1

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	45.58	45.58	0.158	0.159	0.142	-0.016	0.022	15:16:20	No

Sample absorbance is greater than that of the highest standard.

QC out of limits. QC will be repeated based on Retry setting.

=====
 Element: Sb Seq. No.: 77 AS Loc.: 1 Date: 10/02/2001

Sample ID: QCS

µL dispensed: 10 from 36, 5 from 37, 20 from 1

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	45.61	45.61	0.158	0.159	0.145	-0.016	0.020	15:18:58	No

Sample absorbance is greater than that of the highest standard.

Remake QC

wrong dilution
- no

=====
 Element: Sb Seq. No.: 77 AS Loc.: 1 Date: 10/02/2001

Sample ID: QCS

μL dispensed: 10 from 36, 5 from 37, 20 from 1

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.83	10.83	0.037	0.038	0.038	-0.026	0.016	15:24:15	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 78 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.925	9.925	0.034	0.035	0.036	-0.027	0.014	15:26:52	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 79 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.2168	-0.2168	-0.001	0.000	0.012	-0.029	0.012	15:29:29	No

=====
 Element: Sb Seq. No.: 80 AS Loc.: 4 Date: 10/02/2001

Sample ID: LFB

μL dispensed: 10 from 36, 5 from 37, 20 from 4

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	20.06	20.06	0.069	0.070	0.060	-0.028	0.014	15:32:07	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 81 AS Loc.: 5 Date: 10/02/2001

Sample ID: LRB

μL dispensed: 10 from 36, 5 from 37, 20 from 5

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.1150	-0.1150	-0.001	0.000	0.015	-0.030	0.010	15:34:45	No

=====
 Element: Sb Seq. No.: 82 AS Loc.: 6 Date: 10/02/2001

Sample ID: 23023

μL dispensed: 10 from 36, 5 from 37, 20 from 6

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.7625	0.7625	0.002	0.003	0.017	-0.014	0.018	15:37:24	No

=====
 Element: Sb Seq. No.: 83 AS Loc.: 6 Date: 10/02/2001

Sample ID: 23023

µL dispensed: 5 from 37, 10 from 38, 20 from 6

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	7.489	7.489	0.026	0.026	0.033	-0.027	0.009	15:40:01	No

Recovery is 67.3% (outside of specified limits)

=====
 Element: Sb Seq. No.: 84 AS Loc.: 6 Date: 10/02/2001

Sample ID: 23023

µL dispensed: 25 from 36, 5 from 37, 5 from 6

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	3.981	0.9953	0.003	0.004	0.017	-0.029	0.011	15:42:40	No

=====
 Element: Sb Seq. No.: 85 AS Loc.: 6 Date: 10/02/2001

Sample ID: 23023

µL dispensed: 15 from 36, 5 from 37, 10 from 38, 5 from 6

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	30.50	7.625	0.026	0.027	0.027	-0.022	0.019	15:45:26	No

Recovery is 66.3% (outside of specified limits)

=====
 Element: Sb Seq. No.: 86 AS Loc.: 7 Date: 10/02/2001

Sample ID: 23025

µL dispensed: 10 from 36, 5 from 37, 20 from 7

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.9549	0.9549	0.003	0.004	0.018	-0.002	0.019	15:48:07	No

=====
 Element: Sb Seq. No.: 87 AS Loc.: 8 Date: 10/02/2001

Sample ID: 23833

µL dispensed: 10 from 36, 5 from 37, 20 from 8

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.533	1.533	0.005	0.006	0.019	-0.026	0.014	15:50:50	No

=====
 Element: Sb Seq. No.: 88 AS Loc.: 9 Date: 10/02/2001

Sample ID: 23834

µL dispensed: 10 from 36, 5 from 37, 20 from 9

Repl #	SampleConc µg/L	StdndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.1180	-0.1180	-0.001	0.000	0.014	-0.023	0.017	15:53:33	No

=====
 Element: Sb Seq. No.: 89 AS Loc.: 10 Date: 10/02/2001

Sample ID: 23835

μL dispensed: 10 from 36, 5 from 37, 20 from 10

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.432	1.432	0.005	0.005	0.012	-0.023	0.011	15:56:17	No

=====
 Element: Sb Seq. No.: 90 AS Loc.: 11 Date: 10/02/2001

Sample ID: 23836

μL dispensed: 10 from 36, 5 from 37, 20 from 11

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.852	1.852	0.006	0.007	0.016	-0.025	0.009	15:59:02	No

=====
 Element: Sb Seq. No.: 91 AS Loc.: 12 Date: 10/02/2001

Sample ID: 23837

μL dispensed: 10 from 36, 5 from 37, 20 from 12

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.3818	-0.3818	-0.002	-0.001	0.013	-0.021	0.013	16:01:47	No

=====
 Element: Sb Seq. No.: 92 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	8.009	8.009	0.027	0.028	0.031	-0.026	0.014	16:04:30	No

QC out of limits. QC will be repeated based on Retry setting.

=====
 Element: Sb Seq. No.: 93 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.836	9.836	0.034	0.035	0.041	-0.029	0.013	16:07:08	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 94 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	2.127	2.127	0.007	0.008	0.014	-0.034	0.012	16:09:45	No

=====
 Element: Sb Seq. No.: 95 AS Loc.: 13 Date: 10/02/2001

Sample ID: 23838

μL dispensed: 10 from 36, 5 from 37, 20 from 13

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.5283	0.5283	0.001	0.002	0.013	-0.024	0.010	16:12:27	No

=====
 Element: Sb Seq. No.: 96 AS Loc.: 14 Date: 10/02/2001

Sample ID: 23839

μL dispensed: 10 from 36, 5 from 37, 20 from 14

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.5627	-0.5627	-0.002	-0.002	0.018	-0.023	0.012	16:15:13	No

=====
 Element: Sb Seq. No.: 97 AS Loc.: 15 Date: 10/02/2001

Sample ID: 23900

μL dispensed: 10 from 36, 5 from 37, 20 from 15

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.1184	0.1184	0.000	0.001	0.013	-0.023	0.013	16:17:57	No

=====
 Element: Sb Seq. No.: 98 AS Loc.: 16 Date: 10/02/2001

Sample ID: 23901

μL dispensed: 10 from 36, 5 from 37, 20 from 16

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.2543	0.2543	0.000	0.001	0.015	-0.030	0.011	16:20:41	No

=====
 Element: Sb Seq. No.: 99 AS Loc.: 16 Date: 10/02/2001

Sample ID: 23901

μL dispensed: 5 from 37, 10 from 38, 20 from 16

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	6.224	6.224	0.021	0.022	0.032	-0.035	0.009	16:23:25	No

Recovery is 59.7% (outside of specified limits)

=====
 Element: Sb Seq. No.: 100 AS Loc.: 16 Date: 10/02/2001

Sample ID: 23901

μL dispensed: 25 from 36, 5 from 37, 5 from 16

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.822	0.4554	0.001	0.002	0.015	-0.029	0.012	16:26:07	No

=====

Element: Sb Seq. No.: 101 AS Loc.: 16 Date: 10/02/2001

Sample ID: 23901

μL dispensed: 15 from 36, 5 from 37, 10 from 38, 5 from 16

Repl #	Sample Conc μg/L	StdConc μg/L	BlnkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	23.87	5.966	0.020	0.021	0.031	-0.023	0.012	16:28:56	No

Repair 20 Std. for spike

=====

Element: Sb Seq. No.: 101 AS Loc.: 1 Date: 10/02/2001

Sample ID: QCS

μL dispensed: 10 from 36, 5 from 37, 20 from 1

Repl	SampleConc	StndConc	BlnkCorr	Peak	Peak	Bkgnd	Bkgnd	Time	Peak
#	μg/L	μg/L	Signal	Area	Height	Area	Height		Stored
1	9.042	9.042	0.031	0.032	0.043	-0.027	0.012	16:31:52	No

QC value within specified limits.

Method Name: SbHGA Element: Sb
 Method Description: Sbhga SM3113 B

Date: 10/02/2001
 Technique: Furnace Calibration Equation: Calc. Intercept: Linear
 Wavelength: 217.6 nm Slit Width: 0.70 nm
 Lamp Current: 0 Energy: 55
 Sample Info File: Untitled Results Data Set:

Element: Sb Seq. No.: 102 AS Loc.: 16 Date: 10/02/2001
 Sample ID: 23901
 µL dispensed: 10 from 36, 5 from 37, 20 from 16

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.9251	0.9251	0.003	0.004	0.015	-0.028	0.015	16:35:09	No

Element: Sb Seq. No.: 103 AS Loc.: 16 Date: 10/02/2001
 Sample ID: 23901
 µL dispensed: 5 from 37, 10 from 38, 20 from 16

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.52	10.52	0.036	0.037	0.044	-0.034	0.010	16:37:52	No

Recovery is 96.0%

Element: Sb Seq. No.: 104 AS Loc.: 17 Date: 10/02/2001
 Sample ID: 23902
 µL dispensed: 10 from 36, 5 from 37, 20 from 17

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	2.531	2.531	0.008	0.009	0.013	-0.029	0.012	16:40:35	No

Element: Sb Seq. No.: 105 AS Loc.: 18 Date: 10/02/2001
 Sample ID: 23903
 µL dispensed: 10 from 36, 5 from 37, 20 from 18

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.9511	0.9511	0.003	0.004	0.015	-0.022	0.009	16:43:17	No

Element: Sb Seq. No.: 106 AS Loc.: 19 Date: 10/02/2001
 Sample ID: 23904
 µL dispensed: 10 from 36, 5 from 37, 20 from 19

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.3077	0.3077	0.001	0.001	0.015	-0.022	0.017	16:45:57	No

=====
 Element: Sb Seq. No.: 107 AS Loc.: 20 Date: 10/02/2001

Sample ID: 23905

µL dispensed: 10 from 36, 5 from 37, 20 from 20

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.9636	0.9636	0.003	0.004	0.014	-0.022	0.011	16:48:36	No

=====
 Element: Sb Seq. No.: 108 AS Loc.: 21 Date: 10/02/2001

Sample ID: 23906

µL dispensed: 10 from 36, 5 from 37, 20 from 21

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.7211	0.7211	0.002	0.003	0.014	-0.023	0.013	16:51:14	No

=====
 Element: Sb Seq. No.: 109 AS Loc.: 22 Date: 10/02/2001

Sample ID: 23909

µL dispensed: 10 from 36, 5 from 37, 20 from 22

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.430	1.430	0.005	0.005	0.014	0.005	0.121	16:53:51	No

=====
 Element: Sb Seq. No.: 110 AS Loc.: 23 Date: 10/02/2001

Sample ID: 23910

µL dispensed: 10 from 36, 5 from 37, 20 from 23

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.1012	-0.1012	-0.001	0.000	0.013	-0.024	0.012	16:56:28	No

=====
 Element: Sb Seq. No.: 111 AS Loc.: 24 Date: 10/02/2001

Sample ID: 23913

µL dispensed: 10 from 36, 5 from 37, 20 from 24

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.4714	0.4714	0.001	0.002	0.013	-0.025	0.012	16:59:05	No

=====
 Element: Sb Seq. No.: 112 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

µL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	8.607	8.607	0.030	0.030	0.039	-0.025	0.012	17:01:46	No

QC out of limits. QC will be repeated based on Retry setting.

=====
 Element: Sb Seq. No.: 113 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.73	10.73	0.037	0.038	0.043	-0.029	0.010	17:04:23	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 114 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.688	1.688	0.005	0.006	0.013	-0.032	0.012	17:07:01	No

=====
 Element: Sb Seq. No.: 115 AS Loc.: 25 Date: 10/02/2001

Sample ID: 23914

μL dispensed: 10 from 36, 5 from 37, 20 from 25

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.151	1.151	0.004	0.004	0.012	-0.024	0.010	17:09:38	No

=====
 Element: Sb Seq. No.: 116 AS Loc.: 26 Date: 10/02/2001

Sample ID: 23915

μL dispensed: 10 from 36, 5 from 37, 20 from 26

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.6207	0.6207	0.002	0.003	0.015	-0.020	0.016	17:12:19	No

=====
 Element: Sb Seq. No.: 117 AS Loc.: 26 Date: 10/02/2001

Sample ID: 23915

μL dispensed: 5 from 37, 10 from 38, 20 from 26

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	10.91	10.91	0.038	0.038	0.041	-0.023	0.013	17:14:56	No

Recovery is 102.9%

=====
 Element: Sb Seq. No.: 118 AS Loc.: 27 Date: 10/02/2001

Sample ID: 23915dup

μL dispensed: 10 from 36, 5 from 37, 20 from 27

Repl #	SampleConc μg/L	StndConc μg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.783	1.783	0.006	0.007	0.014	-0.020	0.017	17:17:39	No

=====
 Element: Sb Seq. No.: 119 AS Loc.: 28 Date: 10/02/2001

Sample ID: 23947

μL dispensed: 10 from 36, 5 from 37, 20 from 28

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-1.221	-1.221	-0.005	-0.004	0.014	-0.018	0.015	17:20:24	No

=====
 Element: Sb Seq. No.: 120 AS Loc.: 29 Date: 10/02/2001

Sample ID: 23948

μL dispensed: 10 from 36, 5 from 37, 20 from 29

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.256	1.256	0.004	0.005	0.017	-0.022	0.013	17:23:10	No

=====
 Element: Sb Seq. No.: 121 AS Loc.: 30 Date: 10/02/2001

Sample ID: 23993

μL dispensed: 10 from 36, 5 from 37, 20 from 30

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.3965	-0.3965	-0.002	-0.001	0.010	-0.020	0.012	17:25:57	No

=====
 Element: Sb Seq. No.: 122 AS Loc.: 31 Date: 10/02/2001

Sample ID: 23994

μL dispensed: 10 from 36, 5 from 37, 20 from 31

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.162	1.162	0.004	0.004	0.014	-0.022	0.018	17:28:43	No

=====
 Element: Sb Seq. No.: 123 AS Loc.: 32 Date: 10/02/2001

Sample ID: 23995

μL dispensed: 10 from 36, 5 from 37, 20 from 32

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.0774	-0.0774	-0.001	0.000	0.012	-0.016	0.015	17:31:26	No

=====
 Element: Sb Seq. No.: 124 AS Loc.: 33 Date: 10/02/2001

Sample ID: 23996

μL dispensed: 10 from 36, 5 from 37, 20 from 33

Repl #	SampleConc μg/L	StndConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.7207	0.7207	0.002	0.003	0.014	-0.017	0.015	17:34:11	No

=====
 Element: Sb Seq. No.: 125 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.760	9.760	0.034	0.034	0.043	-0.022	0.013	17:36:49	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 126 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.2212	0.2212	0.000	0.001	0.012	-0.026	0.010	17:39:27	No

=====
 Element: Sb Seq. No.: 127 AS Loc.: 34 Date: 10/02/2001

Sample ID: 23997

μL dispensed: 10 from 36, 5 from 37, 20 from 34

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	0.6385	0.6385	0.002	0.003	0.013	-0.021	0.017	17:42:07	No

=====
 Element: Sb Seq. No.: 128 AS Loc.: 35 Date: 10/02/2001

Sample ID: 23998

μL dispensed: 10 from 36, 5 from 37, 20 from 35

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-0.3260	-0.3260	-0.002	-0.001	0.010	-0.020	0.014	17:44:46	No

=====
 Element: Sb Seq. No.: 129 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

μL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	9.310	9.310	0.032	0.033	0.052	-0.021	0.010	17:47:23	No

QC value within specified limits.

=====
 Element: Sb Seq. No.: 130 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

μL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc μg/L	StdConc μg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	1.019	1.019	0.003	0.004	0.012	-0.028	0.012	17:50:00	No