

T1

10/10/01 - no seq # - no time printed

Print all sequences

Rerun cal standards - Rules?

9/20/01 Rerun spike 3x then passed on 4th ?!

ILC failed 3x then passed on 4th !!!

9/20/01 Run is HORRIBLE!

9/24/01 Bad calibration

LET THE INSTRUMENT WARM UP

10/2/01 If go back to instrument - do CCU 1st

EVERY RUN - Analyst says will remake fresh standard. Does he?

10/2/3:28 Stinks. - rerun 10/3/01