

Sample Information File Untitled

Description : T1F POTABLE
 Batch ID :
 Volume Units :
 Weight Units :
 Analyst :
 Sample Volume : 1.000
 Nominal Weight : 1.000000

AS Sample ID Loc	Sample Sample Weight Units	User Dilution	Remarks
---------------------	-------------------------------	------------------	---------

1	QCS		
2	IPC		
3	CB		
4	LFB		
5	LRB		
6	24214		
7	24214DUP		
8	24215		
9	24216		
10	24217		
11	24218		
12	24219		
13	24220		
14	24236		
15	24237		
16	24238		
17	24268		NOT DET
18	24287		
19	24288		
20	24289		
21	24290		
22	24291		
23	24292		
24	24293		
25	24333		
26	24334		
27	24334DUP		
28	24335		
29	24395		
30	24396		
31	24397		
32	24398		
33	24399		
34	24400		
35	24401		

batches 14824-26

QC Reviewed by,
[Signature] 10/16/21

DATE ANALYZED	10/16/21
TIME ANALYZED	AM?
ANALYST INITIALS	LD
REVIEWED BY	see notebook

OK
 still fighting
 w/ calibration

Method Name: TlHGAlow Element: Tl
Method Description: Tlhga EPA 200.9

Date: 10/10/2001
Technique: Furnace Calibration Equation: Calc. Intercept: Linear
Wavelength: 276.8 nm Slit Width: 0.70 nm
Lamp Current: 6 Energy: 52
Sample Info File: Untitled Results Data Set:

Element: Tl AS Loc.: 36 Date: 10/10/2001
Sample ID: Calib Blank
µL dispensed: 10 from 36, 5 from 37, 20 from 36
µL dispensed: 10 from 36, 5 from 37, 20 from 36

Repl #	SampleConc µg/L	Blncorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		-0.008	-0.002	-0.002	09:40:09	No

Auto-zero performed.

Element: Tl AS Loc.: 38 Date: 10/10/2001
Sample ID: S2
µL dispensed: 26 from 36, 5 from 37, 4 from 38
µL dispensed: 26 from 36, 5 from 37, 4 from 38

Repl #	SampleConc µg/L	Blncorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		0.018	0.016	0.016	09:44:22	No

Standard number 1 applied. [2]
Correlation Coefficient: 1.0000 Slope: 0.0091

Element: Tl AS Loc.: 38 Date: 10/10/2001
Sample ID: S4
µL dispensed: 22 from 36, 5 from 37, 8 from 38
µL dispensed: 22 from 36, 5 from 37, 8 from 38

Repl #	SampleConc µg/L	Blncorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		0.038	0.036	0.027	09:48:44	No

Standard number 2 applied. [4]
Correlation Coefficient: 0.9997 Slope: 0.0095

Element: Tl AS Loc.: 38 Date: 10/10/2001
Sample ID: S6
µL dispensed: 18 from 36, 5 from 37, 12 from 38
µL dispensed: 18 from 36, 5 from 37, 12 from 38

Repl #	SampleConc µg/L	Blncorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		0.056	0.054	0.043	09:53:03	No

Standard number 3 applied. [6]
Correlation Coefficient: 0.9998 Slope: 0.0094

=====

Element: Tl AS Loc.: 38 Date: 10/10/2001
 Sample ID: S8
 µL dispensed: 14 from 36, 5 from 37, 16 from 38
 µL dispensed: 14 from 36, 5 from 37, 16 from 38

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		0.069	0.067	0.058	09:57:21	No

Standard number 4 applied. [8]
 Correlation Coefficient: 0.9979 Slope: 0.0088


=====

Element: Tl AS Loc.: 38 Date: 10/10/2001
 Sample ID: S10
 µL dispensed: 10 from 36, 5 from 37, 20 from 38
 µL dispensed: 10 from 36, 5 from 37, 20 from 38

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1		0.076	0.073	0.071	10:01:41	No

The calibration curve may not be linear.
 Standard number 5 applied. [10]
 Correlation Coefficient: 0.9888 Slope: 0.0079

X



Calibration data for Tl

Standard ID	Mean Signal (Pk Area)	Entered Concentration (µg/L)	Calculated Concentration (µg/L)	Standard Deviation	%RSD
Calib Blank	0.000	----	-0.453	----	----
S2	0.018	2	1.859	----	----
S4	0.038	4	4.385	----	----
S6	0.056	6	6.654	----	----
S8	0.069	8	8.366	----	----
S10	0.076	10	9.189	----	----

Correlation Coefficient: 0.98882 Slope: 0.00785 Intercept: 0.004

X

The calibration curve has been edited.

=====
Element: T1 AS Loc.: 1 Date: 10/10/2001
Sample ID: QCS
µL dispensed: 10 from 36, 5 from 37, 20 from 1
µL dispensed: 10 from 36, 5 from 37, 20 from 1
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.272	0.039	0.036	0.037	10:50:11	No

QC out of limits. QC will be repeated based on Retry setting.

=====
Element: T1 AS Loc.: 1 Date: 10/10/2001
Sample ID: QCS
µL dispensed: 10 from 36, 5 from 37, 20 from 1
µL dispensed: 10 from 36, 5 from 37, 20 from 1
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.643	0.042	0.040	0.033	10:54:21	No

QC value within specified limits.

=====
Element: T1 AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.018	0.036	0.034	0.033	10:58:31	No

QC out of limits. QC will be repeated based on Retry setting.

=====
Element: T1 AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.077	0.037	0.035	0.034	11:02:43	No

QC out of limits. Current analytical sequence continued.

Method Name: TlHGAlow Element: Tl
Method Description: Tlhga EPA 200.9

Date: 10/10/2001
Technique: Furnace Calibration Equation: Calc. Intercept: Linear
Wavelength: 276.8 nm Slit Width: 0.70 nm
Lamp Current: 6 Energy: 52
Sample Info File: Untitled Results Data Set:

Element: Tl AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.073	0.046	0.043	0.042	11:07:58	No

QC value within specified limits.

Element: Tl AS Loc.: 3 Date: 10/10/2001
Sample ID: CB
µL dispensed: 10 from 36, 5 from 37, 20 from 3
µL dispensed: 10 from 36, 5 from 37, 20 from 3

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.374	0.004	0.002	-0.005	11:12:07	No

Element: Tl AS Loc.: 4 Date: 10/10/2001
Sample ID: LFB
µL dispensed: 10 from 36, 5 from 37, 20 from 4
µL dispensed: 10 from 36, 5 from 37, 20 from 4

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	8.611	0.077	0.075	0.071	11:16:18	No

Sample absorbance is greater than that of the highest standard.
QC value within specified limits.

Element: Tl AS Loc.: 5 Date: 10/10/2001
Sample ID: LRB
µL dispensed: 10 from 36, 5 from 37, 20 from 5
µL dispensed: 10 from 36, 5 from 37, 20 from 5

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.106	0.000	-0.002	-0.003	11:20:28	No

Element: Tl AS Loc.: 6 Date: 10/10/2001
Sample ID: 24214
µL dispensed: 10 from 36, 5 from 37, 20 from 6
µL dispensed: 10 from 36, 5 from 37, 20 from 6

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.271	-0.001	-0.004	0.037	11:24:40	No

=====
Element: Tl AS Loc.: 6 Date: 10/10/2001
Sample ID: 24214
µL dispensed: 5 from 37, 10 from 38, 20 from 6
µL dispensed: 5 from 37, 10 from 38, 20 from 6
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.188	0.047	0.044	0.072	11:28:47	No

Recovery is 103.8% ✓

=====
Element: Tl AS Loc.: 7 Date: 10/10/2001
Sample ID: 24214DUP
µL dispensed: 10 from 36, 5 from 37, 20 from 7
µL dispensed: 10 from 36, 5 from 37, 20 from 7
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.134	0.000	-0.002	0.036	11:33:04	No

=====
Element: Tl AS Loc.: 8 Date: 10/10/2001
Sample ID: 24215
µL dispensed: 10 from 36, 5 from 37, 20 from 8
µL dispensed: 10 from 36, 5 from 37, 20 from 8
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.257	-0.001	-0.004	0.035	11:37:22	No

=====
Element: Tl AS Loc.: 9 Date: 10/10/2001
Sample ID: 24216
µL dispensed: 10 from 36, 5 from 37, 20 from 9
µL dispensed: 10 from 36, 5 from 37, 20 from 9
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.083	0.000	-0.002	0.060	11:41:45	No

=====
Element: Tl AS Loc.: 10 Date: 10/10/2001
Sample ID: 24217
µL dispensed: 10 from 36, 5 from 37, 20 from 10
µL dispensed: 10 from 36, 5 from 37, 20 from 10
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.182	-0.001	-0.003	0.038	11:46:10	No

=====
Element: Tl AS Loc.: 11 Date: 10/10/2001
Sample ID: 24218
µL dispensed: 10 from 36, 5 from 37, 20 from 11
µL dispensed: 10 from 36, 5 from 37, 20 from 11
=====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.053	0.001	-0.002	0.054	11:50:33	No

=====
Element: Tl AS Loc.: 12 Date: 10/10/2001
Sample ID: 24219
µL dispensed: 10 from 36, 5 from 37, 20 from 12
µL dispensed: 10 from 36, 5 from 37, 20 from 12
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.279	-0.001	-0.004	0.039	11:54:59	No

=====

=====
Element: Tl AS Loc.: 13 Date: 10/10/2001
Sample ID: 24220
µL dispensed: 10 from 36, 5 from 37, 20 from 13
µL dispensed: 10 from 36, 5 from 37, 20 from 13
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.249	0.003	0.001	0.034	11:59:23	No

=====

=====
Element: Tl AS Loc.: 14 Date: 10/10/2001
Sample ID: 24236
µL dispensed: 10 from 36, 5 from 37, 20 from 14
µL dispensed: 10 from 36, 5 from 37, 20 from 14
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.130	0.000	-0.002	0.018	12:03:47	No

=====

=====
Element: Tl AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.721	0.051	0.049	0.039	12:08:02	No

QC out of limits. QC will be repeated based on Retry setting.

=====

=====
Element: Tl AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.881	0.044	0.042	0.038	12:12:12	No

QC value within specified limits.

=====

=====
Element: Tl AS Loc.: 3 Date: 10/10/2001
Sample ID: CB
µL dispensed: 10 from 36, 5 from 37, 20 from 3
µL dispensed: 10 from 36, 5 from 37, 20 from 3
=====

Repl #	SampleConc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.194	-0.001	-0.003	0.001	12:16:22	No

=====

=====
 Element: Tl AS Loc.: 15 Date: 10/10/2001
 Sample ID: 24237
 µL dispensed: 10 from 36, 5 from 37, 20 from 15
 µL dispensed: 10 from 36, 5 from 37, 20 from 15
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.594	-0.004	-0.007	0.258	12:20:46	No

=====

=====
 Element: Tl AS Loc.: 16 Date: 10/10/2001
 Sample ID: 24238
 µL dispensed: 10 from 36, 5 from 37, 20 from 16
 µL dispensed: 10 from 36, 5 from 37, 20 from 16
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.660	-0.005	-0.007	0.022	12:25:08	No

=====

=====
 Element: Tl AS Loc.: 16 Date: 10/10/2001
 Sample ID: 24238
 µL dispensed: 5 from 37, 10 from 38, 20 from 16
 µL dispensed: 5 from 37, 10 from 38, 20 from 16
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	4.611	0.042	0.039	0.061	12:29:30	No

Recovery is 92.2%
 =====

=====
 Element: Tl AS Loc.: 17 Date: 10/10/2001
 Sample ID: 24268
 µL dispensed: 10 from 36, 5 from 37, 20 from 17
 µL dispensed: 10 from 36, 5 from 37, 20 from 17
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.005	0.001	-0.001	0.619	12:33:49	No

=====

=====
 Element: Tl AS Loc.: 18 Date: 10/10/2001
 Sample ID: 24287
 µL dispensed: 10 from 36, 5 from 37, 20 from 18
 µL dispensed: 10 from 36, 5 from 37, 20 from 18
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.196	-0.001	-0.003	0.036	12:38:05	No

=====

=====
 Element: Tl AS Loc.: 19 Date: 10/10/2001
 Sample ID: 24288
 µL dispensed: 10 from 36, 5 from 37, 20 from 19
 µL dispensed: 10 from 36, 5 from 37, 20 from 19
 =====

Repl #	Sample Conc µg/L	Blnk Corr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.233	-0.001	-0.003	0.036	12:42:19	No

=====

=====
Element: Tl AS Loc.: 20 Date: 10/10/2001
Sample ID: 24289
µL dispensed: 10 from 36, 5 from 37, 20 from 20
µL dispensed: 10 from 36, 5 from 37, 20 from 20
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.020	0.001	-0.001	0.040	12:46:33	No

=====

=====
Element: Tl AS Loc.: 21 Date: 10/10/2001
Sample ID: 24290
µL dispensed: 10 from 36, 5 from 37, 20 from 21
µL dispensed: 10 from 36, 5 from 37, 20 from 21
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.100	0.000	-0.002	0.048	12:50:42	No

=====

=====
Element: Tl AS Loc.: 22 Date: 10/10/2001
Sample ID: 24291
µL dispensed: 10 from 36, 5 from 37, 20 from 22
µL dispensed: 10 from 36, 5 from 37, 20 from 22
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.043	0.001	-0.001	0.033	12:54:53	No

=====

=====
Element: Tl AS Loc.: 23 Date: 10/10/2001
Sample ID: 24292
µL dispensed: 10 from 36, 5 from 37, 20 from 23
µL dispensed: 10 from 36, 5 from 37, 20 from 23
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.067	0.002	-0.001	0.035	12:59:02	No

=====

=====
Element: Tl AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.024	0.045	0.043	0.037	13:03:14	No

QC value within specified limits.

=====

=====
Element: Tl AS Loc.: 3 Date: 10/10/2001
Sample ID: CB
µL dispensed: 10 from 36, 5 from 37, 20 from 3
µL dispensed: 10 from 36, 5 from 37, 20 from 3
=====

Repl #	SampleConc µg/L	BlnkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.043	0.001	-0.002	0.000	13:07:25	No

=====

Element: Tl AS Loc.: 24 Date: 10/10/2001
 Sample ID: 24293
 µL dispensed: 10 from 36, 5 from 37, 20 from 24
 µL dispensed: 10 from 36, 5 from 37, 20 from 24

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.419	0.005	0.002	0.117	13:11:37	No

Element: Tl AS Loc.: 25 Date: 10/10/2001
 Sample ID: 24333
 µL dispensed: 10 from 36, 5 from 37, 20 from 25
 µL dispensed: 10 from 36, 5 from 37, 20 from 25

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.378	-0.002	-0.005	0.254	13:15:53	No

Element: Tl AS Loc.: 26 Date: 10/10/2001
 Sample ID: 24334
 µL dispensed: 10 from 36, 5 from 37, 20 from 26
 µL dispensed: 10 from 36, 5 from 37, 20 from 26

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.884	-0.007	-0.009	0.021	13:20:07	No

Element: Tl AS Loc.: 26 Date: 10/10/2001
 Sample ID: 24334
 µL dispensed: 5 from 37, 10 from 38, 20 from 26
 µL dispensed: 5 from 37, 10 from 38, 20 from 26

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.207	0.047	0.045	0.056	13:24:19	No

Recovery is 104.1%

Element: Tl AS Loc.: 27 Date: 10/10/2001
 Sample ID: 24334DUP
 µL dispensed: 10 from 36, 5 from 37, 20 from 27
 µL dispensed: 10 from 36, 5 from 37, 20 from 27

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.142	0.000	-0.003	0.021	13:28:38	No

Element: Tl AS Loc.: 28 Date: 10/10/2001
 Sample ID: 24335
 µL dispensed: 10 from 36, 5 from 37, 20 from 28
 µL dispensed: 10 from 36, 5 from 37, 20 from 28

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.131	0.002	0.000	0.019	13:33:00	No

=====
 Element: Tl AS Loc.: 29 Date: 10/10/2001
 Sample ID: 24395
 µL dispensed: 10 from 36, 5 from 37, 20 from 29
 µL dispensed: 10 from 36, 5 from 37, 20 from 29
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.118	0.000	-0.002	0.042	13:37:25	No

=====

=====
 Element: Tl AS Loc.: 30 Date: 10/10/2001
 Sample ID: 24396
 µL dispensed: 10 from 36, 5 from 37, 20 from 30
 µL dispensed: 10 from 36, 5 from 37, 20 from 30
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	0.086	0.002	-0.001	0.036	13:41:50	No

=====

=====
 Element: Tl AS Loc.: 31 Date: 10/10/2001
 Sample ID: 24397
 µL dispensed: 10 from 36, 5 from 37, 20 from 31
 µL dispensed: 10 from 36, 5 from 37, 20 from 31
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.245	-0.001	-0.003	0.038	13:46:17	No

=====

=====
 Element: Tl AS Loc.: 32 Date: 10/10/2001
 Sample ID: 24398
 µL dispensed: 10 from 36, 5 from 37, 20 from 32
 µL dispensed: 10 from 36, 5 from 37, 20 from 32
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.202	-0.001	-0.003	0.052	13:50:40	No

=====

=====
 Element: Tl AS Loc.: 2 Date: 10/10/2001
 Sample ID: IPC
 µL dispensed: 10 from 36, 5 from 37, 20 from 2
 µL dispensed: 10 from 36, 5 from 37, 20 from 2
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	5.225	0.047	0.045	0.040	13:54:54	No

QC value within specified limits.

=====

=====
 Element: Tl AS Loc.: 3 Date: 10/10/2001
 Sample ID: CB
 µL dispensed: 10 from 36, 5 from 37, 20 from 3
 µL dispensed: 10 from 36, 5 from 37, 20 from 3
 =====

Repl #	Sample Conc µg/L	BlkCorr Signal	Peak Area	Bkgnd Area	Time	Peak Stored
1	-0.254	-0.001	-0.004	0.001	13:59:05	No

=====

=====
Element: Tl AS Loc.: 33 Date: 10/10/2001
Sample ID: 24399
µL dispensed: 10 from 36, 5 from 37, 20 from 33
µL dispensed: 10 from 36, 5 from 37, 20 from 33
=====

Repl	SampleConc	BlnkCorr	Peak	Bkgnd	Time	Peak
#	µg/L	Signal	Area	Area		Stored
1	-0.342	-0.002	-0.004	0.052	14:03:22	No

=====

=====
Element: Tl AS Loc.: 34 Date: 10/10/2001
Sample ID: 24400
µL dispensed: 10 from 36, 5 from 37, 20 from 34
µL dispensed: 10 from 36, 5 from 37, 20 from 34
=====

Repl	SampleConc	BlnkCorr	Peak	Bkgnd	Time	Peak
#	µg/L	Signal	Area	Area		Stored
1	0.008	0.001	-0.001	0.038	14:07:38	No

=====

=====
Element: Tl AS Loc.: 35 Date: 10/10/2001
Sample ID: 24401
µL dispensed: 10 from 36, 5 from 37, 20 from 35
µL dispensed: 10 from 36, 5 from 37, 20 from 35
=====

Repl	SampleConc	BlnkCorr	Peak	Bkgnd	Time	Peak
#	µg/L	Signal	Area	Area		Stored
1	-0.195	-0.001	-0.003	0.033	14:11:54	No

=====

=====
Element: Tl AS Loc.: 2 Date: 10/10/2001
Sample ID: IPC
µL dispensed: 10 from 36, 5 from 37, 20 from 2
µL dispensed: 10 from 36, 5 from 37, 20 from 2
=====

Repl	SampleConc	BlnkCorr	Peak	Bkgnd	Time	Peak
#	µg/L	Signal	Area	Area		Stored
1	4.976	0.045	0.043	0.043	14:16:03	No

QC value within specified limits.

=====

=====
Element: Tl AS Loc.: 3 Date: 10/10/2001
Sample ID: CB
µL dispensed: 10 from 36, 5 from 37, 20 from 3
µL dispensed: 10 from 36, 5 from 37, 20 from 3
=====

Repl	SampleConc	BlnkCorr	Peak	Bkgnd	Time	Peak
#	µg/L	Signal	Area	Area		Stored
1	0.630	-0.005	-0.007	0.005	14:20:14	No

=====