NYC 9/11 Public Portal Document

SN

9/21/01 run — was wet with spike.
tech problem not verified as ok before rerun
~~Blk~~ spikes were low too.
Some data not printed

10/2/01
Bad QC - cup empty? seq 395
Cnd failed too seq 412
no good QC at end.
Reject 23997, 23998

Rerun
10/2/01

NYC-WTC_000065064