

Sample Information File Untitled

Description : SnF POTABLE
 Batch ID :
 Volume Units :
 Weight Units :
 Analyst :
 Sample Volume : 1.000
 Nominal Weight : 1.000000

AS Sample ID Loc	Sample Sample Weight Units	User Dilution	Remarks
---------------------	-------------------------------	------------------	---------

1 QCS			
2 IPC			
3 CB			
4 LFB			
5 LRB			
6 23836			
7 23836DUP			
8 23837			
9 23838			
10 23839			
11 23842			
12 23843			
13 23844			
14 23900			
15 23901			
16 23902			
17 23903			
18 23904			
19 23905			
20 23906			
21 23909			
22 23910			
23 23913			
24 23914			
25 23915			
26 23947			
27 23947DUP			
28 23948			
29 23993			
30 23994			
31 23995			
32 23996			
33 23997			
34 23998			

DATE ANALYZED 10/2/01
 TIME ANALYZED 6:25 PM
 ANALYST INITIALS LD
 REVIEWED BY

QC Reviewed by,
 QB 11/6/01

Method Name: SnHGA Element: Sn
Method Description: Snhga SM3113 B

Date: 10/02/2001

Technique: Furnace

Wavelength: 286.3 nm

Lamp Current: 30

Sample Info File: Untitled

Calibration Equation: Calc. Intercept: Linear

Slit Width: 0.70 nm

Energy: 62

Results Data Set:

Element: Sn Seq. No.: 363 AS Loc.: 36 Date: 10/02/2001

Sample ID: Calib Blank

µL dispensed: 8 from 36, 5 from 37, 12 from 36

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			-0.003	0.000	0.000	0.000	0.000	18:51:42	No

Auto-zero performed.

Element: Sn Seq. No.: 364 AS Loc.: 38 Date: 10/02/2001

Sample ID: s50

µL dispensed: 17 from 36, 5 from 37, 3 from 38

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.098	0.098	0.260	0.017	0.032	18:54:19	No

Standard number 1 applied. [50.0]

Correlation Coefficient: 1.0000

Slope: 0.0020

Element: Sn Seq. No.: 365 AS Loc.: 38 Date: 10/02/2001

Sample ID: s100

µL dispensed: 14 from 36, 5 from 37, 6 from 38

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.180	0.180	0.439	0.034	0.054	18:57:00	No

Standard number 2 applied. [100.0]

Correlation Coefficient: 0.9986

Slope: 0.0018

Element: Sn Seq. No.: 366 AS Loc.: 38 Date: 10/02/2001

Sample ID: s150

µL dispensed: 11 from 36, 5 from 37, 9 from 38

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.256	0.256	0.620	0.046	0.083	18:59:41	No

Standard number 3 applied. [150.0]

Correlation Coefficient: 0.9983

Slope: 0.0017

Element: Sn Seq. No.: 367 AS Loc.: 38 Date: 10/02/2001

Sample ID: s200

µL dispensed: 8 from 36, 5 from 37, 12 from 38

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1			0.325	0.325	0.809	0.059	0.117	19:02:22	No

The calibration curve may not be linear.

Standard number 4 applied. [200.0]

Correlation Coefficient: 0.9977

Slope: 0.0016

=====

Element: Sn Seq. No.: 368 AS Loc.: 1 Date: 10/02/2001

Sample ID: QCS

µL dispensed: 8 from 36, 5 from 37, 12 from 1

Repl	SampleConc	StndConc	BlnkCorr	Peak	Peak	Bkgnd	Bkgnd	Time	Peak
#	µg/L	µg/L	Signal	Area	Height	Area	Height		Stored
1	103.1	103.1	0.177	0.177	0.463	0.038	0.058	19:05:17	No

Pass

QC value within specified limits.

=====

Element: Sn Seq. No.: 369 AS Loc.: 2 Date: 10/02/2001
 Sample ID: IPC
 µL dispensed: 8 from 36, 5 from 37, 12 from 2

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	99.13	99.13	0.170	0.170	0.434	0.038	0.056	19:07:53	No

QC value within specified limits.

=====

Element: Sn Seq. No.: 370 AS Loc.: 3 Date: 10/02/2001
 Sample ID: CB
 µL dispensed: 8 from 36, 5 from 37, 12 from 3

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.441	-6.441	0.000	0.000	0.003	0.015	0.011	19:10:30	No

=====

Element: Sn Seq. No.: 371 AS Loc.: 4 Date: 10/02/2001
 Sample ID: LFB
 µL dispensed: 8 from 36, 5 from 37, 12 from 4

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	196.7	196.7	0.328	0.328	0.730	0.057	0.098	19:13:05	No

QC value within specified limits.

=====

Element: Sn Seq. No.: 372 AS Loc.: 5 Date: 10/02/2001
 Sample ID: LRB
 µL dispensed: 8 from 36, 5 from 37, 12 from 5

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.994	-5.994	0.000	0.000	0.004	0.018	0.013	19:15:42	No

=====

Element: Sn Seq. No.: 373 AS Loc.: 6 Date: 10/02/2001
 Sample ID: 23836
 µL dispensed: 8 from 36, 5 from 37, 12 from 6

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.781	-5.781	0.001	0.001	0.005	0.023	0.020	19:18:20	No

=====

Element: Sn Seq. No.: 374 AS Loc.: 6 Date: 10/02/2001
 Sample ID: 23836
 µL dispensed: 5 from 36, 5 from 37, 3 from 38, 12 from 6

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	52.70	52.70	0.095	0.095	0.276	0.029	0.034	19:21:05	No

Recovery is 105.4%

=====
 Element: Sn Seq. No.: 375 AS Loc.: 7 Date: 10/02/2001
 Sample ID: 23836DUP
 µL dispensed: 8 from 36, 5 from 37, 12 from 7

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.318	-6.318	0.000	0.000	0.004	0.023	0.022	19:23:44	No

=====
 Element: Sn Seq. No.: 376 AS Loc.: 8 Date: 10/02/2001
 Sample ID: 23837
 µL dispensed: 8 from 36, 5 from 37, 12 from 8

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.358	-5.358	0.001	0.001	0.008	0.020	0.020	19:26:24	No

=====
 Element: Sn Seq. No.: 377 AS Loc.: 9 Date: 10/02/2001
 Sample ID: 23838
 µL dispensed: 8 from 36, 5 from 37, 12 from 9

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.995	-5.995	0.000	0.000	0.004	0.019	0.019	19:29:07	No

=====
 Element: Sn Seq. No.: 378 AS Loc.: 10 Date: 10/02/2001
 Sample ID: 23839
 µL dispensed: 8 from 36, 5 from 37, 12 from 10

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.199	-7.199	-0.001	-0.001	0.003	0.022	0.021	19:31:50	No

=====
 Element: Sn Seq. No.: 379 AS Loc.: 11 Date: 10/02/2001
 Sample ID: 23842
 µL dispensed: 8 from 36, 5 from 37, 12 from 11

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.250	-7.250	-0.002	-0.002	0.003	0.039	0.028	19:34:33	No

=====
 Element: Sn Seq. No.: 380 AS Loc.: 12 Date: 10/02/2001
 Sample ID: 23843
 µL dispensed: 8 from 36, 5 from 37, 12 from 12

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.782	-6.782	-0.001	-0.001	0.003	0.021	0.019	19:37:18	No

=====
 Element: Sn Seq. No.: 381 AS Loc.: 13 Date: 10/02/2001
 Sample ID: 23844
 µL dispensed: 8 from 36, 5 from 37, 12 from 13

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.559	-6.559	0.000	0.000	0.003	0.019	0.020	19:40:03	No

=====
 Element: Sn Seq. No.: 382 AS Loc.: 14 Date: 10/02/2001
 Sample ID: 23900
 µL dispensed: 8 from 36, 5 from 37, 12 from 14
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.804	-6.804	-0.001	-0.001	0.003	0.020	0.020	19:42:49	No

=====
 Element: Sn Seq. No.: 383 AS Loc.: 2 Date: 10/02/2001
 Sample ID: IPC
 µL dispensed: 8 from 36, 5 from 37, 12 from 2
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	100.8	100.8	0.173	0.173	0.437	0.037	0.055	19:45:29	No

QC value within specified limits.

=====
 Element: Sn Seq. No.: 384 AS Loc.: 3 Date: 10/02/2001
 Sample ID: CB
 µL dispensed: 8 from 36, 5 from 37, 12 from 3
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.098	-7.098	-0.001	-0.001	0.004	0.016	0.013	19:48:07	No

=====
 Element: Sn Seq. No.: 385 AS Loc.: 15 Date: 10/02/2001
 Sample ID: 23901
 µL dispensed: 8 from 36, 5 from 37, 12 from 15
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.259	-7.259	-0.002	-0.002	0.003	0.022	0.021	19:50:49	No

=====
 Element: Sn Seq. No.: 386 AS Loc.: 16 Date: 10/02/2001
 Sample ID: 23902
 µL dispensed: 8 from 36, 5 from 37, 12 from 16
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.042	-6.042	0.000	0.000	0.003	0.020	0.021	19:53:32	No

=====
 Element: Sn Seq. No.: 387 AS Loc.: 16 Date: 10/02/2001
 Sample ID: 23902
 µL dispensed: 5 from 36, 5 from 37, 3 from 38, 12 from 16
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	52.89	52.89	0.096	0.096	0.281	0.032	0.035	19:56:19	No

Recovery is 105.8%

=====
 Element: Sn Seq. No.: 388 AS Loc.: 17 Date: 10/02/2001
 Sample ID: 23903
 µL dispensed: 8 from 36, 5 from 37, 12 from 17
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.467	-6.467	0.000	0.000	0.003	0.022	0.021	19:59:00	No

=====
 Element: Sn Seq. No.: 389 AS Loc.: 18 Date: 10/02/2001
 Sample ID: 23904
 µL dispensed: 8 from 36, 5 from 37, 12 from 18
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.837	-6.837	-0.001	-0.001	0.002	0.021	0.019	20:01:40	No

=====
 Element: Sn Seq. No.: 390 AS Loc.: 19 Date: 10/02/2001
 Sample ID: 23905
 µL dispensed: 8 from 36, 5 from 37, 12 from 19
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.726	-6.726	-0.001	-0.001	0.002	0.022	0.021	20:04:20	No

=====
 Element: Sn Seq. No.: 391 AS Loc.: 20 Date: 10/02/2001
 Sample ID: 23906
 µL dispensed: 8 from 36, 5 from 37, 12 from 20
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.658	-5.658	0.001	0.001	0.003	0.020	0.020	20:06:57	No

=====
 Element: Sn Seq. No.: 392 AS Loc.: 21 Date: 10/02/2001
 Sample ID: 23909
 µL dispensed: 8 from 36, 5 from 37, 12 from 21
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-4.949	-4.949	0.002	0.002	0.008	0.022	0.021	20:09:33	No

=====
 Element: Sn Seq. No.: 393 AS Loc.: 22 Date: 10/02/2001
 Sample ID: 23910
 µL dispensed: 8 from 36, 5 from 37, 12 from 22
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.889	-6.889	-0.001	-0.001	0.003	0.022	0.021	20:12:09	No

=====
 Element: Sn Seq. No.: 394 AS Loc.: 23 Date: 10/02/2001
 Sample ID: 23913
 µL dispensed: 8 from 36, 5 from 37, 12 from 23
 =====

Repl #	SampleConc µg/L	StdConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.039	-7.039	-0.001	-0.001	0.004	0.021	0.021	20:14:45	No

=====
 Element: Sn Seq. No.: 395 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

µL dispensed: 8 from 36, 5 from 37, 12 from 2

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	37.61	37.61	0.071	0.071	0.150	0.015	0.020	20:17:23	No

QC out of limits. QC will be repeated based on Retry setting.

What happened?

=====
 Element: Sn Seq. No.: 396 AS Loc.: 2 Date: 10/02/2001

Sample ID: IPC

µL dispensed: 8 from 36, 5 from 37, 12 from 2

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	101.6	101.6	0.174	0.174	0.443	0.038	0.054	20:19:59	No

QC value within specified limits.

=====
 Element: Sn Seq. No.: 397 AS Loc.: 3 Date: 10/02/2001

Sample ID: CB

µL dispensed: 8 from 36, 5 from 37, 12 from 3

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.524	-6.524	0.000	0.000	0.003	0.016	0.013	20:22:35	No

=====
 Element: Sn Seq. No.: 398 AS Loc.: 24 Date: 10/02/2001

Sample ID: 23914

µL dispensed: 8 from 36, 5 from 37, 12 from 24

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.063	-6.063	0.000	0.000	0.003	0.019	0.020	20:25:12	No

=====
 Element: Sn Seq. No.: 399 AS Loc.: 25 Date: 10/02/2001

Sample ID: 23915

µL dispensed: 8 from 36, 5 from 37, 12 from 25

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.273	-6.273	0.000	0.000	0.003	0.038	0.029	20:27:50	No

=====
 Element: Sn Seq. No.: 400 AS Loc.: 26 Date: 10/02/2001

Sample ID: 23947

µL dispensed: 8 from 36, 5 from 37, 12 from 26

Repl #	SampleConc µg/L	StndConc µg/L	BlkCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.606	-6.606	-0.001	-0.001	0.003	0.022	0.021	20:30:30	No

=====
 Element: Sn Seq. No.: 401 AS Loc.: 26 Date: 10/02/2001
 Sample ID: 23947
 µL dispensed: 5 from 36, 5 from 37, 3 from 38, 12 from 26
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	54.85	54.85	0.099	0.099	0.274	0.032	0.033	20:33:16	No

Recovery is 109.7%

=====
 Element: Sn Seq. No.: 402 AS Loc.: 27 Date: 10/02/2001
 Sample ID: 23947DUP
 µL dispensed: 8 from 36, 5 from 37, 12 from 27
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.735	-6.735	-0.001	-0.001	0.003	0.021	0.020	20:35:58	No

=====
 Element: Sn Seq. No.: 403 AS Loc.: 28 Date: 10/02/2001
 Sample ID: 23948
 µL dispensed: 8 from 36, 5 from 37, 12 from 28
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.209	-6.209	0.000	0.000	0.003	0.023	0.022	20:38:43	No

=====
 Element: Sn Seq. No.: 404 AS Loc.: 29 Date: 10/02/2001
 Sample ID: 23993
 µL dispensed: 8 from 36, 5 from 37, 12 from 29
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.297	-6.297	0.000	0.000	0.003	0.022	0.023	20:41:28	No

=====
 Element: Sn Seq. No.: 405 AS Loc.: 30 Date: 10/02/2001
 Sample ID: 23994
 µL dispensed: 8 from 36, 5 from 37, 12 from 30
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.379	-6.379	0.000	0.000	0.002	0.021	0.023	20:44:13	No

=====
 Element: Sn Seq. No.: 406 AS Loc.: 31 Date: 10/02/2001
 Sample ID: 23995
 µL dispensed: 8 from 36, 5 from 37, 12 from 31
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.455	-6.455	0.000	0.000	0.003	0.040	0.032	20:46:57	No

=====
 Element: Sn Seq. No.: 407 AS Loc.: 32 Date: 10/02/2001
 Sample ID: 23996
 µL dispensed: 8 from 36, 5 from 37, 12 from 32
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.424	-6.424	0.000	0.000	0.004	0.044	0.030	20:49:40	No

=====
 Element: Sn Seq. No.: 408 AS Loc.: 2 Date: 10/02/2001
 Sample ID: IPC
 µL dispensed: 8 from 36, 5 from 37, 12 from 2
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	108.2	108.2	0.185	0.185	0.454	0.040	0.055	20:52:19	No

QC value within specified limits.

=====
 Element: Sn Seq. No.: 409 AS Loc.: 3 Date: 10/02/2001
 Sample ID: CB
 µL dispensed: 8 from 36, 5 from 37, 12 from 3
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-6.253	-6.253	0.000	0.000	0.004	0.018	0.013	20:54:54	No

=====
 Element: Sn Seq. No.: 410 AS Loc.: 33 Date: 10/02/2001
 Sample ID: 23997
 µL dispensed: 8 from 36, 5 from 37, 12 from 33
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-7.157	-7.157	-0.001	-0.001	0.003	0.022	0.022	20:57:36	No

=====
 Element: Sn Seq. No.: 411 AS Loc.: 34 Date: 10/02/2001
 Sample ID: 23998
 µL dispensed: 8 from 36, 5 from 37, 12 from 34
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	-5.546	-5.546	0.001	0.001	0.005	0.020	0.021	21:00:16	No

=====
 Element: Sn Seq. No.: 412 AS Loc.: 2 Date: 10/02/2001
 Sample ID: IPC
 µL dispensed: 8 from 36, 5 from 37, 12 from 2
 =====

Repl #	SampleConc µg/L	StndConc µg/L	BlncCorr Signal	Peak Area	Peak Height	Bkgnd Area	Bkgnd Height	Time	Peak Stored
1	47.61	47.61	0.087	0.087	0.172	0.017	0.023	21:02:54	No

QC out of limits. Current analytical sequence continued.

FAIL

=====									
Element: Sn		Seq. No.: 413		AS Loc.: 3		Date: 10/02/2001			
Sample ID: CB									
µL dispensed: 8 from 36, 5 from 37, 12 from 3									

Repl	SampleConc	StndConc	BlnkCorr	Peak	Peak	Bkgnd	Bkgnd	Time	Peak
#	µg/L	µg/L	Signal	Area	Height	Area	Height		Stored
1	-5.669	-5.669	0.001	0.001	0.003	0.005	0.011	21:05:30	No