

WinHg Database 1.3

File Utility Help

RN↓ RN↑ ?

Protocol hg-cv Dataset/Proto RUNS1 /hg-cv

Protocol | Line info | Cal Curve | Report | Ctrl Chart | Viewer

Reset

Calib Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B 1.41174e-5

C 8.98963e-4

Rho .999945

Type Linear

Rel Abs. 707458

Accepted

Calibrated

Accepted

Accept

New

Include S1 Rep 1 ☒ 2 ☐ 3 ☐ 4 ☐ 5 ☐

Conc. 10.0

S	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3
01	0.0000	.036	.036	2479	1	2478		
02	.40000	.394	-.006	27814	0%	27814		
03	1.0000	.975	-.025	68970	0%	68970		
04	3.0000	2.95	-.049	208949	0%	208949		
05	5.0000	5.06	.057	358143	0%	358143		
06	10.000	9.99	-.012	707458	0%	707458		

Ready

NUM

11/7/01

NO MSD's
NO blanks as samples

All runs
Reviewed by,
JB 1/6/02

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 91

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 612		14:52:26	07 Nov 01	HG
Hg	.000	ppb	2478					
*** Standard: 2 Rep: 1				Seq: 613		14:54:23	07 Nov 01	HG
Hg	.400	ppb	27814					
*** Standard: 3 Rep: 1				Seq: 614		14:56:54	07 Nov 01	HG
Hg	1.00	ppb	68970					
*** Standard: 4 Rep: 1				Seq: 615		14:58:43	07 Nov 01	HG
Hg	3.00	ppb	208949					
*** Standard: 5 Rep: 1				Seq: 616		15:00:53	07 Nov 01	HG
Hg	5.00	ppb	358143					
*** Standard: 6 Rep: 1				Seq: 617		15:02:32	07 Nov 01	HG
Hg	10.0	ppb	707458					
*** Check Standard: 2 Ck21.0 qc				Seq: 618		15:06:19	07 Nov 01	HG
Line Flag %Rcv. Found True Units						SD/RSD		
Hg		98.4	.984	1.00	ppb	.000		
*** Sample ID: 27007 pot				Seq: 619		15:08:16	07 Nov 01	HG
Hg	-.004	ppb	.000	-.004				
*** Sample ID: 27007d				Seq: 620		15:09:53	07 Nov 01	HG
Hg	-.008	ppb	.000	-.008	✓ LDC			
*** Sample ID: 27007 spk				Seq: 621		15:11:30	07 Nov 01	HG
Hg	4.95	ppb	.000	4.95	✓ 99%			
*** Sample ID: 27008				Seq: 622		15:13:27	07 Nov 01	HG
Hg	.001	ppb	.000	.001				
*** Sample ID: 27009				Seq: 623		15:15:03	07 Nov 01	HG
Hg	.005	ppb	.000	.005				
*** Sample ID: 27010				Seq: 624		15:16:41	07 Nov 01	HG
Hg	.001	ppb	.000	.001				
*** Sample ID: 27011				Seq: 625		15:18:27	07 Nov 01	HG
Hg	-.000	ppb	.000	-.000				

no MSD

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 92

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 27012								
				Seq: 626	15:20:14	07 Nov 01	HG	
Hg	-.010	ppb	.000	-.010				=
*** Sample ID: 27013								
				Seq: 627	15:22:04	07 Nov 01	HG	
Hg	-.006	ppb	.000	-.006				=
*** Sample ID: 27014								
				Seq: 628	15:24:14	07 Nov 01	HG	
Hg	-.010	ppb	.000	-.010				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		108.	1.08	1.00	ppb	.000		=
*** Sample ID: 27015								
				Seq: 630	15:27:34	07 Nov 01	HG	
Hg	.049	ppb	.000	.049				=
*** Sample ID: 27055								
				Seq: 631	15:29:11	07 Nov 01	HG	
Hg	.057	ppb	.000	.057				=
*** Sample ID: 27055d ✓ <i>CD</i>								
				Seq: 632	15:30:51	07 Nov 01	HG	
Hg	.007	ppb	.000	.007				=
*** Sample ID: 27055 spk								
				Seq: 633	15:32:27	07 Nov 01	HG	
Hg	5.39	ppb	107% .000	5.39				=
*** Sample ID: 27056								
				Seq: 634	15:34:08	07 Nov 01	HG	
Hg	-.002	ppb	.000	-.002				=
*** Sample ID: 27188								
				Seq: 635	15:36:14	07 Nov 01	HG	
Hg	-.003	ppb	.000	-.003				=
*** Sample ID: 27189								
				Seq: 636	15:37:56	07 Nov 01	HG	
Hg	.079	ppb	.000	.079				=
*** Sample ID: 27190								
				Seq: 637	15:39:35	07 Nov 01	HG	
Hg	-.015	ppb	.000	-.015				=
*** Sample ID: 27191								
				Seq: 638	15:41:12	07 Nov 01	HG	
Hg	.011	ppb	.000	.011				=

westchester county labs


Folder: RUNS1

Page 93

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 27192								
				Seq: 639	15:42:49	07 Nov 01	HG	
Hg	-.005	ppb	.000	-.005				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		104.	1.04	1.00	ppb	.000		
*** Sample ID: 27193								
				Seq: 641	15:46:43	07 Nov 01	HG	
Hg	.056	ppb	.000	.056				
*** Sample ID: 27194								
				Seq: 642	15:48:56	07 Nov 01	HG	
Hg	.070	ppb	.000	.070				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		106.	1.06	1.00	ppb	.000		



westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 98


Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 665	14:44:57	09 Nov 01	HG	
Hg	.000	ppb	1960					=
*** Standard: 2 Rep: 1								
				Seq: 666	14:46:35	09 Nov 01	HG	
Hg	.400	ppb	29890					=
*** Standard: 3 Rep: 1								
				Seq: 667	14:48:12	09 Nov 01	HG	
Hg	1.00	ppb	74985					=
*** Standard: 4 Rep: 1								
				Seq: 668	14:50:29	09 Nov 01	HG	
Hg	3.00	ppb	222347					=
*** Standard: 5 Rep: 1								
				Seq: 669	14:52:16	09 Nov 01	HG	
Hg	5.00	ppb	388872					=
*** Standard: 6 Rep: 1								
				Seq: 670	14:54:36	09 Nov 01	HG	
Hg	10.0	ppb	769088					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv	Found	True	Seq: 671	15:00:18	09 Nov 01	HG
Hg		95.7	.957	1.00	Units	SD/RSD		
					ppb	.000		=
*** Sample ID: 27232 pot								
				Seq: 672	15:01:56	09 Nov 01	HG	
Hg	.008	ppb	.000	.008				=
*** Sample ID: 27233								
				Seq: 673	15:03:34	09 Nov 01	HG	
Hg	.025	ppb	.000	.025				=
*** Sample ID: 27389								
				Seq: 674	15:06:02	09 Nov 01	HG	
Hg	.042	ppb	.000	.042				=
*** Sample ID: 27389d ✓								
				Seq: 675	15:07:40	09 Nov 01	HG	
Hg	.031	ppb	.000	.031				=
*** Sample ID: 27389 spk ✓ 114%								
				Seq: 676	15:09:19	09 Nov 01	HG	
Hg	5.75	ppb	.000	5.75				=
*** Sample ID: 27390								
				Seq: 677	15:11:01	09 Nov 01	HG	
Hg	.002	ppb	.000	.002				=
*** Sample ID: 27391								
				Seq: 678	15:12:38	09 Nov 01	HG	
Hg	.021	ppb	.000	.021				=

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 99

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 27392								
				Seq: 679	15:14:17	09 Nov 01	HG	
Hg	.022	ppb	.000	.022				=
*** Sample ID: 27393								
				Seq: 680	15:16:18	09 Nov 01	HG	
Hg	.061	ppb	.000	.061				=
*** Sample ID: 27394								
				Seq: 681	15:18:15	09 Nov 01	HG	
Hg	.023	ppb	.000	.023				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.9	.969	1.00	ppb	.000		=
*** Sample ID: 27395								
				Seq: 683	15:22:50	09 Nov 01	HG	
Hg	.030	ppb	.000	.030				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.0	.970	1.00	ppb	.000		=



11/14/01

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 112

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 730		15:09:47	14 Nov 01	HG
Hg	.000	ppb	106					
*** Standard: 2 Rep: 1				Seq: 731		15:11:49	14 Nov 01	HG
Hg	.400	ppb	28890					
*** Standard: 3 Rep: 1				Seq: 732		15:13:25	14 Nov 01	HG
Hg	1.00	ppb	69580					
*** Standard: 4 Rep: 1				Seq: 733		15:15:16	14 Nov 01	HG
Hg	3.00	ppb	235082					
*** Standard: 5 Rep: 1				Seq: 734		15:16:56	14 Nov 01	HG
Hg	5.00	ppb	397638					
*** Standard: 6 Rep: 1				Seq: 735		15:18:47	14 Nov 01	HG
Hg	10.0	ppb	771731					
*** Check Standard: 2 Ck21.0 qc				Seq: 736		15:20:34	14 Nov 01	HG
Line Flag	%Rev	Found	True	Units		SD/RSD		
Hg	98.1	.981	1.00	ppb		.000		
*** Sample ID: 27544 pot				Seq: 737		15:22:23	14 Nov 01	HG
Hg	.116	ppb	.000	.116				
*** Sample ID: 27544d ✓				Seq: 738		15:24:09	14 Nov 01	HG
Hg	.089	ppb	.000	.089				
*** Sample ID: 27544 spk ✓ 99%				Seq: 739		15:25:46	14 Nov 01	HG
Hg	5.05	ppb	.000	5.05				
*** Sample ID: 27545				Seq: 740		15:27:42	14 Nov 01	HG
Hg	.181	ppb	.000	.181				
*** Sample ID: 27546				Seq: 741		15:29:18	14 Nov 01	HG
Hg	.134	ppb	.000	.134				
*** Sample ID: 27547				Seq: 742		15:30:57	14 Nov 01	HG
Hg	.077	ppb	.000	.077				
*** Sample ID: 27548				Seq: 743		15:32:54	14 Nov 01	HG
Hg	.083	ppb	.000	.083				

westchester county labs

Folder: RUNS1


Page 114

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5

*** Sample ID: 27639				Seq: 757	15:58:40	14 Nov 01	HG	
Hg	.048	ppb	.000	.048				=
*** Check Standard: 2 Ck21.0 qc				Seq: 758	16:00:39	14 Nov 01	HG	
Line	Flag	%Rcv	Found	True	Units	SD/RSD		
Hg		88.5	.885	1.00	ppb	.000		=
*** Sample ID: 27678				Seq: 759	16:02:41	14 Nov 01	HG	
Hg	.092	ppb	.000	.092				=
=====								
*** Sample ID: 27679				Seq: 760	16:04:20	14 Nov 01	HG	
Hg	.064	ppb	.000	.064				=
*** Check Standard: 2 Ck21.0 qc				Seq: 769	16:20:15	14 Nov 01	HG	
Line	Flag	%Rcv	Found	True	Units	SD/RSD		
Hg		92.8	.928	1.00	ppb	.000		=



11/16/01

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 124

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 800	15:34:34	16 Nov 01	HG	
Hg	.000	ppb	1954					=
*** Standard: 2 Rep: 1								
				Seq: 801	15:36:11	16 Nov 01	HG	
Hg	.400	ppb	28553					=
*** Standard: 3 Rep: 1								
				Seq: 802	15:37:59	16 Nov 01	HG	
Hg	1.00	ppb	65300					=
*** Standard: 4 Rep: 1								
				Seq: 803	15:39:57	16 Nov 01	HG	
Hg	3.00	ppb	221115					=
*** Standard: 5 Rep: 1								
				Seq: 804	15:41:56	16 Nov 01	HG	
Hg	5.00	ppb	370224					=
*** Standard: 6 Rep: 1								
				Seq: 805	15:43:33	16 Nov 01	HG	
Hg	10.0	ppb	735452					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	Rev.	Found	True	Units	SD/RSD		
Hg		95.4	.954	1.00	ppb	.000		=
*** Sample ID: 27817 pot								
				Seq: 807	15:52:41	16 Nov 01	HG	
Hg	.041	ppb	.000	.041				=
*** Sample ID: 27817d ✓								
				Seq: 808	15:54:20	16 Nov 01	HG	
Hg	.039	ppb	.000	.039				=
*** Sample ID: 27817 spk								
				Seq: 809	15:56:07	16 Nov 01	HG	
Hg	4.62	ppb	.000	4.62				=
*** Sample ID: 27818								
				Seq: 810	15:57:43	16 Nov 01	HG	
Hg	.019	ppb	.000	.019				=
*** Sample ID: 27819								
				Seq: 811	15:59:19	16 Nov 01	HG	
Hg	.019	ppb	.000	.019				=
*** Sample ID: 27823								
				Seq: 812	16:01:28	16 Nov 01	HG	
Hg	.012	ppb	.000	.012				=
*** Sample ID: 27826								
				Seq: 813	16:03:04	16 Nov 01	HG	
Hg	.056	ppb	.000	.056				=

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 125

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 27828								
				Seq: 814	16:04:41	16 Nov 01	HG	
Hg	.013	ppb	.000	.013				=
*** Sample ID: 27831								
				Seq: 815	16:06:20	16 Nov 01	HG	
Hg	.015	ppb	.000	.015				=
*** Sample ID: 27951								
				Seq: 816	16:07:58	16 Nov 01	HG	
Hg	.015	ppb	.000	.015				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Recy.	Found	True	Units	SD/RSD		
Hg		93.0	.930	1.00	ppb	.000		=
*** Sample ID: 27952								
				Seq: 818	16:12:07	16 Nov 01	HG	
Hg	.023	ppb	.000	.023				=
*** Sample ID: 27953								
				Seq: 819	16:13:45	16 Nov 01	HG	
Hg	.016	ppb	.000	.016				=
*** Sample ID: 27954								
				Seq: 820	16:15:25	16 Nov 01	HG	
Hg	.015	ppb	.000	.015				=
*** Sample ID: 27955								
				Seq: 821	16:17:15	16 Nov 01	HG	
Hg	.010	ppb	.000	.010				=
*** Sample ID: 27956								
				Seq: 822	16:19:24	16 Nov 01	HG	
Hg	.013	ppb	.000	.013				=
*** Sample ID: 27957								
				Seq: 823	16:21:02	16 Nov 01	HG	
Hg	.014	ppb	.000	.014				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Recy.	Found	True	Units	SD/RSD		
Hg		92.7	.927	1.00	ppb	.000		=

Missing a page?

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 131

Line	Conc.	Units	SD/RSD	1	2	3	4	5

*** Standard: 1 Rep: 1				Seq: 850		14:48:44	20 Nov 01	HG
Hg	.000	ppb	89					=
*** Standard: 2 Rep: 1				Seq: 851		14:50:41	20 Nov 01	HG
Hg	.400	ppb	24871					=
*** Standard: 3 Rep: 1				Seq: 852		14:52:28	20 Nov 01	HG
Hg	1.00	ppb	69062					=
*** Standard: 4 Rep: 1				Seq: 853		14:54:07	20 Nov 01	HG
Hg	3.00	ppb	214091					=
*** Standard: 5 Rep: 1				Seq: 854		14:55:46	20 Nov 01	HG
Hg	5.00	ppb	352287					=
*** Standard: 6 Rep: 1				Seq: 855		14:58:38	20 Nov 01	HG
Hg	10.0	ppb	635926					=
*** Check Standard: 2 Ck21.0 qc				Seq: 856		15:00:40	20 Nov 01	HG
Line Flag %Rcv. Found True Units SD/RSD								
Hg		90.1	.901	1.00	ppb	.000		=
*** Sample ID: 28043 pot				Seq: 857		15:02:57	20 Nov 01	HG
Hg	-.032	ppb	.000	-.032				=
=====								
*** Sample ID: 28043d				Seq: 858		15:04:47	20 Nov 01	HG
Hg	-.085	ppb	.000	-.085				=
=====								
*** Sample ID: 28043 spk				Seq: 859		15:06:32	20 Nov 01	HG
Hg	5.77	ppb	.000	5.77				=
=====								
*** Sample ID: 28044				Seq: 860		15:08:10	20 Nov 01	HG
Hg	-.120	ppb	.000	-.120				=
=====								
*** Sample ID: 28045				Seq: 861		15:09:50	20 Nov 01	HG
Hg	-.124	ppb	.000	-.124				=
=====								
*** Sample ID: 28046				Seq: 862		15:11:39	20 Nov 01	HG
Hg	-.139	ppb	.000	-.139				=
=====								
*** Sample ID: 28047				Seq: 863		15:13:15	20 Nov 01	HG
Hg	-.126	ppb	.000	-.126				=

westchester county labs

 Folder: RUNS1
 Protocol: hg-cv
 POST-RUN REPORT

Page 132

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 28048								
				Seq: 864	15:15:04	20 Nov 01	HG	
Hg	-.094	ppb	.000	-.094				
=====								
*** Sample ID: 28049								
				Seq: 865	15:16:43	20 Nov 01	HG	
Hg	-.091	ppb	.000	-.091				
=====								
*** Sample ID: 28164								
				Seq: 866	15:18:20	20 Nov 01	HG	
Hg	-.109	ppb	.000	-.109				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		105.	1.05	1.00	ppb	.000		
=====								
*** Sample ID: 28165								
				Seq: 868	15:21:38	20 Nov 01	HG	
Hg	-.136	ppb	.000	-.136				
=====								
*** Sample ID: 28166								
				Seq: 869	15:23:56	20 Nov 01	HG	
Hg	-.110	ppb	.000	-.110				
=====								
*** Sample ID: 28167								
				Seq: 870	15:25:31	20 Nov 01	HG	
Hg	-.101	ppb	.000	-.101				
=====								
*** Sample ID: 28168								
				Seq: 871	15:28:19	20 Nov 01	HG	
Hg	-.100	ppb	.000	-.100				
=====								
*** Sample ID: 28169								
				Seq: 872	15:30:00	20 Nov 01	HG	
Hg	-.115	ppb	.000	-.115				
=====								
*** Sample ID: 28170								
				Seq: 873	15:31:56	20 Nov 01	HG	
Hg	-.138	ppb	.000	-.138				
=====								
*** Sample ID: 28171								
				Seq: 874	15:33:32	20 Nov 01	HG	
Hg	-.130	ppb	.000	-.130				
=====								
*** Sample ID: 28172								
				Seq: 875	15:35:11	20 Nov 01	HG	
Hg	-.136	ppb	.000	-.136				
=====								
*** Sample ID: 28229								
				Seq: 876	15:37:13	20 Nov 01	HG	
Hg	-.089	ppb	.000	-.089				

westchester county labs

Folder: RUNS1
 Protocol: hg-cv
 POST-RUN REPORT

Page 133

Line	Conc.	Units	SD/RSD	1	2	3	4	5	
*** Sample ID: 28230									
				Seq: 877	15:39:22 20 Nov 01			HG	
Hg	-.099	ppb	.000	-.099					=
*** Check Standard: 2 Ck21.0 qc									
				Seq: 878	15:41:18 20 Nov 01			HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg		105.	1.05	1.00	ppb	.000			

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 134

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 888		16:00:46	20 Nov 01	HG
Hg	.000	ppb	2173					=
*** Standard: 2 Rep: 1								
				Seq: 889		16:03:04	20 Nov 01	HG
Hg	.400	ppb	28162					=
*** Standard: 3 Rep: 1								
				Seq: 890		16:04:54	20 Nov 01	HG
Hg	1.00	ppb	74298					=
*** Standard: 4 Rep: 1								
				Seq: 891		16:06:31	20 Nov 01	HG
Hg	3.00	ppb	230687					=
*** Standard: 5 Rep: 1								
				Seq: 892		16:08:19	20 Nov 01	HG
Hg	5.00	ppb	379182					=
*** Standard: 6 Rep: 1								
				Seq: 893		16:09:57	20 Nov 01	HG
Hg	10.0	ppb	676671					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rev	Found	True	Units	SD/RSD		
Hg		92.1	.921	1.00	ppb	.000		=
*** Sample ID: 28308								
				Seq: 895		16:13:22	20 Nov 01	HG
Hg	-.180	ppb	.000					=
*** Sample ID: 28309								
				Seq: 896		16:16:03	20 Nov 01	HG
Hg	-.170	ppb	.000					=
*** Sample ID: 28310								
				Seq: 897		16:18:14	20 Nov 01	HG
Hg	-.148	ppb	.000					=
*** Sample ID: 28322								
				Seq: 898		16:19:51	20 Nov 01	HG
Hg	-.140	ppb	.000					=
*** Sample ID: 28322d ✓								
				Seq: 899		16:22:19	20 Nov 01	HG
Hg	-.142	ppb	.000					=
*** Sample ID: 28322 spk ✓ 98%								
				Seq: 900		16:23:55	20 Nov 01	HG
Hg	4.91	ppb	.000					=
*** Sample ID: 28323								
				Seq: 901		16:25:32	20 Nov 01	HG
Hg	-.166	ppb	.000					=

westchester county labs


Folder: RUNS1

Page 135

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Sample ID: 28324				Seq: 902	16:27:08	20 Nov 01	HG	
Hg	-.153	ppb	.000	-.153				
=====								
*** Sample ID: 28325				Seq: 903	16:29:20	20 Nov 01	HG	
Hg	-.156	ppb	.000	-.156				
=====								
*** Sample ID: 28326				Seq: 904	16:31:07	20 Nov 01	HG	
Hg	-.122	ppb	.000	-.122				
=====								
*** Check Standard: 2 Ck21.0 qc				Seq: 905	16:32:47	20 Nov 01	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.9	.979	1.00	ppb	.000		
=====								
*** Sample ID: 28327				Seq: 906	16:34:53	20 Nov 01	HG	
Hg	-.162	ppb	.000	-.162				
=====								
*** Check Standard: 2 Ck21.0 qc				Seq: 916	16:52:48	20 Nov 01	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		98.5	.985	1.00	ppb	.000		
=====								



Not for any of our runs.