User: Galos, Marie
Westchester Dept. of Labs & Research
Date: 11/28/2001 Time: 10:09:00

Sample: AD25604
Analysis: $GCMS GCMS Scan For Unknowns
Units: ug
Started: 11/28/01 10:08 Ended: 11/28/01 10:08 Analyst: MG
Page: 1

	Component Name	Vol	Result
1	Other unknown compounds		see comment

Comments for Sample AD25604 - Analysis $GCMS

Westchester Dept. of Labs & Research
User: Galos, Marie
Date: 11-28-2001 Time: 10:09:10

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Quantitation Report (QT Reviewed)

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D Vial: 13
Acq On : 17 Nov 2001 2:44 am Operator: 209 GALOS
Sample : gc/ms unknown Inst : GC/MS 209
Misc : Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Nov 17 3:33 2001 Quant Results File: NP103001.RES

Quant Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Last Update : Fri Nov 02 16:41:16 2001
Response via : Initial Calibration
DataAcq Meth : NP103001

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	11.71	152	692057	20.00 ug/l	-0.02
19) Naphthalene-d8	15.32	136	2606579	20.00 ug/l	-0.02
31) Acenaphthene-d10	20.60	164	1289104	20.00 ug/l	-0.02
49) Phenanthrene-d10	25.02	188	1915724	20.00 ug/l	-0.02
59) Chrysene-d12	33.05	240	1437559	20.00 ug/l	-0.02
68) Perylene-d12	37.16	264	1061520	20.00 ug/l	0.00
System Monitoring Compounds					
4) 2-Fluorophenol	0.00	112	0	0.00 ug/l	
Spiked Amount 75.000			Recovery =	0.00%	
5) Phenol-d5	0.00	99	0	0.00 ug/l	
Spiked Amount 75.000			Recovery =	0.00%	
6) 2-Chlorophenol-d4	11.71	132	377	0.01 ug/l	0.47
Spiked Amount 75.000			Recovery =	0.01%	
7) 1,2-Dichlorobenzene-d4	12.23	152	560	0.02 ug/l	-0.02
Spiked Amount 50.000			Recovery =	0.04%	
20) Nitrobenzene-d5	0.00	82	0	0.00 ug/l	
Spiked Amount 50.000			Recovery =	0.00%	
32) 2-Fluorobiphenyl	18.74	172	1778	0.02 ug/l	0.12
Spiked Amount 50.000			Recovery =	0.04%	
33) 2,4,6-Tribromophenol	0.00	330	0	0.00 ug/l	
Spiked Amount 75.000			Recovery =	0.00%	
62) Terphenyl-d14	0.00	244	0	0.00 ug/l	
Spiked Amount 50.000			Recovery =	0.00%	
Target Compounds					Qvalue
2) Pyridine	0.00	79	0	N.D.	
3) N-nitroso-dimethyl amine	0.00	74	0	N.D.	
8) Phenol	11.33	94	104	N.D.	
9) bis(2-Chloroethyl)ether	0.00	93	0	N.D.	
10) 2-Chlorophenol	0.00	128	0	N.D.	
11) 1,3-Dichlorobenzene	0.00	146	0	N.D.	
12) 1,4-Dichlorobenzene	0.00	146	0	N.D.	
13) 1,2-Dichlorobenzene	0.00	146	0	N.D.	
14) 2-Methylphenol	0.00	108	0	N.D.	
15) 2,2'-oxybis(1-Chloropropan	0.00	45	0	N.D.	
16) 3&4-Methylphenol	0.00	108	0	N.D.	
17) N-Nitroso-di-n-propylamine	0.00	70	0	N.D.	
18) Hexachloroethane	0.00	117	0	N.D.	
21) Nitrobenzene	0.00	77	0	N.D.	
22) Isophorone	0.00	82	0	N.D.	

(#) = qualifier out of range (m) = manual integration

Quantitation Report (QT Reviewed)

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D Vial: 13
Acq On : 17 Nov 2001 2:44 am Operator: 209 GALOS
Sample : gc/ms unknown Inst : GC/MS 209
Misc : Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Nov 17 3:33 2001 Quant Results File: NP103001.RES

Quant Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Last Update : Fri Nov 02 16:41:16 2001
Response via : Initial Calibration
DataAcq Meth : NP103001

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
23)	2-Nitrophenol	0.00	139	0	N.D.	
24)	2,4-Dimethylphenol	0.00	107	0	N.D.	
25)	bis(2-Chloroethoxy)methane	0.00	93	0	N.D.	
26)	2,4-Dichlorophenol	0.00	162	0	N.D.	
27)	1,2,4-Trichlorobenzene	0.00	180	0	N.D.	
28)	Naphthalene	15.37	128	238	N.D.	
29)	Hexachlorobutadiene	0.00	225	0	N.D.	
30)	4-Chloro-3-methylphenol	0.00	107	0	N.D.	
34)	Hexachlorocyclopentadiene	0.00	237	0	N.D.	
35)	2,4,6-Trichlorophenol	0.00	196	0	N.D.	
36)	2,4,5-Trichlorophenol	0.00	196	0	N.D.	
37)	2-Chloronaphthalene	0.00	162	0	N.D.	
38)	Dimethylphthalate	0.00	163	0	N.D.	
39)	2,6-Dinitrotoluene	0.00	165	0	N.D.	
40)	Acenaphthylene	0.00	152	0	N.D.	
41)	Acenaphthene	0.00	153	0	N.D.	
42)	2,4-Dinitrophenol	0.00	184	0	N.D.	
43)	4-Nitrophenol	0.00	65	0	N.D.	
44)	2,4-Dinitrotoluene	20.93	165	221	N.D.	
45)	Diethylphthalate	0.00	149	0	N.D.	
46)	4-Chlorophenyl-phenylether	0.00	204	0	N.D.	
47)	Fluorene	0.00	166	0	N.D.	
48)	Azobenzen/ 1,2-Diphenylhyd	0.00	77	0	N.D.	
50)	4,6-Dinitro-2-methylphenol	0.00	198	0	N.D.	
51)	N-Nitrosodiphenylamine	0.00	168	0	N.D.	
52)	4-Bromophenyl-phenylether	0.00	248	0	N.D.	
53)	Hexachlorobenzene	0.00	284	0	N.D.	
54)	Pentachlorophenol	0.00	266	0	N.D.	
55)	Phenanthrene	25.08	178	693	N.D.	
56)	Anthracene	25.08	178	694	N.D.	
57)	Di-n-butylphthalate	27.04	149	1827	N.D.	
58)	Fluoranthene	0.00	202	0	N.D.	
60)	3,3'-Dichlorobenzidine	0.00	252	0	N.D.	
61)	Benzidine	0.00	184	0	N.D.	
63)	Pyrene	0.00	202	0	N.D.	
64)	Butylbenzylphthalate	0.00	149	0	N.D.	
65)	Benzo(a)anthracene	33.05	228	3646	N.D.	
66)	Bis(2-ethylhexyl)phthalate	33.33	149	400	N.D.	
67)	Chrysene	33.05	228	3646	N.D.	
69)	Di-n-Octylphthalate	0.00	149	0	N.D.	
70)	Benzo(b)fluoranthene	0.00	252	0	N.D.	

(#) = qualifier out of range (m) = manual integration

Quantitation Report (QT Reviewed)

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D Vial: 13
Acq On : 17 Nov 2001 2:44 am Operator: 209 GALOS
Sample : gc/ms unknown Inst : GC/MS 209
Misc : Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Nov 17 3:33 2001 Quant Results File: NP103001.RES

Quant Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Last Update : Fri Nov 02 16:41:16 2001
Response via : Initial Calibration
DataAcq Meth : NP103001

Compound	R.T.	QIon	Response	Conc Unit	Qvalue
71) Benzo(k)fluoranthene	0.00	252	0	N.D.	
72) Benzo(a)pyrene	37.16	252	4517	N.D.	
73) Indeno(1,2,3-cd)pyrene	0.00	276	0	N.D.	
74) Dibenz(a,h)anthracene	0.00	278	0	N.D.	
75) Benzo(g,h,i)perylene	0.00	276	0	N.D.	

(#) = qualifier out of range (m) = manual integration

Quantitation Report

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D Vial: 13
Acq On : 17 Nov 2001 2:44 am Operator: 209 GALOS
Sample : gc/ms unknown Inst : GC/MS 209
Misc : Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Nov 17 3:33 2001 Quant Results File: NP103001.RES

Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Last Update : Fri Nov 02 16:41:16 2001
Response via : Initial Calibration



LSC Area Percent Report

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D Vial: 13
Acq On : 17 Nov 2001 2:44 am Operator: 209 GALOS
Sample : gc/ms unknown Inst : GC/MS 209
Misc : Multiplr: 1.00
MS Integration Params: LSCINT.P

Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Smoothing : OFF Filtering: 5
Sampling : 1 Min Area: 1 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	peak area	peak % max.	% of total
1	6.320	135	139	146	rBV	29872	67425	1.30%	0.243%
2	6.421	147	150	161	rVV2	43993	130217	2.52%	0.468%
3	6.558	161	165	173	rVV	21725	56568	1.09%	0.204%
4	6.669	173	177	186	rVB	30786	70605	1.36%	0.254%
5	7.688	284	288	305	rBV	147373	447186	8.64%	1.609%
6	11.709	721	726	745	rBV	1359591	3586526	69.33%	12.903%
7	15.316	1113	1119	1137	rBV	2154303	4752548	91.86%	17.098%
8	20.595	1688	1694	1714	rBV	2357684	5173434	100.00%	18.612%
9	25.020	2169	2176	2188	rBV2	2242950	4987985	96.42%	17.945%
10	33.053	3044	3051	3068	rBV2	1892822	4518786	87.35%	16.257%
11	37.156	3489	3498	3524	rBV2	1389290	4004905	77.41%	14.408%

Sum of corrected areas: 27796185

25604.D NP103001.M Tue Nov 27 09:43:22 2001

LSC Report - Integrated Chromatogram

File : C:\HPCHEM\1\DATA\NOV1601\25604.D
Operator : 209 GALOS
Acquired : 17 Nov 2001 2:44 am using AcqMethod NP103001
Instrument : GC/MS 209
Sample Name: gc/ms unknown
Misc Info :
Vial Number: 13
Quant File :NP103001.RES (RTE Integrator)



25604.D NP103001.M Tue Nov 27 09:43:25 2001

Library Search Compound Report

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D
Acq On : 17 Nov 2001 2:44 am
Sample : gc/ms unknown
Misc :
MS Integration Params: LSCINT.P

Vial: 13
Operator: 209 GALOS
Inst : GC/MS 209
Multiplr: 1.00

Quant Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Library : C:\DATABASE\NBS75K.L

****

Peak Number 1 2-Hexanone Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.42	0.73 ug/l	130217	1,4-Dichlorobenzene-d4	11.71

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	2-Hexanone	100	C6H12O	000591-78-6	91
2	2-Hexanone, 4-hydroxy-3-propyl-	158	C9H18O2	062338-17-4	64
3	Methyl Isobutyl Ketone	100	C6H12O	000108-10-1	58
4	Oxirane, 2-methyl-2-(1-methylethyl)	100	C6H12O	072221-03-5	50



Library Search Compound Report

Data File : C:\HPCHEM\1\DATA\NOV1601\25604.D
Acq On : 17 Nov 2001 2:44 am
Sample : gc/ms unknown
Misc :
MS Integration Params: LSCINT.P

Vial: 13
Operator: 209 GALOS
Inst : GC/MS 209
Multiplr: 1.00

Quant Method : C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title : 209 625/TCLP calibration table
Library : C:\DATABASE\NBS75K.L

****

Peak Number 2 2-Pentanone, 4-hydroxy-4-methy Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
7.69	2.49 ug/l	447186	1,4-Dichlorobenzene-d4	11.71

Hit# of 5	Tentative ID	MW	MolForm	CAS#	Qual
1	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	42
2	Acetone	58	C3H6O	000067-64-1	9
3	Morpholine, 4-methyl-	101	C5H11NO	000109-02-4	9
4	Acetic acid, 1,1-dimethylethyl este	116	C6H12O2	000540-88-5	9



Tentatively Identified Compound (LSC) summary

Operator ID: .209 GALOS Date Acquired: 17 Nov 2001 2:44 am
Data File: C:\HPCHEM\1\DATA\NOV1601\25604.D
Name: gc/ms unknown
Misc:
Method: C:\HPCHEM\1\METHODS\NP103001.M (RTE Integrator)
Title: 209 625/TCLP calibration table
Library Searched: C:\DATABASE\NBS75K.L

TIC Top Hit name	RT	EstConc	Units	Area	IntStd	ISRT	ISArea	ISConc
2-Hexanone	6.42	0.7	ug/l	130217	ISTD01	11.71	3586530	20.0
2-Pentanone, 4-hydro	7.69	2.5	ug/l	447186	ISTD01	11.71	3586530	20.0

25604.D NP103001.M Tue Nov 27 09:43:31 2001